Demystifying back scatter interferometry: a sensitive refractive index detector.

Jepsen, Søren Terpager; Jørgensen, Thomas Martini; Trydal, Torleif; Sørensen, Henrik Schiøtt; Kristensen, Søren Risom

Published in:
Clinical Chemistry and Laboratory Medicine

Link to article, DOI:
[10.1515/cclm-2014-4004](https://doi.org/10.1515/cclm-2014-4004)

Publication date:
2014

Document Version
Peer reviewed version

[Link back to DTU Orbit](#)

Citation (APA):
Jepsen, S. T., Jørgensen, T. M., Trydal, T., Sørensen, H. S., & Kristensen, S. R. (2014). Demystifying back scatter interferometry: a sensitive refractive index detector. *Clinical Chemistry and Laboratory Medicine*, 52(Supplement), [S215]. <https://doi.org/10.1515/cclm-2014-4004>

General rights

Copyright and moral rights for the publications made accessible in the public portal are retained by the authors and/or other copyright owners and it is a condition of accessing publications that users recognise and abide by the legal requirements associated with these rights.

- Users may download and print one copy of any publication from the public portal for the purpose of private study or research.
- You may not further distribute the material or use it for any profit-making activity or commercial gain
- You may freely distribute the URL identifying the publication in the public portal

If you believe that this document breaches copyright please contact us providing details, and we will remove access to the work immediately and investigate your claim.

Advanced technology

Cod: 0014

DEMYSTIFYING BACK SCATTER INTERFEROMETRY - A SENSITIVE REFRACTIVE INDEX DETECTORS.T. Jepsen¹, T.M. Jørgensen², T. Trydal¹, H.S. Sørensen², S.R. Kristensen¹¹Dept. of Clinical Biochemistry, Aalborg University Hospital²Dept. of Photonic Engineering, Technical University of Denmark

BACKGROUND: Back Scatter Interferometry (BSI) is a sensitive method for detecting changes of the refractive index (RI) in small capillaries. The method was originally developed as an off-axial column detector for use in Liquid Chromatography or Capillary Electrophoresis systems, but it has been proposed that this method can also be used to detect molecular binding in a label-free manner. Recent work proposes BSI to be a unique sensor for detecting protein binding with various ligands and other protein interactions in order to obtain relevant binding kinetics. We hypothesize that BSI is actually acting like a common-path interferometer.

METHODS: A HeNe laser is directed at a glass capillary with inner diameter of 1.4 mm and reflected light from air/glass and liquid/glass interfaces interfere to form an RI dependent intensity fringe pattern at a CCD detector. The fringe shift relative to the change of RI of the sample; i.e. the sensitivity; is controlled by the physical interaction length between the interferometric sample beam and the sample itself. Using optical ray-tracing we calculate the sensitivity. We validate these theoretical findings by determining the RI increment (dn/dc) from a set of NaCl standard solutions.

RESULTS: Ray-tracing show that the basic interference pattern recorded with BSI can be fully described by two beams, one reflected from the surface of the capillary and a beam reflected from the back of the capillary wall. In accordance we find that the interferometric interaction length is given by twice the diameter of the capillary. Experimentally we find a sensitivity of 4700 rad/(g/ml) and estimate dn/dc for NaCl to be 0.169 ml/g, which is in accordance with literature. Furthermore we report a minimum detectability of 7×10^{-7} RI Units.

CONCLUSIONS: BSI works like a common-path interferometer. The sensitivity of the BSI system is given by twice the inner diameter of the capillary times the wavenumber of the light source. Our results suggest that Back Scatter Interferometry does not provide a unique measurement principle for sensing biochemical bindings compared to what should be possible using many commercial available refractometers.