Sensitivity analysis of magnetic field measurements for magnetic resonance electrical impedance tomography (MREIT)

Göksu, Cihan; Scheffler, Klaus; Ehses, Philipp; Hanson, Lars G. ; Thielscher, Axel

Published in:
Magnetic Resonance in Medicine

Link to article, DOI:
[10.1002/mrm.26727](https://doi.org/10.1002/mrm.26727)

Publication date:
2017

Document Version
Peer reviewed version

[Link back to DTU Orbit](#)

Citation (APA):
Göksu, C., Scheffler, K., Ehses, P., Hanson, L. G., & Thielscher, A. (2017). Sensitivity analysis of magnetic field measurements for magnetic resonance electrical impedance tomography (MREIT). *Magnetic Resonance in Medicine*, 79(2), 748–760. <https://doi.org/10.1002/mrm.26727>

General rights

Copyright and moral rights for the publications made accessible in the public portal are retained by the authors and/or other copyright owners and it is a condition of accessing publications that users recognise and abide by the legal requirements associated with these rights.

- Users may download and print one copy of any publication from the public portal for the purpose of private study or research.
- You may not further distribute the material or use it for any profit-making activity or commercial gain
- You may freely distribute the URL identifying the publication in the public portal

If you believe that this document breaches copyright please contact us providing details, and we will remove access to the work immediately and investigate your claim.

Sensitivity Analysis of Magnetic Field Measurements for Magnetic Resonance Electrical Impedance Tomography (MREIT)

Cihan Göksu,^{1,2} Klaus Scheffler,^{3,4} Philipp Ehses,^{3,4} Lars G. Hanson,^{1,2†} and Axel Thielscher^{1,2,3†*}

Purpose: Clinical use of magnetic resonance electrical impedance tomography (MREIT) still requires significant sensitivity improvements. Here, the measurement of the current-induced magnetic field ($\Delta B_{z,c}$) is improved using systematic efficiency analyses and optimization of multi-echo spin echo (MESE) and steady-state free precession free induction decay (SSFP-FID) sequences.

Theory and Methods: Considering T_1 , T_2 , and T_2^* relaxation in the signal-to-noise ratios (SNRs) of the MR magnitude images, the efficiency of MESE and SSFP-FID MREIT experiments, and its dependence on the sequence parameters, are analytically analyzed and simulated. The theoretical results are experimentally validated in a saline-filled homogenous spherical phantom with relaxation parameters similar to brain tissue. Measurement of $\Delta B_{z,c}$ is also performed in a cylindrical phantom with saline and chicken meat.

Results: The efficiency simulations and experimental results are in good agreement. When using optimal parameters, $\Delta B_{z,c}$ can be reliably measured in the phantom even at injected current strengths of 1 mA or lower for both sequence types. The importance of using proper crusher gradient selection on the phase evolution in a MESE experiment is also demonstrated.

Conclusion: The efficiencies observed with the optimized sequence parameters will likely render in-vivo human brain MREIT feasible. **Magn Reson Med** 000:000–000, 2017. © 2017 International Society for Magnetic Resonance in Medicine.

Key words: efficiency analysis; magnetic resonance electrical impedance tomography; multi-echo spin echo; steady-state free precession; sequence optimization

INTRODUCTION

Magnetic resonance current density imaging (MRCDI) and magnetic resonance electrical impedance tomography (MREIT) are two emerging imaging modalities, which combine MRI with externally applied currents (either direct current or alternating current at low frequencies combined with repeated refocusing pulses (1)) to reconstruct the current density distribution and ohmic conductivity variation inside body tissue (2–7). This may open up novel ways to characterize pathological tissue (8). In addition, better knowledge of the conductivity distribution would allow improving the accuracy of source localization methods for electroencephalography and magnetoencephalography (9) and enable better spatial targeting of neurostimulation methods (10,11). However, MRCDI and MREIT are still hampered by their low sensitivity, which prevents their clinical usage.

In both modalities, electrical current is applied in synchrony with the MRI pulse sequence. The current flow induces a magnetic field distribution in the body, and the component of the induced magnetic field ($\Delta B_{z,c}$) which is parallel to the main magnetic field (B_0) creates a phase perturbation in the MRI signal that can be measured (5). The sensitivity of the $\Delta B_{z,c}$ measurement directly affects the accuracy and quality of the reconstructed current and conductivity distributions (12). However, a reliable $\Delta B_{z,c}$ measurement in in-vivo situations is crucial and challenging as only weak currents can be applied to the human body in the low frequency range, e.g. around 1–2 mA for brain studies (13). Optimized MR sequences which allow for efficient $\Delta B_{z,c}$ measurements within clinically relevant scan times are thus important to enable in-vivo applications of MRCDI and MREIT.

Up to now, single-echo spin echo (SE), multi-echo spin echo (MESE), gradient recalled echo, echo planar imaging, and steady-state free precession free induction decay (SSFP-FID) MREIT experiments have been performed (5,14–19). Sequences with refocusing pulses are more robust to main field inhomogeneities and have a higher signal-to-noise ratio (SNR), but imaging time is prolonged. On the other hand, the gradient-echo sequences are more vulnerable to main field inhomogeneities and have less SNR attributed to T_2^* decay, but are generally faster.

In this study, systematic efficiency analyses of two sensitive sequences (MESE and SSFP-FID) are performed, thereby considering the impact of T_1 , T_2 , and T_2^* relaxation and radiofrequency (RF) imperfections on

¹Danish Research Center for Magnetic Resonance, Center for Functional and Diagnostic Imaging and Research, Copenhagen University Hospital, Hvidovre, Denmark.

²Center for Magnetic Resonance, DTU Elektro, Technical University of Denmark, Kgs Lyngby, Denmark.

³High-Field Magnetic Resonance Center, Max-Planck-Institute for Biological Cybernetics, Tübingen, Germany.

⁴Department of Biomedical Magnetic Resonance, University of Tübingen, Tübingen, Germany.

Grant sponsor: Lundbeck Foundation; Grant numbers: R118-A11308; R59 A5399 (PI Hartwig Siebner).

*Correspondence to: Axel Thielscher, Danish Research Center for Magnetic Resonance, Center for Functional and Diagnostic Imaging and Research, Copenhagen University Hospital Hvidovre, Section 714, Kettegaard Allé 30, 2650 Hvidovre, Denmark. E-mail: axelt@drbcmr.dk

[†]These authors contributed equally to this work.

Received 30 November 2016; revised 2 March 2017; accepted 29 March 2017

DOI 10.1002/mrm.26727

Published online 00 Month 2017 in Wiley Online Library (wileyonlinelibrary.com).

© 2017 International Society for Magnetic Resonance in Medicine

FIG. 1. (a) Diagram of the MESE MREIT pulse sequence with equal and symmetric echo spacing. The sequence is composed of a 90° excitation pulse preceding repetitive 180° refocusing pulses, so that multiple echoes are created. Crusher gradients are used to preserve the desired echo pathways, while eliminating unwanted ones caused by nonideal refocusing pulses (27). At the end of the sequence, phase encoding rewinder and spoiler gradients are used to eliminate unwanted effects of remaining transverse magnetization. This is followed by a dead time T_D after which the slice (or the next slice for multi-slice measurements; see subfigure c) is excited again. The injected bipolar electrical current is synchronized with the RF pulses, so that the phase of the continuous complex transverse magnetization (χ_μ) increases linearly over time. (b) Sequence diagrams of the two SSFP-FID variants. An SSFP sequence is composed of repetitive constant tip angle and in-phase excitation pulses, where the interval T_R between each RF pulse is constant. These conditions are enough to reach a steady state (30). In case of a bipolar electrical current injection in synchrony with the SSFP-FID sequence, the continuous transverse magnetization phase evolves in opposite directions in odd and even T_R periods, which induces two different steady-state conditions with opposite current-induced phases. Please note that unlike in the original study of Lee et al (19), we decided to inject electrical current until T_E for SSFP-FID_{PCI} in order to test its most efficient case. On the other hand, the current is injected within the entire T_R period in SSFP-FID_{FCI}. (c) Interleaved multi-slice acquisition of MESE.

SNR. All results are experimentally validated in a saline-filled homogenous spherical phantom with relaxation parameters similar to brain tissue. For MESE (Fig. 1a,c), it is simulated how efficiency depends on the relevant sequence parameters, which are shown to be the echo spacing T_{ES} , the number of echoes N_{echo} , and the dead time T_D . The efficiency change for multi-slice acquisition is subsequently assessed. In addition, the importance of selecting the proper echo pathways on the phase evolution is demonstrated. Furthermore, two different SSFP-FID variants (Fig. 1b) are simulated and compared. The more efficient variant is subsequently optimized with respect to the utilized tip angle α , echo time T_E , and repetition time T_R . In final experiments, the efficiencies of the optimized MESE and SSFP-FID sequences are directly compared, and $\Delta B_{z,c}$ measurements are performed for both MESE and SSFP-FID for a nonhomogeneous phantom.

THEORY

Efficiency of an MREIT Experiment

We use the following notation of efficiency η_{seq} to characterize the performance of a sequence (20), thereby relating the SNR of the acquired $\Delta B_{z,c}$ image to the required total scan time T_{tot} (Eq. 1):

$$\eta_{seq} = \frac{SNR_{\Delta B_{z,c}}}{\sqrt{T_{tot}}} = \frac{|\Delta B_{z,c}|}{\sigma_{\Delta B_{z,c}} \sqrt{T_{tot}}} \quad [1]$$

$|\Delta B_{z,c}|$ is the magnitude of the current-induced magnetic field and $\sigma_{\Delta B_{z,c}}$ the noise standard deviation of $\Delta B_{z,c}$. Please note that η_{seq} varies spatially, because $\Delta B_{z,c}$ depends on the injected current strength, electrode placement, electrode geometry, and conductivity distribution. In addition, $\sigma_{\Delta B_{z,c}}$ depends on the SNR of the MR image and the phase sensitivity of MRI sequence. In the following, we derive how the

efficiency depends on sequence and tissue relaxation parameters. The resulting equations are then used to determine the optimal parameter settings by numerical simulations.

MESE MREIT

The pulse sequence of a standard MESE MREIT sequence with equal and symmetric echo spacing is shown in Figure 1a,c. For current injection within periods free of RF pulses, the measured $\Delta B_{z,c}$ from each single echo and its noise variance were reported previously (17,21,22) and are given by Equations [2] and [3]:

$$\Delta B_{z,c}^n = \frac{\angle M_n^+ - \angle M_n^-}{2\gamma[(T_{ES} - \tau_\pi)n - 0.5\tau_{\pi/2}]} \quad [2]$$

$$\text{Var}\{\Delta B_{z,c}^n\} = \frac{1}{4\gamma^2 \text{SNR}_n^2 [(T_{ES} - \tau_\pi)n - 0.5\tau_{\pi/2}]^2} \quad [3]$$

$\angle M_n^+$ and $\angle M_n^-$ are the phase of the complex MR images from the n th echo with positive (+) and negative (−) constant current injection, SNR_n is the SNR of the magnitude image from the n th echo, and γ denotes the gyromagnetic ratio. τ_π and $\tau_{\pi/2}$ are the durations of the 180° and 90° RF pulses where current is not applied.

The $\Delta B_{z,c}$ measurements can be optimally combined across echoes by weighting each by the inverse of its variance. Normalizing by a common factor to ensure that the weights across all echoes sum to 1, and adding the weighted images, the noise variance of the combined $\Delta B_{z,c}$ is then given by Equation [4] (17):

$$\text{Var}\{\Delta B_{z,c}^{\text{comb}}\} = \frac{1}{\sum_{n=1}^{N_{\text{echo}}} 4\gamma^2 \text{SNR}_n^2 [(T_{ES} - \tau_\pi)n - 0.5\tau_{\pi/2}]^2} \quad [4]$$

Applying Equation [4] to Equation [1] finally gives the efficiency of measuring the combined $\Delta B_{z,c}$ (Eq. [5]):

$$\eta_{\text{MESE}} = \frac{|\Delta B_{z,c}^{\text{comb}}|}{\sqrt{T_{\text{tot}}}} \sqrt{\sum_{n=1}^{N_{\text{echo}}} 4\gamma^2 \text{SNR}_n^2 [(T_{ES} - \tau_\pi)n - 0.5\tau_{\pi/2}]^2} \quad [5]$$

In order to further relate the efficiency stated in Equation [5] to the sequence and tissue relaxation parameters, the 1D case is considered. The continuous complex transverse magnetization $\mu(x, t)$ depends on T_1 , T_2 , T_2^* relaxations and a signal loss factor caused by imperfect refocusing. Defining β_{RF} to be the fraction of preserved signal after each refocusing pulse, and under an assumption of a Lorentzian spectral density distribution (Eq. [6]),

$$\mu^n(x, t) = \mu_0(x) [1 - e^{-T_{\text{rec}}/T_1(x)}] e^{-nT_{ES}/T_2(x)} e^{-|t-nT_{ES}|/T_2^*(x)} \beta_{\text{RF}}^n \quad \text{for } (n - 0.5)T_{ES} < t < (n + 0.5)T_{ES} \quad [6]$$

$\mu_0(x)$ is the equilibrium magnetization distribution and T_{rec} is the T_1 recovery period between nulling of longitudinal magnetization after the last refocusing pulse and re-excitation of the same slice (the factor β_{RF}

expresses the accumulated effect of imperfect refocusing pulses in later echoes). The acquired signal $S(k_x, t)$ for the n th MESE echo can then be expressed as Equation [7]:

$$S^n(k_x, t) = \int_{\text{object}} \mu^n(x, t) e^{-j2\pi k_x x} dx \quad [7]$$

The object can conceptually be considered a distribution of point sources. Combining Equations [6] and [7] and assuming an idealized single point distribution ($\mu_0(x) = \delta(x)$), constant relaxation times, β_{RF} , and noise σ , and a standard k -space trajectory $k_x(t) = \frac{\gamma G_x}{2\pi} (t - nT_{ES})$ results in the conclusion that the SNR_n in Equation [5] is proportional to attenuation factors (α_{T_1} , α_{T_2} , $\alpha_{T_2^*}$ and α_{RF}) caused by the T_1 , T_2 , T_2^* relaxations and RF imperfections, which can be expressed as Equation [8]:

$$\begin{aligned} \text{SNR}_n &= \frac{|M_n|}{\sigma} \propto \alpha_{T_1} \alpha_{T_2} \alpha_{T_2^*} \alpha_{\text{RF}} \\ \alpha_{T_1} &= 1 - e^{-T_{\text{rec}}/T_1}, \quad \alpha_{T_2} = e^{-nT_{ES}/T_2}, \quad \alpha_{\text{RF}} = \beta_{\text{RF}}^n \\ \alpha_{T_2^*} &= \frac{1}{N_x \Delta k_x} \int_{-N_x \Delta k_x/2}^{N_x \Delta k_x/2} e^{-2\pi |k_x|/\gamma G_x T_2^*} dk_x = \frac{2T_2^* (1 - e^{-T_s/2T_2^*})}{T_s} \end{aligned} \quad [8]$$

$|M_n|$ is the noise-free reconstructed MR magnitude image, which is proportional to μ^n given in Equation [6]. G_x is the readout gradient strength, N_x the readout matrix size, T_s the readout period, and Δk_x the spatial frequency resolution. The recovery of the longitudinal magnetization is almost linear for $T_{ES} \ll T_1$ within the period between refocusing pulses. Therefore, it can be assumed that the longitudinal magnetization is nulled at each echo, and T_{rec} can be approximated as shown by Equation [9]:

$$T_{\text{rec}} \approx (N_{\text{slice}} - 1)N_{\text{echo}}T_{ES} + N_{\text{slice}}T_D \quad [9]$$

N_{slice} and T_D are number of slices and dead time, respectively. In combination, Equations [5], [8], and [9] characterize the dependency of the efficiency of a MESE MREIT experiment on the sequence and tissue parameters.

SSFP-FID MREIT

Lee et al have previously studied different SSFP variants for MREIT (19). Here, we investigate their most sensitive variant further, in which the current is applied before the readout period (SSFP-FID_{PCI} with partial current injection; Fig. 1b). In addition, we propose a novel variant in which the current is injected within the entire T_R period (SSFP-FID_{FCI} with full current injection; Fig. 1b). The analytical solutions for the steady-state magnetization immediately after excitation with bipolar current injection have been derived by Lee et al (19) (Eq. [10]):

$$\mu_{ss1}(t=0^+) = \frac{\mu_0(1-E_1)\sin(\alpha)}{D} \begin{pmatrix} A_1 e^{-2j(\varphi_g+\varphi_b)} + A_2 e^{-2j\varphi_c} + A_3 e^{j(\varphi_g+\varphi_b-\varphi_c)} \\ + A_4 e^{-j(\varphi_g+\varphi_b-\varphi_c)} + A_5 e^{-j(\varphi_g+\varphi_b+\varphi_c)} + A_6 \end{pmatrix}$$

$$\mu_{ss2}(t=0^+) = \frac{\mu_0(1-E_1)\sin(\alpha)}{D} \begin{pmatrix} A_1 e^{-2j(\varphi_g+\varphi_b)} + A_2 e^{2j\varphi_c} + A_3 e^{j(\varphi_g+\varphi_b+\varphi_c)} \\ + A_4 e^{-j(\varphi_g+\varphi_b+\varphi_c)} + A_5 e^{-j(\varphi_g+\varphi_b-\varphi_c)} + A_6 \end{pmatrix}$$

with

$$A_1 = E_2^2(1+E_1)(1+\cos(\alpha)), A_2 = E_2^2(1-E_1)(1-\cos(\alpha))$$

$$A_3 = -E_2(1+E_1)(1+\cos(\alpha)), A_4 = E_2(1-E_1)(1-\cos(\alpha)) \quad [10]$$

$$A_5 = 2E_2^3(E_1+\cos(\alpha)), A_6 = -2(1+E_1\cos(\alpha))$$

$$D = E_2^2(1-E_1^2)[(\cos(\alpha)+1)^2 \cos(2(\varphi_g+\varphi_b)) + (\cos(\alpha)-1)^2 \cos(2\varphi_c)] +$$

$$2E_1E_2(1-E_2^2)(\cos(2\alpha)-1)\cos(\varphi_g+\varphi_b)\cos(\varphi_c) +$$

$$2(E_1\cos(\alpha)+1)(E_1\cos(\alpha)-1) + 2E_2^4(E_1+\cos(\alpha))(E_1-\cos(\alpha))$$

and

$$E_1 = e^{-T_R/T_1}, E_2 = e^{-T_R/T_2}$$

Here, μ_{ss1} and μ_{ss2} are the alternating first and second steady-state transversal magnetizations; μ_0 is the thermal equilibrium magnetization, α the tip angle, φ_g the gradient-induced phase, φ_b the B_0 inhomogeneity-induced phase, and φ_c the current-induced phase. The steady-state magnetization at T_E becomes (Eq. [11]):

$$\mu_{SS1}^{FID_1}(\Delta B_{z,c}, t=T_E) = \mu_{ss1}(\varphi_c = \gamma\Delta B_{z,c}T_c, t=0^+)e^{-T_E/T_2}e^{j\gamma(\Delta B_0+\Delta B_{z,c})T_E}$$

$$\mu_{SS2}^{FID_1}(\Delta B_{z,c}, t=T_E) = \mu_{ss2}(\varphi_c = \gamma\Delta B_{z,c}T_c, t=0^+)e^{-T_E/T_2}e^{j\gamma(\Delta B_0-\Delta B_{z,c})T_E}$$

$$\mu_{SS1}^{FID_2}(\Delta B_{z,c}, t=T_E) = \mu_{ss1}(\varphi_c = \gamma\Delta B_{z,c}T_R, t=0^+)e^{-T_E/T_2}e^{j\gamma(\Delta B_0+\Delta B_{z,c})T_E}$$

$$\mu_{SS2}^{FID_2}(\Delta B_{z,c}, t=T_E) = \mu_{ss2}(\varphi_c = \gamma\Delta B_{z,c}T_R, t=0^+)e^{-T_E/T_2}e^{j\gamma(\Delta B_0-\Delta B_{z,c})T_E}$$

[11]

where T_c is the injected current pulse width and ΔB_0 the local B_0 inhomogeneity. Assuming sufficiently strong spoiler gradients creating a uniform intravoxel phase distribution at the end of each repetition, the SSFP-FID signal is equal to the integral of the steady-state magnetization with respect to φ_g over a 2π interval (19,23). Therefore, constant phase shifts attributed to RF phase imperfections or local B_0 inhomogeneity do not influence the steady-state signal.

In contrast to MESE, SSFP-FID has a nonlinear dependence of $\Delta B_{z,c}$ on the phase of the transverse magnetization. However, for weak currents, this can be well approximated by a linear relationship (Eq. [12]):

$$\Delta B_{z,c} = \frac{\angle M_{SS1} - \angle M_{SS2}}{m_{seq}} \quad [12]$$

where $\angle M_{SS1}$ and $\angle M_{SS2}$ are the phases of the first and second steady-state complex MR images and $m_{seq} = \partial(\angle M_{SS1} - \angle M_{SS2}) / \partial \Delta B_{z,c}$ express the field dependence on the phase change. The standard deviation of

the $\Delta B_{z,c}$ estimate and the efficiency can then be calculated as Equation [13]:

$$\sigma_{\Delta B_{z,c}} = \frac{1}{m_{seq} \text{SNR}},$$

$$\eta_{SSFP-FID} = \frac{|\angle M_{SS1} - \angle M_{SS2}|}{\sqrt{T_{tot}}} \text{SNR} \quad [13]$$

with SNR being the SNR of the magnitude image.

METHODS

In this section, the numerical simulation methods are introduced, which were used to systematically evaluate the efficiencies of MESE and SSFP-FID MREIT based on the above theory. This is followed by a description of the experimental methods used to validate the theory and simulations.

Simulations

The efficiency of MESE was simulated based on Equations [5], [8], and [9]. Relaxation times of $T_1 = 1.1$ seconds, $T_2 = 100$ ms, and $T_2^* = 50$ ms were used, similar to those of brain tissue (24). The RF pulse durations and their efficiency were set to $\tau_{\pi/2} = 2.048$ ms, $\tau_{\pi} = 2.56$ ms, and $\beta_{RF} = 0.86$ to match those of the clinical 3T scanner used in the MESE experiments (the MESE section also describes the measurement of β_{RF}). The longest crusher gradient duration was set to $T_{crush} = 7.5$ ms, and this value was determining the minimal echo spacing in the simulations. Because the SNR of an MR image scales with the square root of data acquisition time (21), the lowest possible sampling bandwidths (BW) were chosen in all simulations.

FIG. 2. (a) Photograph of the current source. (b) Phantom 1 was spherical with a diameter of 10 cm, filled with saline (1.45 g/L of NaCl) and doped with 0.1 mM of MnCl_2 to reach relaxation times of $T_1 = 1.1$ s and $T_2 = 100$ ms (31). T_1 values were determined by repeating an inversion recovery gradient recalled echo (IR-GRE) sequence for different inversion times. T_2 values were measured by repeating a spin echo sequence for a range of echo times. Also, the tip angle variation over the imaging region was investigated using a double-angle method (32). The tip angle map was created by repeating an RF spoiled fast low angle shot sequence with two different tip angles ($\alpha = 30^\circ$ and $\alpha = 60^\circ$; $T_E = 5$ ms; $T_R = 5$ seconds). The tip angle deviation over the imaging region was around 10%. (c) Phantom 2 was cylindrical with 10 cm in diameter and 3 cm in height, filled with the same saline solution and with a piece of organic chicken meat placed in its center. The relaxation parameters in Phantom 2 were around $T_1 = 1.05$ seconds, $T_2 = 110$ ms in the saline region, and $T_1 = 1.1$ seconds, $T_2 = 50$ ms in the chicken meat.

- For a single-slice acquisition, the SNR and efficiency of $\Delta B_{z,c}$ were simulated for a fixed dead time $T_D = 510$ ms to demonstrate the effects of N_{echo} and T_{ES} . The simulations were performed for $T_{ES} = [20-160]$ ms and $N_{\text{echo}} = [1-8]$.
- For multi-slice acquisitions, the dependence of the efficiency on T_D was obtained for different numbers of slices $N_{\text{slice}} = [1-5, 15]$, whereas the T_{ES} and N_{echo} giving the highest efficiency were selected for each T_D . The simulations were performed for $T_D = [0.1-10]$ s, $T_{ES} = [20-200]$ ms, and $N_{\text{echo}} = [1-8]$ (even for $T_{ES} = 200$ ms, the assumption of a linear recovery of the longitudinal magnetization causes an error of less than 9% when reading out the optimal T_D).

The SSFP-FID simulations were performed by using 3D rotation and relaxation matrices (25) and were cross-checked by the analytically derived Equations [10] to [13]. The number of isochromates in the simulations was 100, instantaneous RF pulses were assumed, and the spoiler gradients were modeled as creating 4π intravoxel phase dispersion. Relaxation times of $T_1 = 1.1$ s, $T_2 = 100$ ms, and $T_2^* = 50$ ms were used.

- First, the dependence of the steady-state transverse magnetization magnitude and phase on $\Delta B_{z,c}$ were simulated for both SSFP-FID variants and compared with spin echo. The simulation parameters were $\alpha = 60^\circ$, $TR = 20$ ms, $TE = 10$ ms, and a range of $\Delta B_{z,c} = [-100 \text{ to } 100]$ nT was covered.
- For the more efficient variant SSFP-FID_{FCI}, SNR and efficiency of $\Delta B_{z,c}$ measurements were simulated in order to demonstrate the effect of T_R and T_E . The simulation parameters were $\alpha = 20^\circ$, $\Delta B_{z,c} = 1$ nT, $T_R = [20-260]$ ms, and $T_E = [10-140]$ ms. The RF pulse width, prephaser, and spoiler gradient durations were set to $\tau_\alpha = 2$ ms, $T_{\text{pre}} = 0.5$ ms, and $T_{\text{sp}} = 0.6$ ms, respectively. Impossible combinations of T_E and T_R (i.e., $T_E > T_R$) were ignored. The image SNR was adjusted according to a choice of lowest possible sampling BWs.
- To find the most efficient parameters settings for SSFP-FID_{FCI}, the effect of the tip angle on the efficiency was also investigated. The simulation

parameters were $\alpha = [5^\circ-90^\circ]$, $T_E = [10-120]$ ms, and $T_R = [20-1500]$ ms. For each tip angle, the normalized maximal efficiency and the corresponding T_E and T_R were selected and the results plotted with respect to α .

As a last step, we explored via simulations the loss in efficiency when using multi-gradient-echo summation by means of multiple monopolar or bipolar readout gradients to prevent image distortions resulting from using low BWs at long T_{ES} (MESE) or T_R (SSFP-FID). The number of summed echoes during a readout period N_m was varied in the range [1-16]. For MESE, $T_{ES} = 80$ ms was selected and the duration of the added prephaser gradients was $T_{\text{pre}} = 0.5$ ms. The other parameters were kept unchanged from the prior simulations. For SSFP-FID, $T_R = 160$ ms was used.

Experiments

All experiments were performed on a 3T MRI scanner (MAGNETOM Prisma; Siemens Healthcare, Erlangen, Germany) equipped with a 64-channel head coil (an adaptive combine algorithm (26) was used to combine the received MRI signals from each coil element). The current waveforms were created by an arbitrary waveform generator (33500B; Keysight Technologies, Santa Clara, CA, USA) and a home-made voltage-to-current converter (Fig. 2a), and injected into a phantom by recessed copper electrodes (Fig. 2b,c). Two different phantoms were used: Phantom 1 was spherical and filled with doped saline having relaxation times similar to brain tissue (Fig. 2b). Phantom 2 was cylindrical, filled with similar doped saline and having a piece of organic chicken meat placed in its center (Fig. 2c).

For MESE, the following experiments were performed:

- The importance of properly designed crusher gradients to prevent the impact of nonideal RF refocusing pulses on the phase evolution was demonstrated in Phantom 1. Three different MESE pulse sequences were tested and their current-induced phase evolutions over echoes were compared. In the first two sequences, the momentum of the crusher gradients

were kept constant and refocusing RF pulses of 150° and 180° were used to explore effects of B_1 inhomogeneity. In the third sequence, the momentum of the crusher gradients was systematically changed and 180° refocusing RF pulses were used: The crusher momentums were either doubled between subsequent echoes or the crusher gradient direction was switched. In all sequences, the first crusher gradient was optimized for creating a 4π intravoxel phase dispersion (27). The other sequence parameters were: field of view (FOV) = $200 \times 200 \text{ mm}^2$, image matrix = 128×128 , slice thickness $\Delta z = 3 \text{ mm}$, $N_{\text{avg}} = 1$, $N_{\text{echo}} = 7$, injected current magnitude $I_c = 1 \text{ mA}$, $T_{\text{ES}} = [40, 60] \text{ ms}$, and $T_D = 510 \text{ ms}$.

- Efficiency measurements were performed in Phantom 1. The measurement parameters were FOV = $300 \times 300 \text{ mm}^2$, image matrix = 256×256 , $\Delta z = 5 \text{ mm}$, $T_D = 510 \text{ ms}$, $N_{\text{avg}} = 1$, $N_{\text{echo}} = [1-8]$, and $I_c = 0.5 \text{ mA}$. The measurements were repeated by varying the echo spacing $T_{\text{ES}} = [20-160] \text{ ms}$ with 20-ms intervals. In each experiment, the lowest possible bandwidth (BW) was used. The experiments were repeated with opposite polarity bipolar current injection in order to eliminate systematic phase artifacts and to increase the SNR of the experiment (21). The phase evolution over echoes, combined $\Delta B_{z,c}$ across echoes, SNR of the combined $\Delta B_{z,c}$, and efficiency were determined from the measurements (the root-mean-square SNR of the combined $\Delta B_{z,c}$ and the efficiency values were calculated in the region of interest ROI shown in Figs. 4a and 7a). To estimate the preserved signal ratio β_{RF} influenced by RF inhomogeneity, the signal decay across multiple echoes for $T_{\text{ES}} = 20 \text{ ms}$ was compared with the real T_2 decay determined from the first echoes when varying T_{ES} from 20 ms to 160 ms.

For SSFP-FID, the following experiments were performed:

- SSFP-FID measurements were repeated for different current magnitudes to validate the simulated dependency on $\Delta B_{z,c}$ of the transverse magnetization phase. The measurements were performed with both SSFP-FID_{PCI} and SSFP-FID_{FCI} sequences in Phantom 1 and their phase sensitivities were compared. The sequence parameters were FOV = $375 \times 375 \text{ mm}^2$, image matrix = 256×256 , $\Delta z = 3 \text{ mm}$, $\alpha = 60^\circ$, $N_{\text{avg}} = 16 \times 2$ (16 separate averages for each steady state). The experiments were repeated for three different repetition times $T_R = [10, 30, 50] \text{ ms}$ (with $T_E = T_R/2$) and for different current magnitudes $I_c = [-10 \text{ to } 10] \text{ mA}$ with 2-mA intervals. The lowest possible BW was always selected to maximize SNR.
- Efficiency measurements were performed in Phantom 1 for SSFP-FID_{FCI}. The measurement parameters were FOV = $192 \times 192 \text{ mm}^2$, image matrix = 128×128 , $\Delta z = 3 \text{ mm}$, $\alpha = 20^\circ$, $N_{\text{avg}} = 2 \times 2$, and $I_c = 1 \text{ mA}$. The experiment was repeated for different echo times $T_E = [10-140] \text{ ms}$ with 10-ms intervals and repetition times $T_R = [20-260] \text{ ms}$ with 20-ms intervals. Impossible combinations of T_E and T_R (i.e., $T_E > T_R$) were

ignored. In each measurement, bipolar currents were injected to create dual steady states with opposite current-induced phases. From these steady-state data, phase difference images were calculated and $\Delta B_{z,c}$ was reconstructed by using $m_{\text{seq}} = \partial(\angle M_{\text{SS1}} - \angle M_{\text{SS2}}) / \partial \Delta B_{z,c}$ in the simulations. The SNR of the $\Delta B_{z,c}$ images and the efficiency were then determined.

In addition, two experiments with the optimized MESE and SSFP-FID_{FCI} sequences were performed in Phantom 1 in order to directly compare their efficiencies. The sequence parameters were FOV = $256 \times 256 \text{ mm}^2$, image matrix = 128×128 , $N_{\text{avg}} = 1 \times 2$, and $I_c = 1 \text{ mA}$. The optimized parameters were selected as $T_{\text{ES}} = 80 \text{ ms}$, $T_D = 1.5 \text{ s}$, and $N_{\text{echo}} = 3$ for MESE; and $T_E = 60 \text{ ms}$, $T_R = 120 \text{ ms}$, and $\alpha = 30^\circ$ for SSFP-FID_{FCI}.

Finally, the MESE and SSFP-FID_{FCI} experiments were performed in Phantom 2 to demonstrate the sequence performance for a nonhomogenous geometry involving a chicken meat piece. The experiments were performed for both vertical and horizontal electrical current injection. The MESE measurement parameters were $T_{\text{ES}} = 80 \text{ ms}$, $T_D = 510 \text{ ms}$, $N_{\text{echo}} = 3$, FOV = $192 \times 192 \text{ mm}^2$, image matrix = 128×128 , $\Delta z = 3 \text{ mm}$, BW = 100 Hz/pixel, $N_{\text{avg}} = 1$, and $I_c = 1 \text{ mA}$. The SSFP-FID_{FCI} measurement parameters were $\alpha = 20^\circ$, $T_E = 15 \text{ ms}$, $T_R = 30 \text{ ms}$, FOV = $192 \times 192 \text{ mm}^2$, image matrix = 128×128 , $\Delta z = 3 \text{ mm}$, BW = 100 Hz/pixel, $N_{\text{avg}} = 16 \times 2$, and $I_c = 1 \text{ mA}$.

RESULTS

MESE

As a first step, the importance of properly chosen crusher gradients is demonstrated. When keeping the crusher gradients constant, the stimulated echo pathways caused by the nonideal refocusing pulses have a clear impact on the phase evolution (Fig. 3a,b). This effect is more prominent for 150° refocusing pulses (Fig. 3a), but is also clearly visible in the later echoes for 180° refocusing pulses (Fig. 3b). In contrast, systematically doubling the area of the crusher gradients between consecutive echoes in combination with changing crusher direction (27) successfully eliminates the unwanted echo pathways, resulting in the expected linear phase increase over echoes (Fig. 3c).

The results of the efficiency simulations and measurements for a fixed T_D are shown in Figure 4. As an example, Figure 4a shows the combined $\Delta B_{z,c}$ image for eight echoes ($T_{\text{ES}} = 20 \text{ ms}$, $T_R = 670 \text{ ms}$, BW = 125 Hz/pixel, and $I_c = 0.5 \text{ mA}$). The measured $\Delta B_{z,c}$ pattern is in agreement with the current flowing from top to bottom. As expected, the weak current strength did not cause significant geometric distortions despite being applied throughout the readout periods. Figure 4b shows the measured phase evolution across echoes for $T_{\text{ES}} = [20-160] \text{ ms}$, confirming the linear phase evolution for the optimized crusher gradients. The simulated and measured dependencies of the efficiency of the combined $\Delta B_{z,c}$ on N_{echo} and T_{ES} are shown in Figure 4e,f. Because the simulations give only relative efficiency values, both plots are normalized to their individual maxima. The

FIG. 3. Phase evolution for MESE across echoes, tested for two different echo spacings $T_{\text{ES}} = [40, 60]$ ms. (a) The refocusing pulse tip angle is 150° , and the gradient areas and axes are kept identical across echoes. This results in both primary and stimulated echo pathways. (b) The refocusing pulse tip angle is 180° , and the gradients are kept identical. This also causes primary and stimulated echo pathways. (c) The refocusing pulse tip angle is 180° and the gradients are systematically varied, resulting in the selection of only the primary echo pathway and a linear phase accumulation.

simulations and experimental results are in good agreement. The corresponding results for the SNR are shown in Supporting Figure S1. While the SNR increases with the number of acquired echoes, the highest efficiency occurs for $N_{\text{echo}} = [2, 3]$. This indicates that the later echoes contribute only weakly to the combined $\Delta B_{z,c}$ image. Interestingly, the highest efficiency is found for rather long echo times of $T_{\text{ES}} = [80-100]$ ms. In order to make a comparison with our results and the literature, single-echo SE with $T_E = 60$ ms is selected as a reference (12,21). The selection of the most efficient sequence parameters results in an efficiency increase of 41%.

So far, the efficiency was only assessed for a single slice and a fixed T_D value. Figure 5a shows the simulated efficiency also with respect to T_D and different number of slices, normalized to the maximum across all simulations. For each T_D , the most efficient T_{ES} and N_{echo} were selected. In addition, the corresponding T_{ES} and N_{echo} for a single slice are shown in Figure 5b. For a single slice, the efficiency peaks for a rather long T_D of around 1.5 seconds, indicating that a substantial recovery of the longitudinal magnetization before re-excitation is optimal. Interestingly, the maximal efficiency can still be reached for three to four slices (as expected, T_D reaches 0 in this case) and a clear drop occurs only for a higher number of slices. This shows that multi-slice MESE MREIT is feasible without losing efficiency. The optimized T_{ES} is around 80 ms and the best-performing N_{echo} increases from 2 to 4 when increasing T_D .

SSFP-FID

Simulated dependencies of the transverse magnetization phase on $\Delta B_{z,c}$ are shown in Figure 6a for both sequence variants. The results indicate that the dependency of the steady-state phase on $\Delta B_{z,c}$ can be well linearized for weak injection currents. Judging from the slope of the phase dependencies around 0, SSFP-FID_{PCI} is 37% and SSFP-FID_{FCI} is 73% more sensitive compared to the standard spin echo case. Measured

dependencies of the steady-state phase on the injected current strength I_c are shown in Figure 6b (SSFP-FID_{PCI}) and 6c (SSFP-FID_{FCI}). The measured steady-state phase depends linearly on $\Delta B_{z,c}$ for both variants. SSFP-FID_{FCI} is 26% more sensitive than SSFP-FID_{PCI} for $T_R = 50$ ms. This is in good agreement with the simulations (Fig. 6a), using the linear relationship between $\Delta B_{z,c}$ and I_c . In contrast to the phase, the steady-state magnitude has a flat dependency on $\Delta B_{z,c}$ for both variants for the weak injected current strengths tested here ($\Delta B_{z,c}$ close to 0), both in the simulations (Supporting Fig. S2a) and measurements (Supporting Fig. S2b,c).

The results of the efficiency simulations and measurements for SSFP-FID_{FCI} are shown in Figure 7. Figure 7a shows the reconstructed $\Delta B_{z,c}$ image from the averaged phase difference images between the two alternating steady states for $\alpha = 20^\circ$, $T_R = 20$ ms, $T_E = 10$ ms, and $I_c = 1$ mA. The image is in agreement with the current flowing from top to bottom, and, as expected, the weak current strength did not cause geometric distortions despite being applied throughout the readout periods. However, significant signal drop attributed to T_2^* decay is observed in the poorly shimmed regions, such as near the electrodes and phantom edges. Figure 7b shows the measured phase evolution for $T_E = [10-140]$ ms and $T_R = [20-260]$ ms. The steady-state phase increases linearly with increasing T_E . There is no significant phase change observed for different T_R values when T_E is kept constant.

Figure 7c,d shows the simulation and experimental results for the efficiencies of $\Delta B_{z,c}$, normalized to their individual maxima (Supporting Fig. S3a,b depicts the corresponding SNR plots). Simulations and experimental results agree well. The maximal efficiency occurs for $T_E = [60-90]$ ms and $T_R = [120-180]$ ms. The highest efficiency is mostly observed when $T_E = T_R/2$, attributed to the symmetric data acquisition. Interestingly, the highest efficiency occurs for rather long echo times. This indicates that the increased signal strength attributed to

FIG. 4. MESE simulation and measurement results. (a) Measured combined $\Delta B_{z,c}$ image for $N_{\text{slice}} = 1$, $N_{\text{avg}} = 1$, $N_{\text{echo}} = 8$, $T_{\text{ES}} = 20$ ms, $T_D = 510$ ms, $BW = 125$ Hz/pixel, and $I_c = 0.5$ mA. The current is injected in a vertically downward direction. The ROI used to calculate the SNR and the efficiency is shown by the dashed lines. (b) Measured phase evolution over echo numbers for different T_{ES} . (c) Simulated efficiency. (d) Measured efficiency. The results in (c) and (d) are normalized relative to their maximal values. The measurement and simulation parameters in (b–d) are $\text{FOV} = 300 \times 300 \text{ mm}^2$, image matrix = 256×256 , $\Delta z = 5$ mm, $N_{\text{slice}} = 1$, $N_{\text{avg}} = 1$, $N_{\text{echo}} = [1-8]$, $T_{\text{ES}} = [20-160]$ ms, $T_D = 510$ ms, $T_1 = 1.1$ seconds, $T_2 = 100$ ms, $T_2^* = 50$ ms, and $I_c = 0.5$ mA. In both measurements and simulations, the lowest possible BW is selected to maximize the SNR of the MR magnitude image.

increased T_1 recovery and higher phase accumulation outweighs the stronger impact of T_2^* decay at longer T_E .

So far, the results were assessed for a fixed tip angle α of 20° . Figure 8 shows the simulated efficiency also with respect to changes in the tip angle for SSFP-FID_{FCL}, normalized to the maximum across all simulations. The

most efficient T_E and T_R values were selected for each tip angle. The maximal efficiency occurs around $\alpha = 30^\circ$ and decreases slightly for higher tip angles (Fig. 8a). The corresponding optimized T_E and T_R values are shown in Figure 8b,c. The optimized echo time T_E varies in the range between 50 and 80 ms (i.e., it is roughly in the

FIG. 5. (a) Efficiency of MESE with respect to T_D , assessed for slices $N_{\text{slice}} = [1-5, 15]$ and normalized to the peak across all simulations. For each T_D , T_{ES} , BW , and N_{echo} were optimized. (b) Corresponding echo spacing T_{ES} and number of echoes N_{echo} for $N_{\text{slice}} = 1$.

FIG. 6. (a) Simulated dependency of phase of the steady-state transverse magnetization. (b) SSFP-FID_{FCI}. Measured dependencies of the phase of the transverse magnetization. (c) SSFP-FID_{FCI}. Measured dependencies of the phase of the transverse magnetization. (b,c) The results were obtained for $T_R = [10, 30, 50]$ ms and $I_c = [-10 \text{ to } 10]$ mA.

range of the selected T_2^*) and reaches a plateau for higher tip angles. This is mainly attributed to the large signal loss for sampling times much longer than T_2^* . The optimized T_R increases with tip angle.

Comparison Between MESE and SSFP-FID

The efficiencies of MESE and SSFP-FID_{FCI} with optimized sequence parameters were directly compared. SSFP-FID_{FCI} has a 0.07% higher SNR for $\Delta B_{z,c}$ compared to MESE, but gives a 3 times higher efficiency. This suggests that SSFP-FID may be very useful for rapid imaging. However, the SSFP-FID causes a significant efficiency decrease in inhomogeneous regions and the image is significantly distorted, whereas MESE can preserve both. In addition, multi-slice SSFP-FID applications will cause significant efficiency decrease, whereas MESE preserves the efficiency.

Maximal Efficiency for Multi-Gradient-Echo Acquisition

The most efficient parameter ranges in both MESE and SSFP-FID experiments result in very low BWs, which cause geometric image distortions attributed to B_0 inhomogeneities. This effect can be prevented by acquiring multiple gradient echoes during each readout period at a higher BW, which are then added (17). Here, the efficiency decrease attributed to the time required for the additional prephaser gradients and gradient switching and corresponding BW were simulated. For both monopolar (Fig. 9a) and bipolar readout gradients (Fig. 9b), only a moderate loss of efficiency of less than 10% occurred for up to 16 gradient echoes. This indicates that the summation of multiple gradient echoes may be a suitable way for preventing geometric distortions caused by otherwise low BWs while maintaining acquisition efficiency.

Experiments in a Phantom With Inhomogeneous Geometry

MESE and SSFP-FID_{FCI} images were obtained in Phantom 2 containing a piece of chicken meat to assess the sequence performance for nonuniform structures. The sequence parameters were chosen in pilot trials to optimize efficiency as far as possible while maintaining image quality at an acceptable level. The results are reported for vertical and horizontal directions of current injection. For MESE, the combined MR magnitude image is shown in Figure 10a, and the combined $\Delta B_{z,c}$ images for horizontal and vertical current injection are depicted in Figure 10b,c. For SSFP-FID_{FCI}, the averaged MR magnitude image is shown in Figure 10d, and the $\Delta B_{z,c}$ images are given in Figure 10e,f. Both sequences allow accurate $\Delta B_{z,c}$ measurements for the saline regions of the phantom, despite using a low current magnitude of $I_c = 1$ mA. The impact of the chicken piece on the $\Delta B_{z,c}$ distribution is clearly visible in particular for the horizontal current injection. In MESE, the SNR of combined $\Delta B_{z,c}$ image is lower in the region of the chicken meat, which can be explained by the chosen T_{ES} (80 ms), which exceeds T_2 in this region (50 ms) and results in a low signal magnitude (Fig. 10a). This is less of an issue for SSFP-FID_{FCI}, where a short T_E (15 ms) was chosen.

DISCUSSION AND CONCLUSIONS

Successful in-vivo applications of MRCDI and MREIT will require that magnetic fields created by weak injection

FIG. 7. Simulations and measurement results for SSFP-FID_{FCI}. (a) Measured $\Delta B_{z,c}$ image for $I_c = 1$ mA, $\alpha = 20^\circ$, $T_R = 20$ ms, and $T_E = 10$ ms. The ROI used to calculate the SNR and the efficiency is shown by the dashed lines. (b) Measured phase evolution. (c) Simulated efficiency (normalized to the maximum) of the reconstructed $\Delta B_{z,c}$ image. (d) Measured efficiency (normalized to the maximum) of the reconstructed $\Delta B_{z,c}$ image. The measurement and simulation parameters in (b–d) are $N_{\text{avg}} = 2 \times 2$ (two separate averages for each steady state), $T_E = [10-140]$ ms, $T_R = [20-260]$ ms, $T_1 = 1.1$ seconds, $T_2 = 100$ ms, $T_2^* = 50$ ms, voxel size $= 1.5 \times 1.5 \times 3$ mm³, image matrix $= 256 \times 256$, and $I_c = 1$ mA. In both measurements and simulations, readout was symmetrical around T_E and the lowest possible BW is selected to maximize SNR of the MR magnitude image. For impossible combinations of T_E and T_R (i.e., $T_E > T_R$), the SNR and efficiency were set to 0.

currents of 1 to 2 mA are reliably measured in clinically relevant acquisition times. We therefore performed systematic sensitivity analyses to optimize the efficiency of two MREIT pulse sequences based on MESE and SSFP-FID, respectively, while assuming relaxation times similar

to human brain tissue at 3T. For both sequence types, the current injection was extended into the readout periods to maximize sensitivity. Considering the low targeted current strengths, we suggest that this is feasible without causing relevant image distortions so that correction

FIG. 8. Simulated efficiencies for different tip angles for SSFP-FID_{FCI}. (a) Normalized maximal efficiency dependence on the tip angle. (b) Corresponding optimal T_E values. (c) Corresponding optimal T_R values.

FIG. 9. Efficiency loss and corresponding BW increase in case of multi-echo acquisition: (a) monopolar readout gradient and (b) bipolar readout gradient.

strategies (28) are not needed (the distortions depend on the ratio between current-induced magnetic field and the readout gradient magnitude). In line with this, our simulations and measurements indicate that the steady-state magnitude response is only insignificantly affected by weak $\Delta B_{z,c}$, and there was no observable distortions in the magnitude images.

For MESE, the highest efficiencies were reached at echo spacings of $T_{ES} = [80-100]$ ms when using two to three echoes and a rather long dead time of $T_D = 1.5$ s. This is interesting, because it highlights the importance of allowing for sufficient T_1 recovery to boost signal intensity and by that also the SNR and efficiency of the $\Delta B_{z,c}$ images. It further opens up the possibility to use the dead time to acquire additional slices without decreasing efficiency. The parameters giving highest efficiency depend on the chosen RF pulse width, crusher gradients duration, and the efficiency of refocusing pulses. In particular, increasing the efficiency of the refocusing pulses above the 86% achieved in our phantom experiments may result in higher efficiencies with shorter T_{ES} and more echoes. This might be feasible for some human applications attributed to a better RF field homogeneity, for example, in the upper part of the brain. It is important to note that efficiency improvements by the combination of multiple echoes depend on a proper design of the crusher gradients to allow a linear phase

accumulation over echoes. The systematic arrangement of crusher gradients in this study (doubling up gradient area or changing direction) guarantees elimination of unwanted echoes, at a cost of large crusher widths. This may cause small signal loss attributed to diffusion weighting, eddy currents, or concomitant magnetic fields, which are not quantified in this study. Alternative methods, such as random crusher variation, do not guarantee the complete elimination of unwanted echoes.

Two different SSFP-FID variants were considered, with the current being injected until T_E (as originally investigated in a previous work (19)) and within the entire T_R period, respectively. Because the later variant exhibited increased phase sensitivity, it was considered further in the efficiency analyses. The maximal efficiency occurred for echo times of $T_E = [60-90]$ ms, repetition times of $T_R = [120-180]$ ms, and tip angles of $\alpha = 30^\circ$.

Our main focus was on determining optimal parameter ranges. For this, relative, rather than absolute, efficiency values were sufficient, as obtained in the simulations. However, we also directly compared the measured absolute efficiencies between optimized MESE and SSFP-FID sequences. The results demonstrate that SSFP-FID has 3 times higher efficiency compared to MESE. The SNR of the reconstructed $\Delta B_{z,c}$ images are in a similar range, but the total scan time is substantially shorter for SSFP-FID.

FIG. 10. Results for Phantom 2. (a) Combined magnitude image for MESE. (b) Combined $\Delta B_{z,c}$ images for MESE (horizontal current injection). (c) Combined $\Delta B_{z,c}$ images for MESE (vertical current injection). (d) Averaged magnitude image for SSFP-FID_{FCI}. (e) Reconstructed $\Delta B_{z,c}$ image for SSFP-FID_{FCI} (horizontal current injection). (f) Reconstructed $\Delta B_{z,c}$ image for SSFP-FID_{FCI} (vertical current injection). The parameters for MESE were FOV = $192 \times 192 \text{ mm}^2$, image matrix = 128×128 , $\Delta z = 3 \text{ mm}$, $N_{\text{echo}} = 3$, BW = 100 Hz/pixel, $N_{\text{avg}} = 1$, $T_{\text{ES}} = 80 \text{ ms}$, and $T_D = 510 \text{ ms}$. The parameters for SSFP-FID_{FCI} were FOV = $192 \times 192 \text{ mm}^2$, image matrix = 128×128 , $\Delta z = 3 \text{ mm}$, BW = 100 Hz/pixel, $N_{\text{avg}} = 16 \times 2$, $\alpha = 20^\circ$, $T_E = 15 \text{ ms}$, and $T_R = 30 \text{ ms}$. The injected current magnitude was $I_c = 1 \text{ mA}$.

On the other hand, MESE has a better image quality, is robust to B_0 inhomogeneities, and is better suited for multi-slice experiments. Attributed to the robustness to B_0 inhomogeneities, MESE might perform better than SSFP-FID in regions with very short T_2^* (Supporting Fig. S4e,f).

Our results show that the efficiency is maximized for rather long echo spacings (for MESE) and echo times (for SSFP-FID), respectively. This also implies low readout bandwidths to optimize the SNR, resulting in considerable image distortions attributed to B_0 inhomogeneities. We suggest that this problem can be ameliorated without substantial decrease in efficiency when multiple gradient echoes are acquired at a higher BW during each readout period and are subsequently added (17). This strategy should result in a good image quality for MESE, for which the signal evolution is robust to B_0 inhomogeneities. SSFP-FID sequences are generally more susceptible to local B_0 inhomogeneities, so that the T_R (and thus also T_E), which can be achieved in practice, might be lower than the one required to maximize efficiency.

To summarize, in our phantom study, the optimized MESE and SSFP-FID sequences allowed for a reliable measurement of the magnetic field created by currents of 1 mA or below. This is promising for the exploration of these sequences for in-vivo brain imaging applications. Future sequence optimizations might use multi-gradient-echo readouts to combine high efficiencies with good image quality. Also, other sequences might further improve the efficiency, for example, balanced SSFP MREIT attributed to its very high phase sensitivity (29). Further studies are needed to evaluate the image quality in-vivo, which also depends on the sensitivity of the

sequence, for example, to physiological noise and subject motion.

REFERENCES

1. Mikac U, Demsar F, Beravs K, Sersa I. Magnetic resonance imaging of alternating electric currents. *Magn Reson Imaging* 2001;19:845–856.
2. Eyüboğlu BM. Magnetic resonance current density imaging. In *Wiley Encyclopedia of Biomedical Engineering*, Vol. 4, Metin Akay, ed. Hoboken, NJ, USA: Wiley-Interscience; 2006:2147–2153.
3. Eyüboğlu BM. Magnetic resonance electrical impedance tomography. In *Wiley Encyclopedia of Biomedical Engineering*, Vol. 4, Metin Akay, ed. Hoboken, NJ, USA: Wiley-Interscience; 2006:2154–2162.
4. Seo JK, Woo EJ. Magnetic resonance electrical impedance tomography. *Soc Ind Appl Math* 2011;53:40–68.
5. Scott GC, Joy ML, Armstrong RL, Henkelman RM. Measurement of nonuniform current density by magnetic resonance. *IEEE Trans Med Imaging* 1991;10:362–374.
6. Woo EJ, Lee SY, Mun CW. Impedance tomography using internal current density distribution measured by nuclear magnetic resonance. *SPIE* 1994;2299:377–385.
7. Seo JK, Woo EJ. Electrical tissue property imaging at low frequency using MREIT. *IEEE Trans Biomed Eng* 2014;61:1390–1399.
8. Fear EC, Hagness SC, Meaney PM, Okoniewski M, Stuchly MA. Enhancing breast tumor detection with near-field imaging. *IEEE Microw Mag* 2002;3:48–56.
9. Mosher JC, Leahy RM, Lewis PS. EEG and MEG: forward solutions for inverse methods. *IEEE Trans Biomed Eng* 1999;46:245–259.
10. Opitz A, Paulus W, Will S, Antunes A, Thielscher A. Determinants of the electric field during transcranial direct current stimulation. *Neuroimage* 2015;109:140–150.
11. Thielscher A, Opitz A, Windhoff M. Impact of the gyral geometry on the electric field induced by transcranial magnetic stimulation. *Neuroimage* 2011;54:234–243.
12. Göksu C, Sadighi M, Eroğlu HH, Eyüboğlu M. Realization of magnetic resonance current density imaging at 3 Tesla. In: *Annual International Conference of the IEEE Engineering in Medicine and Biology Society, EMBC*, 2014. pp. 1115–1118.
13. Utz KS, Dimova V, Oppenländer K, Kerkhoff G. Electrified minds: transcranial direct current stimulation (tDCS) and Galvanic Vestibular

- Stimulation (GVS) as methods of non-invasive brain stimulation in neuropsychology—a review of current data and future implications. *Neuropsychologia* 2010;48:2789–2810.
14. Minhas AS, Jeong WC, Kim YT, Han Y, Kim HJ, Woo EJ. Experimental performance evaluation of multi-echo ICNE pulse sequence in magnetic resonance electrical impedance tomography. *Magn Reson Med* 2011;66:957–965.
 15. Sadleir R, Grant S, Zhang SU, Oh SH, Lee B Il, Woo EJ. High field MREIT: setup and tissue phantom imaging at 11 T. *Physiol Meas* 2006;27:S261–S270.
 16. Hamamura MJ, Muftuler LT. Fast imaging for magnetic resonance electrical impedance tomography. *Magn Reson Imaging* 2008;26:739–745.
 17. Nam HS, Kwon OI. Optimization of multiply acquired magnetic flux density $B(z)$ using ICNE-Multiecho train in MREIT. *Phys Med Biol* 2010;55:2743–2759.
 18. Birgül Ö, Eyüboğlu BM, Ider YZ. Magnetic resonance-conductivity imaging using 0.15 Tesla MRI scanner. In: 2001 Proceedings of the 23rd Annual EMBS International Conference, 2001. pp. 3384–3387.
 19. Lee H, Jeong WC, Kim HJ, Woo EJ, Park J. Alternating steady state free precession for estimation of current-induced magnetic flux density: a feasibility study. *Magn Reson Med* 2016;75:2009–2019.
 20. Pohmann R, von Kienlin M, Haase A. Theoretical evaluation and comparison of fast chemical shift imaging methods. *J Magn Reson* 1997;129:145–160.
 21. Scott GC, Joy MLG, Armstrong RL, Henkelman RM. Sensitivity of magnetic-resonance current-density imaging. *J Magn Reson* 1992;97:235–254.
 22. Sadleir R, Grant S, Zhang SU, et al. Noise analysis in magnetic resonance electrical impedance tomography at 3 and 11 T field strengths. *Physiol Meas* 2005;26:875–884.
 23. Scheffler K, Maderwald S, Ladd ME, Bieri O. Oscillating steady states. *Magn Reson Med* 2006;55:598–603.
 24. Wansapura JP, Holland SK, Dunn RS, Ball WS. NMR relaxation times in the human brain at 3.0 Tesla. *J Magn Reson Imaging* 1999;9:531–538.
 25. Jaynes ET. Matrix treatment of nuclear induction. *Phys Rev* 1955;98:1099–1105.
 26. Walsh DO, Gmitro AF, Marcellin MW. Adaptive reconstruction of phased array MR imagery. *Magn. Reson. Med.* 2000;43:682–690.
 27. Bernstein MA, King KF, Zhou XJ (eds). *Handbook of MRI Pulse Sequences*. Oxford, UK: Elsevier Academic; 2004.
 28. Park C, Lee B Il, Kwon O, Woo EJ. Measurement of induced magnetic flux density using injection current nonlinear encoding (ICNE) in MREIT. *Physiol Meas* 2007;28:117–127.
 29. Bieri O, Maderwald S, Ladd ME, Scheffler K. Balanced alternating steady-state elastography. *Magn Reson Med* 2006;55:233–241.
 30. Scheffler K. A pictorial description of steady-states in rapid magnetic resonance imaging. *Concepts Magn Reson* 1999;11:291–304.
 31. Tofts PS. QA: quality assurance, accuracy, precision and phantoms. *Quant MRI Brain* 2003:55–81.
 32. Insko E, Bolinger L. Mapping of the radiofrequency field. *J Magn Reson* 1993;103:82–85.

SUPPORTING INFORMATION

Additional supporting information can be found in the online version of this article.

Fig. S1. MESE results. (a) Simulated and (b) measured dependence of $SNR_{ABZ,C}$ on the acquired number of echoes and on T_{ES} . Results are normalized relative to their maximal values. The measurement and simulation parameters are $FOV = 300 \times 300 \text{ mm}^2$, image matrix = 256×256 , $\Delta z = 5 \text{ mm}$, $N_{\text{slice}} = 1$, $N_{\text{avg}} = 1$, $N_{\text{echo}} = [1-8]$, $T_{ES} = [20-160] \text{ ms}$, $T_D = 510 \text{ ms}$, $T_1 = 1.1 \text{ s}$, $T_2 = 100 \text{ ms}$, $T_2^* = 50 \text{ ms}$, and $I_c = 0.5 \text{ mA}$. In both measurements and simulations, the lowest possible BW is selected to maximize the SNR of the MR magnitude image.

Fig. S2. (a) Simulated dependency of the magnitude of the steady-state transverse magnetization on $\Delta B_{z,C}$ for SSFP-FID. (b) Measured dependencies of the magnitude of the transverse magnetization on the injected current strength for SSFP-FID_{FCI}. As a side note, a decrease in the signal magnitude when increasing T_E is usually expected for SSFP-FID sequences attributed to T_2 decay. However, this is only the case when holding the BW fixed. Here, the experiments were performed for the lowest possible BW (with T_E adjusted to $T_R/2$), which caused increases in the signal magnitude up to $T_E = [60-80] \text{ ms}$. (c) Measured dependencies of the magnitude of the transverse magnetization for SSFP-FID_{FCI}. The distortion in the flat response at $T_R = 10 \text{ ms}$ and $I_c = 6 \text{ mA}$ may have been caused by hardware imperfection. The results were obtained for $T_R = [10, 30, 50] \text{ ms}$ and $I_c = [-10 \text{ to } 10] \text{ mA}$.

Fig. S3. SSFP-FID_{FCI} results. (a) Simulated and (b) measured dependence of $SNR_{ABZ,C}$ on T_E and T_R . The measurement and simulation parameters are $N_{\text{avg}} = 2 \times 2$ (two separate averages for each steady state), $T_E = [10-140] \text{ ms}$, $T_R = [20-260] \text{ ms}$, $T_1 = 1.1 \text{ seconds}$, $T_2 = 100 \text{ ms}$, $T_2^* = 50 \text{ ms}$, voxel size = $1.5 \times 1.5 \times 3 \text{ mm}^3$, image matrix = 256×256 , and $I_c = 1 \text{ mA}$. In both measurements and simulations, the readout is symmetrical around T_E and the lowest possible BW is selected to maximize SNR of the MR magnitude image. For impossible combinations of T_E and T_R (i.e., $T_E > T_R$), the SNR values were set to 0.

Fig. S4. Dependence of the maximal efficiency of MESE (a,c,e) and SSFP-FID_{FCI} (b,d,f) on the relaxation parameters T_1 , T_2 , and T_2^* . The simulations were performed by varying one of the relaxation parameters while keeping the other two fixed and close to the parameters of brain tissue ($T_1 = 1.1 \text{ seconds}$, $T_2 = 100 \text{ ms}$, and $T_2^* = 50 \text{ ms}$). The simulations are normalized to their maxima.