An Error Analysis of Structured Light Scanning of Biological Tissue

Jensen, Sebastian Hoppe Nesgaard; Wilm, Jakob; Aanæs, Henrik

Published in:
Image Analysis

Link to article, DOI:
[10.1007/978-3-319-59126-1_12](https://doi.org/10.1007/978-3-319-59126-1_12)

Publication date:
2017

Document Version
Peer reviewed version

[Link back to DTU Orbit](#)

Citation (APA):
Jensen, S. H. N., Wilm, J., & Aanæs, H. (2017). An Error Analysis of Structured Light Scanning of Biological Tissue. In *Image Analysis* (pp. 135-145). Springer. Lecture Notes in Computer Science Vol. 10269 https://doi.org/10.1007/978-3-319-59126-1_12

General rights

Copyright and moral rights for the publications made accessible in the public portal are retained by the authors and/or other copyright owners and it is a condition of accessing publications that users recognise and abide by the legal requirements associated with these rights.

- Users may download and print one copy of any publication from the public portal for the purpose of private study or research.
- You may not further distribute the material or use it for any profit-making activity or commercial gain
- You may freely distribute the URL identifying the publication in the public portal

If you believe that this document breaches copyright please contact us providing details, and we will remove access to the work immediately and investigate your claim.

An Error Analysis of Structured Light Scanning of Biological Tissue

Sebastian Nesgaard Jensen, Jakob Wilm and Henrik Aanaes
{snje, jakw, aanes}@dtu.dk

Technical University of Denmark
Department of Applied Mathematics and Computer Science
Richard Petersens Plads B321, Kongens Lyngby, Denmark

Abstract. This paper presents an error analysis and correction model for four structured light methods applied to three common types of biological tissue; skin, fat and muscle.

Despite its many advantages, structured light is based on the assumption of direct reflection at the object surface only. This assumption is violated by most biological material e.g. human skin, which exhibits subsurface light reflection. In this study, we find that in general, structured light scans of biological tissue deviate significantly from the ground truth. We show that a large portion of this error can be predicted with a simple, stochastic linear model based on the scan geometry. As such, scans can be corrected without introducing any specially designed pattern strategy or hardware. We can effectively reduce the error in a structured light scanner applied to biological tissue by as much as factor of two or three.

Keywords: 3D Reconstruction, Error Modeling, Structured Light

1 Introduction

Structured light has proven to be very useful for 3D scene acquisition. This is due to its high speed, precision and versatility. As such a wide array of related techniques have been developed in the past decades, facilitating everything from high precision metrology to real-time guidance of automation [8].

Structured light uses a calibrated camera-projector pair as shown in Fig. 1. A series of time multiplexed patterns is projected onto the scene, which can be used for matching and triangulation with the camera. This active approach makes correspondence searching much simpler than passive stereo approaches, and is applicable to scenes with poor texturing. A very important application for structured light is 3D scanning of biological materials, especially human tissue. Examples include head tracking for medical motion correction [22], vision guided surgery [18][23], medical diagnostics [4][1][28] and automation in agriculture and farming [21][25][7]. While the progress in the field has been impressive, one must understand that many target materials are quite problematic. Indeed, they violate the inherent assumption of direct, diffuse surface reflection that most structured light methods are built on. The Fresnel equations predict that when light transitions from one media to another a portion is directly reflected and another is transmitted into the media itself. In the media the light is scattered one or multiple times

until it is absorbed or retransmitted into the environment. The proportion between reflected and refracted light is determined by the specific media's optical properties. For example only 5-7% of human skin reflectance is direct, the remainder is emitted via subsurface scattering [14]. It is therefore of paramount importance that the effect of this violation on structured light is understood and quantified.

In this study, we show that in general, a structured light scan of biological tissue deviates significantly from reference measurements, even with patterns designed specifically to reduce these effects. A large portion of the error can be predicted with a simple, stochastic linear model based on the incident ray geometry. Scans can then be corrected without the need for advanced pattern strategies or special hardware. We can effectively reduce the error in any structured light scanner applied to biological tissue by as much as factor of two or three.

Our study focuses on three types of biological tissue (fat, muscle and skin) with an emphasis on human applications. However we are using porcine materials as a substitute due to its availability and optical similarity to human tissue [26][27]. Through empirical study we quantify the error induced in structured light by the biological material's optical properties. This results in a linear error model based on the view geometry fitted to each method, material combination that can be used to predict and correct for the deterministic scan error.

Fig. 1. The structured light principle: a number of patterns are projected onto the scene, and images are captured by a camera. Correspondences are determined by different encoding algorithms, and used to triangulate points on the object surface. In this example, 3-step phase shifting patterns are shown.

2 Related Work

The issue of global lighting effects in the context of structured light has been recognized by many authors, e.g. in the acquisition of a human face reflectance field [6].

In order to reduce these effects, hardware modifications such as polarization have been used [2]. Recent attempts have been to design structured light encoding patterns such that they are less susceptible to global lighting effects. The underlying observation is, that with high-frequent patterns, global lighting effects can be considered constant,

and invariant to a spatial shift of the pattern. This allows for efficient separation of the observed light intensities into direct and global light [20]. In modulated phase shifting [3], structured light patterns are modulated by means of carrier patterns, such that they become high-frequent in both spatial dimensions, thereby improving their separation power. Micro Phase Shifting [10] makes use of sinusoidal patterns in a narrow high-frequency band, promising robustness to global lighting effects and stable phase unwrapping with an optimal number of patterns. It should be noted, that the decoding process in conventional phase shifting methods (e.g. [13]) also implicitly performs direct/global light separation. This is true in particular for high frequency scene coding patterns. Since lower frequency phase unwrapping patterns are affected differently by global lighting effects, this can lead to gross outliers. Hence, the advantage of micro phase shifting is not in higher accuracy, but rather in improved robustness (fewer outliers), and more efficient use of information in the encoding patterns.

A newer approach is unstructured light [5], in which the pattern frequency can be high in both dimensions. However the number of patterns is not ideal, and the matching procedure rather inefficient. For binary encoding methods, exclusively high or low-frequency pattern schemes can be considered robust against different global illumination effects [9].

An approach to compensate for the measurement error in isotropic semi-transparent material caused by subsurface scattering was presented in [16]. Similarly to our approach, this work empirically determines the measurement error and explains it by means of a single variable (the projected light angle), albeit only with a single verification object and structured light method. In [15], a Monte-Carlo simulation of the measurement situation was presented, which gives some insight into the error forming process.

In [11], an analytical derivation of the measurement error is given for the phase shifting method. This error model predicts the error to decrease with increased spatial frequency of the pattern. The model does not however take into account the loss of amplitude at higher frequency patterns, which increases noise in the measurement data. Furthermore it requires precise knowledge about the scanned material's optical properties (extinction coefficient, phase function and index of refraction), all of which can be difficult to find or estimate.

Computer simulations of structured light scans were performed in [19] to benchmark encoding methods with respect to various parameters, and were found to have similar robustness with respect to subsurface effects.

To our knowledge, no study has thus far quantified the amount of error in scans of biological tissue, or provided a means of correcting for it.

3 Statistical Error Model

Our principle assumption is that the error is composed of a deterministic part, which once determined can be subtracted from future scans, in order to improve the accuracy. Previous work gives some hints as to which parameters to include in a statistical error model [16][11].

Fig. 2. The structured light scan geometry with the parameters of our error model. The surface normal is n , view direction v , light direction l and the projector-surface distance is d .

Considering the scan geometry, as shown in Fig. 2, we include three variables in our error model: the view angle (given by $n \cdot v$), the light angle (given by $n \cdot l$) and the distance from projector to object, d . We then formulate the following error model:

$$y = [1 \ n \cdot v \ n \cdot l \ d] \begin{bmatrix} \beta_0 \\ \beta_1 \\ \beta_2 \\ \beta_3 \end{bmatrix}, \quad (1)$$

where

y is the predicted error in mm,

β_i is a weight,

n , v , l and d are shown in Fig. 2.

We also tried including many other variables, including reflected light to view angle and coding direction to normal vector angles. These variables are inspired by the analytical error model of Holroyd [11], but did not explain sufficient variance to justify their inclusion in our model. We also fitted Holroyd's error model directly, but our linear model provided more explanatory power.

4 Experiments

In order to gather data for the error quantification we scanned surfaces made of one of three porcine tissue types; fat, muscle or skin. All samples were raw and unprocessed with 8 samples of each type. The samples were placed individually in the scan volume and spanned many view and light angles. Their distance to the projector also varied from approximately 200 mm to 400 mm. Each scan produced approximately $5 \cdot 10^5$ data points resulting in millions for each tissue type.

In optical metrology it is common practice to prepare optically challenging surface with a spray [12]. This makes the surface optically diffuse while preserving the geometry. The method was used to acquire a ground truth surface to which each scan was compared. Specifically, after each scan the object was sprayed and covered with a thin

Fig. 3. Structured light patterns used in our experiments. In each case, 12 patterns were used.

Fig. 4. Fine grained binary structured light pattern projected onto various types of tissues. The effect of subsurface scattering is clearly seen the pattern becomes blurred without chalk coating.

layer of chalk. Then the reference scan was obtained. While we cannot assume that the chalk coated surfaces to be perfect, we consider them ground truth as they provide very clear contrast with virtually no global illumination. In order to verify that this procedure does not alter surface geometry, we applied two separate layers of chalk to a sample object, and compared the scan result after each layer. The mean signed distance was 0.037 mm, indicating that chalk spraying the surfaces does not significantly bias the result. As can be seen in Figure 4 the effect of chalk spraying is relatively pronounced, increasing reflectance and counteracting the pattern blurring caused by subsurface scattering.

In our experiments, we used four different structured light methods:

- Binary Gray coding [24]: one completely lit and one completely dark image were used to define the binary threshold individually in each camera pixel. The remaining patterns were used to encode $2^{10} = 1024$ individual lines on the object surface.

- N-step phase shifting was used with 9 shifts of a high-frequency sinusoid of frequency $1/76 \text{ px}^{-1}$, corresponding to approximately $1/10 \text{ mm}$ on the object surface. Three additional patterns were used for phase-unwrapping [13].
- Micro phase shifting [10] using frequencies in the band $[1/80.00 - 1/70.00] \text{ px}^{-1}$. These frequencies corresponds to a spatial frequency on the object surface of approximately $1/10 \text{ mm}$. Slightly different from [10], the specific values were determined using a derivative free non-linear pattern search.
- Modulated phase shifting [3] with three shifts of a sinusoid of frequency $1/76 \text{ px}^{-1}$ ($1/10 \text{ mm}$ on the object surface). Each of these sinusoids was modulated in the orthogonal direction using a sinusoidal carrier with the same frequency. Three additional patterns were used for phase-unwrapping.

For the sake of brevity these will henceforth be referred respectively to as; Gray, PS, Micro PS and Mod PS. The former two are standard methods of structured light and can be expected to perform very similar to many derived methods. The latter two are state-of-the-art and have been specifically designed to mitigate the effects of global illumination, as described in Sec. 2. A pattern budget¹ of 12 was settled on for each method as it provided a reasonable balance in acquisition time and accuracy. For all phase-shifting methods, pattern frequency was set so that each period would be approximately 10 mm on the object surface. The remaining frequencies needed in micro phase-shifting were determined using simplex optimization as suggested in the original paper [10]. Fig. 3 shows the pattern sequences used in our experiments.

For every sample, we defined a binary mask within which all possible surface points were reconstructed. This ensured that the exact same surface region was used in the evaluation of each method.

The error of each surface point was quantified by determining its signed distance to the corresponding point in the chalk sprayed reference. For Gray code scans we define the corresponding points as being the pair with the smallest absolute normal distance. With the other methods, we compared points using their position in the pixel grid.

5 Results and Discussion

The parameters obtained after fitting the error model to our data are seen in Table 1, 2 and 3. These shows the estimated parameters as well as the RMS of data compared to the chalk coated reference before and after correction (respectively RMS_{raw} and RMS_{cor}) in units of mm . The two latter were evaluated through a process of leave-one-out k-fold cross validation with 5 partitions. In addition we have also estimated the degree of variance explained, R^2 , as well as the P -values for the statistical significance of our model against a constant model. All model dependencies were subject to an analysis of variance (ANOVA) [17].

In general the model provides a significant reduction in RMS for all methods with the greatest effect for muscle and skin. It is interesting to note that R^2 is in general relatively low; at best 13% and at worst 0.8%. Such measure might dispute model's validity, but the statistical test versus a constant model proves otherwise. In all cases we

¹ Pattern budget is the number of projected patterns allowed in a single scan.

Fig. 5. Signed distance (sd) between scan and reference on a single sample of muscle. Top row: before applying the linear correction model. Bottom row: after correction.

Fig. 6. Signed distance (sd) between scan and reference on a single sample of skin. Top row: before applying the linear correction model. Bottom row: after correction.

Fig. 7. Signed distance (sd) between scan and reference on a single sample of fat. Top row: before applying the linear correction model. Bottom row: after correction.

Table 1. Muscle model estimate and regression quality.

	β_0	β_1	β_2	β_3	RMS_{raw}	RMS_{cor}	R^2	P
Gray	0.13	0.15	-0.026	2.3×10^{-4}	0.42	0.27	0.0082	0
Phase Shifting	0.25	0.47	-0.18	-2.5×10^{-5}	0.5	0.21	0.06	0
Micro PS	0.21	0.36	-0.12	-4.1×10^{-6}	0.45	0.23	0.034	0
Modulated PS	0.27	0.077	0.053	-9.7×10^{-5}	0.42	0.26	0.0037	0

Table 2. Skin model estimate and regression quality.

	β_0	β_1	β_2	β_3	RMS_{raw}	RMS_{cor}	R^2	P
Gray	-0.48	0.018	0.43	1.3×10^{-3}	0.4	0.19	0.069	0
Phase Shifting	0.27	0.28	0.26	-5.9×10^{-4}	0.54	0.17	0.13	0
Micro PS	0.45	0.27	0.21	-1.0×10^{-3}	0.52	0.19	0.13	0
Modulated PS	0.34	0.1	0.27	-6.7×10^{-4}	0.46	0.22	0.054	0

Table 3. Fat model estimate and regression quality.

	β_0	β_1	β_2	β_3	RMS_{raw}	RMS_{cor}	R^2	P
Gray	-0.12	0.13	0.039	2.0×10^{-4}	0.26	0.24	0.016	0
Phase Shifting	-0.18	0.31	-0.11	3.9×10^{-4}	0.22	0.16	0.084	0
Micro PS	-0.13	0.2	-0.043	3.0×10^{-4}	0.2	0.16	0.043	0
Modulated PS	-0.06	0.15	-0.029	1.6×10^{-4}	0.2	0.17	0.018	0

can conclude that our model is statistical significant within almost a 100% confidence interval, as indicated by the P -values tested against a constant model. While this might seem improbably low, bear in mind that the models was estimated using millions of points which assists in obtaining a statistical significant results. The model estimate itself is rather stable, yielding almost the same error measure for every iteration in the cross validation. This is to be expected due to the high number of training samples and the low dimensionality of the model.

It is seen that most methods have a positive intercept, meaning that regardless of measurement conditions the surface seems to be further away from the camera. The phase shifting methods are especially affected by this bias. This effect is further amplified under ideal scanning conditions, where view and light angle are approximately perpendicular to the measured surface. Since β_1 and β_2 are in most cases positive it will further add to positive surface bias. It is also interesting to note that for phase shifting methods distance adds a negative weight. This means that distant measurement will effectively have less of a positive bias than close ones. The worst bias can be observed in standard phase-shifting applied to skin were error can climb to approximately 0.75 mm.

This positive trend can be illustrated by visualizing the per point error as a heat map upon an obtained point cloud, an interesting trend can be observed. Fig. 5, 6 and 7 shows the signed error on a single sample visually before and after applying the correction model. All have a positive bias which is very strong for muscle and skin. This alludes to a general trend, subsurface scattering causes the estimated surface to lie further away

from the scanner. This is intuitively correct as subsurface scattering is caused by light entering the material for a bit before it is reflected.

In all cases the application of the proper linear model reduces the error's RMS significantly. With a relatively low reduction for fat and a high reduction for skin and muscle. Skin seems to be especially interesting for application as it has the highest error RMS and also receives the largest reduction from error prediction. The remaining unmodeled variance can probably be attributed to variance in chalk thickness, material inhomogeneity and slight vibrations in the recording environment.

6 Conclusion

Structured light is greatly affected by the optical properties of biological materials such as subsurface scattering. By comparing structured light scans of a biological object with scans of the same object covered with a thin chalk layer, we have successfully quantified the resulting error. Our study shows a general positive bias resulting in a surface that lies further away from the scanner than an identical diffuse surface. Due to this positive bias, the RMS of the error can be as high as 0.54 mm. We described the error by fitting a stochastic linear model based on view geometry to the obtained data. Using it, a large portion of the error can be predicted and compensated for. For instance, applying this model to phase-shifting scans of skin reduces error RMS from 0.54 mm to 0.17 mm.

As opposed to the solutions to global illumination proposed in [10][3] our approach requires no specially designed pattern strategy or hardware. It can simply be applied directly to the obtained geometry. Additionally our methodology can be applied to any given structured light method and subsurface scattering material. From a pragmatic view, one must conclude that standard phase-shifting is the superior choice for scanning biological tissue. Not because it shows the lowest error, but rather because the error can be predicted well and compensated using our method.

Bibliography

- [1] ACKERMAN, J. D., KELLER, K., AND FUCHS, H. Surface reconstruction of abdominal organs using laparoscopic structured light for augmented reality. *Proc. SPIE 4661* (2002), 39–46.
- [2] CHEN, T., LENSCH, H. P. A., FUCHS, C., AND SEIDEL, H. P. Polarization and phase-shifting for 3D scanning of translucent objects. *Proc. IEEE CVPR* (2007).
- [3] CHEN, T., SEIDEL, H.-P., AND LENSCH, H. P. Modulated phase-shifting for 3D scanning. *Proc. IEEE CVPR* (2008), 1–8.
- [4] CLANCY, N. T., LIN, J., ARYA, S., HANNA, G. B., AND ELSON, D. S. Dual multispectral and 3d structured light laparoscope. *Proc. SPIE 9316* (2015), 93160C.
- [5] COUTURE, V., MARTIN, N., AND ROY, S. Unstructured light scanning robust to indirect illumination and depth discontinuities. *Int. Journal on Computer Vision* 108, 3 (2014), 204–221.
- [6] DEBEVEC, P., HAWKINS, T., TCHOU, C., DUIKER, H.-P., SAROKIN, W., AND SAGAR, M. Acquiring the reflectance field of a human face. *Proc. SIGGRAPH* (2000), 145–156.
- [7] FENG, Q. C., CHENG, W., ZHOU, J. J., AND WANG, X. Design of structured-light vision system for tomato harvesting robot. *Int. Journal of Agricultural and Biological Engineering* 7, 2 (2014), 19–26.
- [8] GENG, J. Structured-light 3D surface imaging: a tutorial. *Advances in Optics and Photonics* 160, 2 (2011), 128–160.
- [9] GUPTA, M., AGRAWAL, A., VEERARAGHAVAN, A., AND NARASIMHAN, S. G. A Practical Approach to 3D Scanning in the Presence of Interreflections, Sub-surface Scattering and Defocus. *Int. Journal on Computer Vision* 102, 1-3 (aug 2012), 33–55.
- [10] GUPTA, M., AND NAYAR, S. K. Micro Phase Shifting. *Proc. IEEE CVPR* (2012), 813–820.
- [11] HOLROYD, M., AND LAWRENCE, J. An Analysis of Using High-Frequency Sinusoidal Illumination to Measure the 3D Shape of Translucent Objects. *Proc. IEEE CVPR* (2011), 2985–2991.
- [12] HUANG, Z., NI, J., AND SHIH, A. J. Quantitative evaluation of powder spray effects on stereovision measurements. *Measurement Science and Technology* 19, 2 (2008).
- [13] HUNTLEY, J. M., AND SALDNER, H. Temporal phase-unwrapping algorithm for automated interferogram analysis. *Applied Optics* 32, 17 (1993), 3047–3052.
- [14] KRISHNASWAMY, A., AND BARANOSKI, G. A biophysically-based spectral model of light interaction with human skin. *Computer Graphics Forum* 23, 3 (2004), 331–340.
- [15] LUTZKE, P., HEIST, S., KÜHMSTEDT, P., KOWARSCHIK, R., AND NOTNI, G. Monte Carlo simulation of three-dimensional measurements of translucent objects. *Optical Engineering* 54, 8 (2015).

- [16] LUTZKE, P., KÜHMSTEDT, P., AND NOTNI, G. Measuring error compensation on three-dimensional scans of translucent objects. *Optical Engineering* 50, 6 (2011), 063601.
- [17] MADSEN, H., AND THYREGOD, P. *Introduction to general and generalized linear models*. CRC Press, Taylor & Francis Group, 2011.
- [18] MAURICE, X., ALBITAR, C., DOIGNON, C., AND DE MATHELIN, M. A structured light-based laparoscope with real-time organs' surface reconstruction for minimally invasive surgery. *Proc. Int. Conf. of IEEE EMBS 2012* (2012), 5769–5772.
- [19] MEDEIROS, E., DORAISWAMY, H., BERGER, M., AND SILVA, C. T. Using Physically Based Rendering to Benchmark Structured Light Scanners. *Pacific Graphics* 33, 7 (2014).
- [20] NAYAR, S. K., KRISHNAN, G., GROSSBERG, M. D., AND RASKAR, R. Fast separation of direct and global components of a scene using high frequency illumination. *ACM Trans. on Graphics* 25, 3 (2006), 935.
- [21] NGUYEN, T., SLAUGHTER, D. C., MAX, N., MALOOF, J. N., AND SINHA, N. Structured light-based 3d reconstruction system for plants. *Sensors* 15, 8 (2015), 18587–18612.
- [22] OLESEN, O. V., PAULSEN, R. R., HØJGAARD, L., ROED, B., AND LARSEN, R. Motion tracking for medical imaging: a nonvisible structured light tracking approach. *IEEE Trans. Medical Imaging* 31, 1 (2012), 79–87.
- [23] PAQUIT, V., PRICE, J. R., SEULIN, R., MERIAUDEAU, F., FARAH, R. H., TOBIN, K. W., AND FERRELL, T. L. Near-infrared imaging and structured light ranging for automatic catheter insertion. *Proc. SPIE 6141* (2006).
- [24] POSDAMER, J., AND ALTSCHULER, M. Surface measurement by space-encoded projected beam systems. *Computer Graphics and Image Processing* 18 (1982), 1–17.
- [25] ROSELL-POLO, J. R., CHEEIN, F. A., GREGORIO, E., ANDJAR, D., PUIGDOMNECH, L., MASIP, J., AND ESCOL, A. Chapter three - advances in structured light sensors applications in precision agriculture and livestock farming. In *Advances in Agronomy*, vol. 133. Academic Press, 2015, pp. 71 – 112.
- [26] TFAILLI, S., GOBINET, C., JOSSE, G., ANGIBOUST, J.-F., MANFAIT, M., AND PIOT, O. Confocal raman microspectroscopy for skin characterization: a comparative study between human skin and pig skin. *Analyst* 137, 16 (2012), 3673–3682.
- [27] WEIGMANN, H.-J., SCHANZER, S., PATZELT, A., BAHABAN, V., DURAT, F., STERRY, W., AND LADEMANN, J. Comparison of human and porcine skin for characterization of sunscreens. *Journal of Biomedical Optics* 14, 2 (2009), Article No.: 024027.
- [28] WISSEL, T., STÜBER, P., WAGNER, B., BRUDER, R., SCHWEIKARD, A., AND ERNST, F. Enriching 3d optical surface scans with prior knowledge: tissue thickness computation by exploiting local neighborhoods. *Int. Journal of Computer Assisted Radiology and Surgery* (2015).