



Towards Predicting Room Acoustical Effects on Sound-Field ASSR from Stimulus Modulation Power

Zapata Rodriguez, Valentina; Laugesen, Søren; Jeong, Cheol-Ho; Brunskog, Jonas; Harte, James Michael

Publication date:
2018

Document Version
Publisher's PDF, also known as Version of record

[Link back to DTU Orbit](#)

Citation (APA):
Zapata Rodriguez, V., Laugesen, S., Jeong, C-H., Brunskog, J., & Harte, J. M. (2018). *Towards Predicting Room Acoustical Effects on Sound-Field ASSR from Stimulus Modulation Power*. Abstract from 41st Midwinter Meeting of the Association for Research in Otolaryngology , San Diego, California, United States.

General rights

Copyright and moral rights for the publications made accessible in the public portal are retained by the authors and/or other copyright owners and it is a condition of accessing publications that users recognise and abide by the legal requirements associated with these rights.

- Users may download and print one copy of any publication from the public portal for the purpose of private study or research.
- You may not further distribute the material or use it for any profit-making activity or commercial gain
- You may freely distribute the URL identifying the publication in the public portal

If you believe that this document breaches copyright please contact us providing details, and we will remove access to the work immediately and investigate your claim.

Cheol-Ho Jeong

From: meetings@aro.org
Sent: 23. september 2017 23:04
To: valr@iru.interacoustics.com
Cc: slau@iru.interacoustics.com; Cheol-Ho Jeong; Jonas Brunskog; jmha@iru.interacoustics.com
Subject: ARO Abstract Confirmation 41st MidWinter Meeting

Call for Abstracts: ARO 41st MidWinter Meeting

Access Key (password): **JRQT5964**

You can access your abstract any time by [clicking here](#).

Abstract Type: Poster Only

Member Status: Non-Member

Abstract Status: Complete

Abstract ID: 379658

Abstract Title: *Towards Predicting Room Acoustical Effects on Sound-Field ASSR from Stimulus Modulation Power*

Author(s)

[Valentina Zapata-Rodriguez](#) - PhD student, Interacoustics Research Unit (Role: Presenting Author)

[Søren Laugesen](#) - Senior Research Engineer, Interacoustics Research Unit (Role: Co-Author)

[Cheol-Ho Jeong](#) - Associate Professor, Acoustic Technology group, Department of Electrical Engineering, Technical University of Denmark (Role: Co-Author)

[Jonas Brunskog](#) - Associate Professor, Acoustic Technology group, Department of Electrical Engineering, Technical University of Denmark (Role: Co-Author)

[James M. Harte](#) - Research manager, Interacoustics Research Unit (Role: Co-Author)

Topic

Auditory Prostheses

Abstract

One of the most important goals in early intervention of hearing loss is to ensure the child's access to speech. This can enable hearing impaired infants to develop language skills to a level comparable to normal-hearing infants. Hearing-aid fitting validation is important to ensure an appropriate amplification. However, this becomes challenging in pre-lingual infants because they do not respond to behavioral tests. For this reason, there is a growing interest in using objective electrophysiological measures for hearing-aid validation. Here, an approach based on the auditory steady-state response (ASSR) is considered. Instead of using insert earphones to deliver the stimuli, as is customary, the auditory signals are reproduced from a loudspeaker placed in front of the subject, so as to include the hearing aid in the transmission path. Loudspeaker presentation of the stimulus can lower its effective modulation depth due to reverberation and background noise in the measurement room. This could be critical for the quality of the measurement as ASSR magnitude is dependent on the amount of modulation in the stimulus. Previous studies have shown a reduction in the response magnitude as the modulation depth decreases, indicating a slope of about $s = 0.8$

between response magnitude and modulation depth both in dB (Boettcher et al. , 2001; Rønne, 2012; Bharadwaj et al. , 2015). However, the relation between observed sound-field ASSR magnitude and changes to stimulus modulation brought about by the acoustical properties of the measurement room has not been considered. The present work explores the relation between the stimulus modulation power and the ASSR amplitude in a simulated sound-field ASSR data set with varying reverberation time. Three rooms were simulated using the Green's function approach, and the impulse responses were convolved with narrow-band (NB) CE-Chirps centered at the octave-bands of 0. 5, 1. 0, 2. 0 and 4. 0 kHz. Fifteen normal-hearing adults were presented with the auralized stimuli, as well as an unmodified 'dry' version using insert earphones. The modulation power analysis is done based on the physiological input/output curves previously mentioned. This study discusses to what extent this modulation-growth function can be used as a prediction model to determine the changes in the ASSR amplitude based on the stimulus modulation in the room.

Authors will be formally notified of acceptance and the scheduled format and date/time of their presentation will be provided at a later date.

NOTE: Travel Awards - If you are a Post Doc Graduate Student, Medical or Resident or Audiologist, make sure to look for information pertaining to the application site opening on September 1st.

If you have any questions about your abstract, please Contact Wendy Stevens at (856) 423-0041, option 2.

Don't forget the abstract process closes 9/15/2017. Any changes, revisions or additions must be made prior to that date.

Technical Support

Email: Help@ConferenceAbstracts.com

Phone: (410) 638-9239