Determination of Ultralow Level ^{135}Cs and $^{135}\text{Cs}/^{137}\text{Cs}$ Ratio in Environmental Samples by Chemical Separation and Triple Quadrupole ICP-MS

Zhu, Liuchao; Hou, Xiaolin; Qiao, Jixin

Published in:
Analytical Chemistry

Link to article, DOI:
[10.1021/acs.analchem.0c01153](https://doi.org/10.1021/acs.analchem.0c01153)

Publication date:
2020

Document Version
Publisher's PDF, also known as Version of record

[Link back to DTU Orbit](#)

Citation (APA):
Zhu, L., Hou, X., & Qiao, J. (2020). Determination of Ultralow Level ^{135}Cs and $^{135}\text{Cs}/^{137}\text{Cs}$ Ratio in Environmental Samples by Chemical Separation and Triple Quadrupole ICP-MS. *Analytical Chemistry*, 92(11), 7884–7892. <https://doi.org/10.1021/acs.analchem.0c01153>

General rights

Copyright and moral rights for the publications made accessible in the public portal are retained by the authors and/or other copyright owners and it is a condition of accessing publications that users recognise and abide by the legal requirements associated with these rights.

- Users may download and print one copy of any publication from the public portal for the purpose of private study or research.
- You may not further distribute the material or use it for any profit-making activity or commercial gain
- You may freely distribute the URL identifying the publication in the public portal

If you believe that this document breaches copyright please contact us providing details, and we will remove access to the work immediately and investigate your claim.

Determination of Ultralow Level ^{135}Cs and $^{135}\text{Cs}/^{137}\text{Cs}$ Ratio in Environmental Samples by Chemical Separation and Triple Quadrupole ICP-MS

Liuchao Zhu, Xiaolin Hou,* and Jixin Qiao

Cite This: <https://dx.doi.org/10.1021/acs.analchem.0c01153>

Read Online

ACCESS |

Metrics & More

Article Recommendations

Supporting Information

ABSTRACT: An analytical method was developed for the determination of ultralow level ^{135}Cs in environmental samples by chromatographic separation of cesium with AMP-PAN and AG50W-X8 columns and sensitive measurement of cesium isotopes with triple quadrupole inductively coupled plasma mass spectrometry (ICP-MS/MS). Cesium was simply released by acid leaching using *aqua regia* from environmental solid samples and preconcentrated on AMP-PAN column. The cesium adsorbed on the column was effectively eluted with NH_4Cl solution without dissolving the AMP. The excessive amount of NH_4Cl in the eluate was removed by sublimation in the presence of small amount of LiCl . The remaining barium and other interfering elements such as Mo , Sn , Sb , and Li were efficiently removed using cation exchange chromatography (AG50W-X8). The decontamination factors of this procedure are above 4×10^7 for barium and 4×10^5 for molybdenum; the chemical yields of cesium are more than 85% for samples of less than 10 g. This method enables to separate cesium from large size of samples for the determination of ultralow level ^{135}Cs , avoiding the problem of removal of a huge amount of Mo in the dissolved AMP. Intrinsic ^{137}Cs in the environmental samples measured by gamma spectrometry before and after separation was used as internal isotope dilution standard for quantitative determination of ^{135}Cs without complete release and recover of radiocesium. The interference of barium (^{135}Ba and ^{137}Ba) to the ICP-MS measurement of ^{135}Cs and ^{137}Cs was further suppressed to 8×10^{-5} by using N_2O as the reaction gas in ICP-MS/MS at a flow rate of 0.7 mL/min, so a total suppression of 2×10^{-12} for Ba was achieved, making the isobaric interference of Ba isotopes to the measurement of ^{135}Cs and ^{137}Cs in environmental samples negligible. A detection limit of 9.1×10^{-17} g/g for ^{135}Cs and ^{137}Cs was achieved for 60 g samples. The developed method was validated by analysis of standard reference materials (IAEA-375, IAEA-330, and IAEA-385) and successfully applied for the determination of ^{135}Cs concentrations and $^{135}\text{Cs}/^{137}\text{Cs}$ ratios in soil samples collected from Denmark, Sweden, and Ukraine. The $^{135}\text{Cs}/^{137}\text{Cs}$ isotopic ratios in Danish soil (2.08–2.68) were significantly higher than that from Sweden and Ukraine (0.65–0.71), indicating different sources of radiocesium. This work demonstrated the application of $^{135}\text{Cs}/^{137}\text{Cs}$ as a unique fingerprint for discriminating the sources of radioactive contamination and estimating their contribution to the total inventory, which will be useful for nuclear forensics and environmental tracer studies.

Radioisotopes of cesium are important in environmental protection due to their high production and volatile feature of cesium; among them, ^{137}Cs and ^{134}Cs are the major concern. ^{135}Cs is one of important radionuclides in the waste repository due to its long half-life (2.3×10^6 years) and relatively large inventory. These radioisotopes of cesium are mainly produced by fission of nuclear fuel with some neutron activation production of ^{134}Cs , and released to the environment mainly by atmospheric nuclear weapons tests, nuclear accidents, spent nuclear fuel reprocessing plants, and nuclear power and research reactors.¹ Based on the source specific $^{134}\text{Cs}/^{137}\text{Cs}$ ratios, they have been used as tracers for identification of sources of radioactive substances,^{2,3} estimation of sedimentation rate, and circulation of water mass in the ocean.^{4,5} However, due to the short half-life of ^{134}Cs (2.1

years), the $^{134}\text{Cs}/^{137}\text{Cs}$ ratio can only be used for a limited time (<10 years). The Fukushima derived ^{134}Cs in the environment has already been difficult to measure in the present environmental samples.

In the production of ^{135}Cs by fission of nuclear fuel, besides direct fission production of ^{135}Cs , most ^{135}Cs is produced by the decay of other short-lived fission products of isobars

Received: March 16, 2020

Accepted: May 5, 2020

Published: May 5, 2020

(^{135}Te , ^{135}I and ^{135}Xe). Due to an extremely large neutron capture cross section (2×10^6 b), part of the formed ^{135}Xe is converted to ^{136}Xe instead of decaying to ^{135}Cs , causing the accumulated fission yield of ^{135}Cs to vary with neutron flux and irradiation time. While, the fission yield of ^{137}Cs does not change because of no such neutron poison in the fission chain production of ^{137}Cs . The $^{135}\text{Cs}/^{137}\text{Cs}$ ratio therefore shows a source dependent value and is related to the neutron flux and irradiation time of the nuclear fuel (^{235}U or ^{239}Pu),⁶ which makes it a fingerprint for source identification or indicator of nuclear power plant operation.⁷ The $^{135}\text{Cs}/^{137}\text{Cs}$ atomic ratio at the time of explosion of nuclear weapons was estimated to be 1.0, and the global fallout $^{135}\text{Cs}/^{137}\text{Cs}$ atomic ratio was measured to be 3.7 until 2020.⁸ Much lower $^{135}\text{Cs}/^{137}\text{Cs}$ ratios of 0.28–0.32 for the Chernobyl derived radiocesium (decay corrected to April 1986)⁹ and 0.33–0.38 for the Fukushima derived radiocesium (decay corrected to March 2011)¹⁰ were reported, while various ratios of $^{135}\text{Cs}/^{137}\text{Cs}$ of 0.4–1.6 in the discharges from reprocessing plants and influent from nuclear reactors were reported,¹¹ reflecting the influence of neutron flux and irradiation time, as well as the time interval between production and analysis. This shows the potential to replace the short-lived ^{134}Cs with ^{135}Cs in tracer applications.^{12–16}

As a low-energy beta emitter, the measurement of ^{135}Cs by radiometric method is usually hindered by the coexistence of high-energy beta emitter ^{137}Cs , which is often several orders of magnitude higher radioactivity compared to that for ^{135}Cs . Mass spectrometry is therefore a practical technique for ^{135}Cs measurement in the environmental samples. Among them, resonance ionization mass spectrometry (RIMS),¹⁷ thermal ionization mass spectrometry (TIMS),^{18–20} sector field inductively coupled plasma-mass spectrometry (SF-ICP-MS),^{21,22} electric thermal vaporization ICP-MS (ETV-ICP-MS),²³ and triple quadrupole inductively coupled plasma mass spectrometry (ICP-MS/MS)^{12–14} have been used for this purpose. ICP-MS/MS is becoming the most attractive mass spectrometric technique for the measurement of ^{135}Cs due to its highest popularity, high abundance sensitivity, and excellent removal of barium interference by reaction cell technique.

Due to the low concentration of ^{135}Cs in ordinary environmental samples (<1 mBq/kg), cesium needs to be preconcentrated and separated from the sample matrix and interferences before measurement. It has been widely reported that cesium is strongly bound in minerals in the soil and sediment,¹ and it is therefore important to effectively release the binding radiocesium from sample matrix to the solution before chemical separation. The elimination of isobaric interference is the most critical issue for the accurate measurement of ^{135}Cs and ^{137}Cs by ICP-MS/MS, because the mass concentrations of stable isobars (^{135}Ba and ^{137}Ba) are more than 8 orders of magnitude higher than ^{135}Cs and ^{137}Cs in the environmental samples. The interference of polyatomic ions formed by Mo, Sn, and Sb with isotopes of light elements (e.g., O, S, Ar, etc.) is another critical issue and has to be resolved. Due to the specific adsorption of Cs, AMP powder and AMP-PAN resin have been used to isolate Cs from the sample matrix, and the AMP matrix with adsorbed Cs was usually dissolved with ammonia for further purification of cesium. A combined anion and cation exchange chromatographic (AG MP-1 M and AG 50W-X8) separation procedure has been reported,^{13,14} in which MoO_4^{2-} in the sample solution was removed by anion exchange column, and the Cs remained in the effluent was purified using a cation exchange

column. This method has been successfully applied for analysis of Fukushima accident derived environmental samples.^{13,14}

However, it might be not suitable for analysis of large size of environmental background samples (e.g., >100 g soil or sediment and large volume of seawater) with ultralow level ^{135}Cs . In this case, a large amount of AMP powder or AMP-PAN resin needs to be used to preconcentrate cesium, consequently a massive amount of Mo and Ba is introduced to the sample solution by complete dissolution of AMP using ammonia, which will make the following purification of Cs from Mo and Ba very challenging.

This work aims to develop a method for accurate determination of ultralow level ^{135}Cs in large size environmental solid samples. The work focuses on effective release of cesium from environmental solid samples, selective separation of cesium from sample matrix using a AMP-PAN column without introducing Mo into sample solution, complete elimination of barium isobaric interference, and sensitive measurement of ^{135}Cs and ^{137}Cs using ICP-MS/MS.

■ EXPERIMENTAL SECTION

Instrumentation. A triple-quadrupole ICP-MS instrument (Agilent 8800, U.S.A.) was employed for the measurement of ^{135}Cs and ^{137}Cs in this work, which was equipped with an octopole dynamic reaction cell (DRC) between two quadrupole mass filters. The first quadrupole (Q1) selects the target ions with certain m/z values (e.g., $m/z = 135$ and 137) and introduces it into the DRC. The collision/reaction gas injected to the DRC reacts with the entered ions to remove the interfering ions by forming new ions. The analyte ions (original or newly formed) are selected in the second quadrupole (Q2) and measured in the following detector. N_2O gas (99.999% purity) was injected to the DRC as collision/reaction gas. Apex-Q sample introduction system (Elemental Scientific Instruments, U.S.A.) was used to improve the sensitivity of cesium isotopes.

Reagents and Samples. All reagents used in this work were of analytical reagent grade. Ultrapure water (18.2 M Ω cm, Sartorius water purification system, arium pro VF) and distilled HNO_3 (by Savillex DST-1000 system, U.S.A.) were used for preparation of cesium solution from AG 50W-X8 column for ICP-MS/MS measurement. AMP-PAN (100–600 μm , Triskem International) resin and strong acidic cation exchange resin AG 50W-X8 (100–200 mesh, H^+ form, analytical grade, Bio-Rad laboratories) were used for separation and purification of cesium.

Reference materials IAEA-375 (soil, collected from Novozybkov, Bryansk, Russia in 1990), IAEA-385 (Irish sea sediment, collected in 1995), and IAEA-330 (spinach, collected from Poleskoe, Kiev, Ukraine in 1990) were purchased from the International Atomic Energy Agency (IAEA). Two Danish soil samples were analyzed in this study, one (Soil-1) was collected from 0 to 5 cm using a stainless steel cylinder in Roskilde, and another one (Soil-2) is a mixed surface soil (0–5 cm), which was collected from 8 different sites (Tylstrup, Borris, Askov, St. Jyndevad, Tystofte, Tornbygård) in Denmark. One soil from Gavle, Sweden and three soils from Feofaniya, Ukraine were collected from 0 to 5 cm depth in 1992. The sampling sites are shown in Figure S1.

Release of Cesium from Solid Samples. Soil or sediment samples were first oven-dried at 80 $^\circ\text{C}$ and then ashed in a muffle oven at 450 $^\circ\text{C}$ for 8 h to decompose the

organic substances. A total of 1–60 g of soil samples were used in this work depending on the measured ^{137}Cs level. Two methods were applied to release cesium from soil/sediment samples, one was acid leaching using aqua regia and the other one was fusion using LiBO_2 . For acid leaching, aqua regia was added to the ashed soil or sediment with a ratio of 8 for acid/sample (mL/g). A small volume of acid (HCl) was slowly added to the sample to dissolve carbonate to avoid a sudden release of a large amount of CO_2 , which could cause a spillover of the sample out of the beaker. The sample with acid solution in a beaker was covered with a watch glass and heated at $180\text{ }^\circ\text{C}$ for 2 h on a hot plate. The leachate was separated by filtration through a glass fiber filter, the residue was washed with diluted HCl and the washing solution was combined with leachate. For fusion method, LiBO_2 was mixed with the ashed soil in a graphite crucible with a mass ratio of 1:2 for sample/ LiBO_2 . The sample with LiBO_2 was fused at $950\text{ }^\circ\text{C}$ for 20 min. The obtained molten button of sample was dissolved with aqua regia, and the solution was used for separation of cesium. The residue was dried and measured for the remaining ^{137}Cs to monitor the recovery in this step. The leachate was evaporated to near-dryness and dissolved in 1.0 M HNO_3 .

Separation of Cesium Using AMP-PAN Column.

Cesium was separated and purified from the sample matrix and interferences using AMP-PAN column followed by a cation exchange (AG 50W-X8) chromatography. The prepared sample solution was loaded into an AMP-PAN column (1.0 mL resin packed in a 2 mL column, ϕ 7 mm $\times$ 5 cm), which has been preconditioned with 40 mL of 4 M NH_4NO_3 , 20 mL of ultrapure H_2O , and 10 mL of 1 M HNO_3 . The column was rinsed with 30 mL of 1 M HNO_3 and 20 mL of deionized H_2O to remove matrix elements. Cesium adsorbed on the column was eluted with selected ammonium salt (e.g., NH_4Cl , NH_4NO_3 , and $(\text{NH}_4)_2\text{SO}_4$).

Removal of NH_4Cl Salt from the Eluate of Cesium.

Before further purification of cesium (eluted from AMP-PAN column) using cation exchange chromatography, NH_4Cl needs to be removed because of the competitive adsorption of NH_4^+ with Cs^+ on the cation exchange resin. The NH_4Cl eluate with cesium was heated to dryness on a hot plate at $180\text{ }^\circ\text{C}$ after adding LiCl (or other protective reagents). The residue in the beaker was then heated at different temperature (e.g., 350, 400, and $450\text{ }^\circ\text{C}$) for about 1–2 h in a muffle oven to remove NH_4Cl . After cooling to room temperature, the remaining residue was dissolved with 10 mL of 1.5 M $\text{NH}_3\cdot\text{H}_2\text{O}$. For method development, a small fraction of the solution (0.1 mL) was taken for the measurement of Cs content using ICP-MS to monitor the recovery of cesium in this step.

Purification of Cesium from Li, Ba, and Mo. The obtained sample solution from the last step was loaded onto a cation exchange column (10 mL of AG 50W-X8, ϕ 1.0 cm $\times$ 20 cm, H^+ form), which was preconditioned with 50 mL of 5.0 M HNO_3 , 20 mL of deionized H_2O , and 20 mL of 1.5 M $\text{NH}_3\cdot\text{H}_2\text{O}$. The column was rinsed with 50 mL of 1.5 M $\text{NH}_3\cdot\text{H}_2\text{O}$ and 20 mL of deionized H_2O (and 70 mL of 1 M HNO_3 for optimized procedure). Cesium remaining in the column was eluted with different concentrations of HNO_3 (e.g., 0.5, 1.0, 1.5, and 2.0 M). For method development, each 10 mL of eluate was collected and measured for Li, Cs, and Ba using ICP-MS to obtain an elution curve. For analysis of samples, the collected eluate of cesium (e.g., 70 mL of 1 M HNO_3) was evaporated to dryness and redissolved with 2 mL of 0.30 M HNO_3 for measurement of $^{135}\text{Cs}/^{137}\text{Cs}$ ratio using ICP-MS/

MS. The schematic diagram of the optimized analytical procedure is shown in Figure S2.

Measurement of ^{137}Cs Using Gamma Spectrometry.

Radioactivity of ^{137}Cs in all samples was measured using an HPGe gamma spectrometer (with a relative counting efficiency of 35% and a resolution of 1.85 keV for 1332 keV gamma peak of ^{60}Co) by counting its 661.6 keV γ rays. The dried samples were sealed into a plastic container and put on the top of the detector (nearest distance). Gennie 2000 software (Canberra) was used to acquire gamma spectra and perform spectra analysis. The gamma spectrometry was calibrated for counting efficiency using a mixed gamma standard; the measurement results of ^{137}Cs were corrected for counting geometry, self-absorption (density and sample components), and sum coincidence.

Measurement of Cesium Isotopes (^{135}Cs and ^{137}Cs) with ICP-MS/MS. Cesium isotopes (^{135}Cs and ^{137}Cs) in the prepared solution were measured using ICP-MS/MS. The instrument was tuned using stable ^{133}Cs (1.0 ng/mL) prepared using cesium standard solution (CPI international, Amsterdam, Netherlands). The blanks, standards, and samples were prepared in 0.30 M HNO_3 for measurement. The sample solution was introduced using MicroMist nebulizer with a peristaltic pump at 0.1 mL/min for ^{133}Cs measurement and optimization of the instrument setup. Pt skimmer cone, s-lens, and hot plasma were used in the measurement of cesium isotopes. N_2O (99.999% purities) was injected to the DRC as reaction gas in the measurement of ^{135}Cs and ^{137}Cs . For measurement of ^{135}Cs and ^{137}Cs in low-level environmental samples, an Apex-Q sample introduction system was employed to improve nebulization efficiency, measurement sensitivity of cesium isotopes and analytical precision. For measurement of ^{135}Cs , both Q1 and Q2 were set to $m/z = 135$, and for ^{137}Cs , they were set to 137.

RESULTS AND DISCUSSION

Extraction of Cesium from Environmental Solid Samples. Both acid leaching using aqua regia and complete decomposition using LiBO_2 fusion followed by acid dissolution were investigated for releasing radiocesium (^{135}Cs and ^{137}Cs) from the soil/sediment to solution. Since both ^{135}Cs and ^{137}Cs in the environment were originated from human nuclear activities, and their chemical properties and environmental behaviors are similar, ^{137}Cs was therefore measured in the samples before and after decomposition using gamma spectrometry and the efficiency of the decomposition methods was evaluated. The results (Table S1) show that a high recovery of more than 93% for ^{137}Cs was observed using LiBO_2 fusion. This should be attributed to the complete decomposition of the sample indicated by the entire melt of the sample with flux at high temperature. A small loss of ^{137}Cs in this method might be attributed to the following dissolution of the molten button. It is interesting that acid leaching using aqua regia also showed a high leaching efficiency of ^{137}Cs from soil/sediment, for less than 10 g of samples, more than 85% of ^{137}Cs was leached, and even for a sample of up to 60 g, the ^{137}Cs recovery was higher than 60%. This indicates that acid leaching is sufficiently good for releasing ^{135}Cs and ^{137}Cs from soil/sediment. Since the acid leaching is simple, easy operation, less time-consuming, and more suitable for treatment of a large size sample compared with fusion method, acid leaching using aqua regia with a sample/acid ratio of 1:8 at $180\text{ }^\circ\text{C}$ for 2 h is

recommended and used in this work. The content of ^{137}Cs in the original samples and the separated cesium samples were measured using gamma spectrometry to calculate the recovery of ^{135}Cs in the chemical separation procedure. $^{135}\text{Cs}/^{137}\text{Cs}$ atomic ratios in the final prepared samples were measured using ICP-MS/MS, which was used in combination with the measured ^{137}Cs concentrations in the original samples for calculation of ^{135}Cs concentrations in the samples, i.e., ^{137}Cs was used as an internal isotope dilution standard, as well as a chemical yield tracer. Meanwhile the incomplete release and recover of ^{135}Cs in the sample preparation does not influence the final analytical results of ^{135}Cs concentration.¹² It has been widely reported that ^{137}Cs deposited in soil from human nuclear activities was firmly bound to soil grains and mainly in the clay minerals and micas component by occupying/replacing the site of potassium in the crystal of clay minerals.^{23–26} The fractionation analysis of soil and sediment samples has shown that only a small percentage of ^{137}Cs presents as exchangeable, or associated with carbonate, organic substances or oxides, but mainly associated with minerals,²⁷ and *aqua regia* could leach 70–95% of ^{137}Cs from the soil and sediment. The leaching efficiency of ^{137}Cs is also highly related to the type and properties of the sample. For soil/sediment sample, ashing at relatively low temperature (450 °C) is necessary for decomposition of organic substances and avoiding loss of cesium at high temperature due to its volatile property.

Preconcentration of Cesium from Sample Solution by AMP-PAN Column. The acid leachate of the sample contains high level of matrix elements (cations), a preconcentration of cesium from the sample matrix is needed before further purification. Due to the selective adsorption of cesium, AMP powder or AMP-PAN column were widely applied for preconcentration of cesium from seawater or sample solution.¹ Our experiment showed that more than 97% cesium in the acidic solution (e.g., 1 M HNO_3) can be adsorbed on AMP powder or AMP-PAN resin. For the determination of ^{135}Cs , the adsorbed cesium on the AMP or AMP-PAN resin needs to be eluted for further purification and measurement of ^{135}Cs using ICP-MS/MS. Ammonia has been used to remove cesium from AMP powder.¹⁴ In this case, all AMP was dissolved. Consequently, all interference elements (e.g., Ba, Mo, Sn, Sb, etc.) also entered into the sample solution, which will influence removal of these interferences. Meanwhile, large amounts of molybdate, phosphate, and potassium were also dissolved into the solution, which significantly influence the further purification of cesium from interferences. The highly selective adsorption of Cs^+ on AMP-PAN is based on the replacement of NH_4^+ in the AMP by Cs^+ . Therefore, a high concentration of NH_4^+ solution might be possible to replace and elute Cs^+ adsorbed on the AMP-PAN without dissolving the AMP component. The elution characteristics of Cs^+ from the AMP-PAN column with different ammonium salts were investigated. The results show that NH_4HCO_3 , $(\text{NH}_4)_2\text{CO}_3$, NH_4F , $\text{NH}_4\text{C}_2\text{O}_4$, NH_4Ac , and NH_4Cit solution can directly dissolve AMP component, similar as ammonia solution, therefore these reagents could not be used for this purpose. Ammonium thiocyanate solution can react with AMP and reduce it to phosphorus molybdenum blue, as indicated by the color change of the AMP-PAN resin from yellow to blue. NH_4Cl , NH_4NO_3 , and $(\text{NH}_4)_2\text{SO}_4$ solutions do not dissolve AMP but could elute Cs^+ from the AMP-PAN resin. In the eluate of these reagents, the high concentration of NH_4^+ will strongly

influence the further separation of Cs^+ using cation exchange chromatography because of its competitive adsorption with Cs^+ on the resin and has to be removed before separation. Considering the easy removal of NH_4Cl by the heating method based on its sublimation at low temperatures (338 °C), the risk for dangerous explosion of NH_4NO_3 during heating, and difficulties in removal of sulfate, NH_4Cl was selected to elute Cs^+ from the AMP-PAN resin.

The elution profiles of Cs^+ from AMP-PAN column (1.0 mL) with different concentrations of NH_4Cl solution (Figure 1) show that cesium can be efficiently eluted with high

Figure 1. Elution profiles of Cs with different concentrations of NH_4Cl solution from 1 mL of AMP-PAN column.

concentration of NH_4Cl , and a recovery of >95% for cesium was obtained using 40 mL of 4 M or 5 M NH_4Cl . While, a large volume and long tailing were observed when eluted with 3 M or lower concentrations of NH_4Cl . In considering the amount of NH_4Cl in the eluate, 40 mL of 4.0 M of NH_4Cl was applied to elute cesium from 1 mL of AMP-PAN column.

In addition, the AMP-PAN column can efficiently remove Ba, Sb, and Sn by factors of 7300, 20, and 1400, respectively, which are much better than that using AMP powder.¹⁴ This is very critical to achieve a better separation of Cs from Ba and other interfering elements in a large size soil or sediment samples (e.g., 60 g) for ultralow-level ^{135}Cs measurement.

Removal of Ammonium from Cs Eluate. Cation exchange chromatography demonstrated an effective technique for purifying cesium from Ba and other interfering elements. However, due to the competitive adsorption, almost no adsorption of Cs^+ on the AG 50W-X8 column (10 mL) was observed when directly loading the cesium eluate in 4 M NH_4Cl medium. The massive NH_4^+ ions in the eluate have to be removed prior to the purification of cesium by cation exchange chromatography. However, due to the similar chemical properties of Cs^+ and NH_4^+ , it is difficult to remove NH_4^+ without loss of Cs^+ . Based on the low sublimation temperature of NH_4Cl (338 °C) but not CsCl , a simple heating method was developed to remove NH_4Cl by sublimation. The eluate was first evaporated to dryness on a hot plate; the residue was then heated at 350–450 °C to remove NH_4Cl . The results show that almost all NH_4Cl was sublimated at a temperature of above 350 °C for more than 2 h. However, it was observed that most of cesium (>60%) was

also lost when NH_4Cl was sublimated, even heated at only 350 °C (Table 1).

Table 1. Recoveries of Cesium during the Sublimation of NH_4Cl Eluate at Different Temperatures and Different Protective Reagents

sample no.	temperature/°C	protective reagent	amount of protective reagent	recovery of Cs/%
1	350	no		31 ± 8 ($n = 3$)
2	400	no		32 ± 8 ($n = 3$)
3	450	no		34 ± 9 ($n = 3$)
4	350	H_3PO_4	1.5 mmol	67 ± 7 ($n = 3$)
5	350	LiCl	8 mg	47 ± 6 ($n = 3$)
6	350	LiCl	24 mg	76 ± 5 ($n = 3$)
7	350	LiCl	40 mg	91 ± 3 ($n = 3$)
8	350	LiCl	60 mg	99 ± 3 ($n = 3$)
9	350	LiCl	80 mg	100 ± 3 ($n = 3$)
10	350	NaCl	80 mg	91 ± 5 ($n = 3$)
11	350	CaCl_2	80 mg	89 ± 5 ($n = 3$)

Cesium chloride is not a volatile species, with a melting point of 646 °C and a boiling point of 1297 °C. The loss of cesium in the heating process might be attributed to its small amount ($<20 \mu\text{g}$) compared to the extremely large amount of NH_4Cl , resulting in a $\text{NH}_4\text{Cl}/\text{CsCl}$ mass ratio of more than 5×10^5 . During the sublimation process, a massive amount of NH_4Cl gas (possibly also with some NH_3 and HCl) was released, and a trace amount of CsCl wrapped in the NH_4Cl might be carried out in this process. It was noticed that, by adding H_3PO_4 into the eluate before evaporation, the recovery of cesium was significantly improved to 67% (Table 1). Addition of other salts such as LiCl , NaCl , and CaCl_2 , the recovery of cesium was also significantly improved; almost quantitative recovery of Cs was obtained when more than 60 mg of LiCl was added. LiCl , NaCl , and CaCl_2 are not volatile compounds; CsCl might be adsorbed on the particles of these salts during the sublimation of NH_4Cl , thus preventing the loss of CsCl . The increased recoveries of Cs with the amount of LiCl of less than 60 mg (Table 1) confirmed that the sufficient amount of particles/salt are important for preventing loss of Cs during sublimation of NH_4Cl . However, cations of Li^+ , Na^+ , and Ca^{2+} also compete with Cs^+ to adsorb on the cation exchange resin during the purification of cesium. Due to much low relative selectivity of Li (0.85) on the cation exchange resin (AG 50W) compared to Cs (2.7), Na (1.5), and Ca (3.9), LiCl has much less influence on the adsorption of Cs^+ on the cation exchange resin. Therefore 60 mg of LiCl was selected in this study.

Elimination of Ba and Other Interfering Elements Using Cation Exchange Chromatography. Although AMP-PAN is a specific resin for separation of Cs from the sample matrix, the removal efficiencies for most of interfering elements, such as Ba, Mo, Sb and Sn, are not sufficient high for the low-level ^{135}Cs measurement by ICP-MS/MS. Besides the

interfering elements, Li^+ added in the last step also needs to be removed due to matrix effect in ICP-MS/MS measurement. Based on the different affinity/relative selectivity of Cs^+ and these interfering elements, strong acidic cation exchange resin (AG 50W-X8) was applied for purification of cesium. The affinity of Cs^+ is much higher than that of Li^+ but lower than that of Ba^{2+} . Li^+ is first removed from the column; Cs^+ is then eluted but keeping Ba^{2+} on the column. The decontamination factor of Ba relies on the eluent and its concentration. After removal of NH_4Cl by sublimation, the cesium sample was prepared in a small volume of 1.5 M $\text{NH}_3 \cdot \text{H}_2\text{O}$ (10 mL). In this alkaline media, Mo, Sn, and Sb can be converted to anion species (MoO_4^{2-} , $\text{Sn}(\text{OH})_6^{2-}$ and SbO_3^-), which are not adsorbed on the cation exchange resin. After loading the Cs sample solution to AG 50W-X8 column, and rinsing with 50 mL of 1.5 M $\text{NH}_3 \cdot \text{H}_2\text{O}$ and 20 mL of H_2O , anion species of Mo, Sn, and Sb can be well eliminated. HNO_3 was then used for elution of Cs, which can avoid introduction of any salt to the final solution and benefit for sensitive measurement of cesium isotopes. Meanwhile, the eluate can be evaporated to dryness and prepared in a small volume solution for ICP-MS/MS measurement.

The elution profiles (Figure 2) show that 2.0 M HNO_3 can quickly elute Cs from the column, but Ba can be also quickly

Figure 2. Elution profiles of Cs, Ba, and Li on 10 mL of AG50-X8 column ($\phi 1.0 \times 20$ cm, H^+ form) with different concentrations of HNO_3 .

eluted and overlap with Cs eluate. By eluting with lower concentrations of HNO_3 (e.g., 0.5 M), Cs can be completely separated from Ba, but a large volume of eluent is needed. The optimal elution was achieved using 1 M HNO_3 , and in this case, Cs can be completely separated from Ba and Li with 70 mL of 1 M HNO_3 , meanwhile most of K^+ and Rb^+ can be also removed due to their lower affinity compared to Cs^+ .¹⁴ A 10 mL cation exchange resin (AG50W-X8) in a column of $\phi 1.0 \times 20$ cm was employed for obtaining a better separation of Cs from Ba, Rb, and K.

The decontamination factors of Ba, Li, Mo, Sn, and Sb in the cation exchange chromatographic separation procedure were monitored by determination of the concentrations of these elements in the final eluted Cs fraction using ICP-MS and compared with the initiate amounts of these elements in the leached sample solution. The overall decontamination factors

of 4×10^7 for Ba, 4×10^6 for Li, 4×10^5 for Mo, 3×10^5 for Sn, and 2×10^5 for Sb were achieved, which are better or comparable compared to the reported values using combined chromatographic purification (AMP-PAN, AGMP-1 M, and AG 50W-X8).^{14,18} In consideration of all steps including ICP-MS measurement, the entire analysis of a batch of 8 samples can be completed within 3 days. The overall chemical yields of ^{135}Cs were measured by counting the ^{137}Cs in the original samples and the final separated cesium solution by gamma spectrometry to be more than 85% for soils from Sweden and Ukraine. The recoveries of radiocesium in each step were also measured by counting ^{137}Cs , being about 90% for acid leaching step, 97% for AMP-PAN column separation, 99% for NH_4Cl sublimation process, and 99% for cation exchange resin purification step. A slightly lower chemical yields of 60–70% for cesium was obtained for Danish soil samples, which is mainly attributed to the lower leaching rate of cesium in the step of *aqua regia* leaching of big soil samples (60 g). However, this can be improved by dividing big sample into a few aliquots for leaching to increase the leaching efficiency.

For the analysis of real environmental samples, the Cs eluate in 1.0 M HNO_3 was evaporated to dryness and then prepared in 2.0 mL of 0.3 M HNO_3 for ICP-MS/MS measurement. The concentrations of Ba, Mo, Sb, and Sn in the final Cs fraction were less than 0.1, 1, 0.2, and 0.2 ng/mL respectively.

Measurement of Low-Level ^{135}Cs and ^{137}Cs by ICP-MS/MS. The major challenges in the measurement of ultralow level ^{135}Cs and ^{137}Cs (<2 fg/g) in environmental samples are elimination of isobaric interference of stable ^{135}Ba and ^{137}Ba , interference of polyatomic ions including $^{95}\text{Mo}^{40}\text{Ar}^+$, $^{97}\text{Mo}^{40}\text{Ar}^+$, $^{119}\text{Sn}^{16}\text{O}^+$, and $^{121}\text{Sb}^{16}\text{O}^+$ and suppressing the tailing of stable ^{133}Cs . In this work, two sequential quadrupole mass filters equipped in the Agilent 8800 was used to reduce the tailing influence of ^{133}Cs ; collision reaction technique with N_2O as reaction/collision gas was used to eliminate Ba interference, as well as polyatomic ions interference. The results (Figure 3) show that the signal intensities of Ba^+ ions were significantly decreased by 5 orders of magnitude when the flow rates of N_2O gas injected to the DRC were increased from 0 to 0.9 mL/min., while the signal intensities of Cs^+ ions were only slightly reduced with the increased flow rates of N_2O .

Figure 3. Variation of intensities of Cs^+ and Ba^+ signal and ratios of $^{133}\text{Cs}/^{135}\text{Ba}$ with the flow rates of N_2O injected to the DRC. The standard solutions of Cs (1.0 ng/mL) and Ba (5.0 ng/mL) were used in this experiment.

This is attributed to the high affinity of Ba^+ ion with oxygen atom (92.8) compared to Cs^+ ion (14) in N_2O gas,²⁸ causing formation of BaO^+ but not CsO^+ . With 0.70 mL/min of N_2O as reaction gas, the measured $^{133}\text{Cs}/^{135}\text{Ba}$ ratios were improved by a factor of more than 2×10^4 , while the sensitivity of Cs was decreased by a factor of 8. At this condition, the measured sensitivity of Cs is 3.4×10^5 cps/ppb with the Apex-Q system, while the contributions of 0.2 ng/mL Ba at the $m/z = 135$ and 137 were less than 0.3 cps.

The interferences from polyatomic ions on $m/z = 135$ and 137, such as $^{95}\text{Mo}^{40}\text{Ar}^+$, $^{119}\text{Sn}^{16}\text{O}^+$, $^{97}\text{Mo}^{40}\text{Ar}^+$, and $^{121}\text{Sb}^{16}\text{O}^+$ were investigated by measuring the signal intensities at $m/z = 135$ and 137 for different concentrations of single element standard solutions of Mo, Sn, and Sb at the optimal condition of ICP-MS/MS with 0.7 mL/min. N_2O as reaction gas. The results (Figure S3) show that the signal intensities at $m/z = 135$ and 137 were very low but increased with the concentrations of these elements in the standard solution, indicating the polyatomic ions formed by these elements were effectively destroyed in the DRC by the injected reaction/collision gas. For 50 ng/mL solution of Sn, Sb, and Mo, the measured intensities at m/z of 135 and 137 were 4 cps for Sn, 70 cps for Sb, and 1 cps for Mo. Although the concentration of Mo in the final Cs eluate was relatively higher (1 ng/mL) compared to other elements, the interference from Mo (<0.02 cps) is still negligible. Because the concentrations of Sn and Sb in the final purified sample solution were less than 0.2 ng/mL, the interferences from these elements on the measurement of ^{135}Cs and ^{137}Cs (<0.3 cps) are also negligible. Meanwhile, it also indicates that the effective removal of these elements from Cs fraction is critical for accurate measurement of the low-level ^{135}Cs due to their relatively high concentrations in the soil and sediment samples (about 1–3 $\mu\text{g/g}$ Mo, Sb and Sn) compared to ^{135}Cs ($<3 \times 10^{-15}$ g/g).

The abundance sensitivity is another critical interference for the measurement of low-level ^{135}Cs due to the relatively high stable cesium concentration in environmental samples (1–3 $\mu\text{g/g}$) and low atomic ratio of $^{135}\text{Cs}/^{133}\text{Cs}$ ($<10^{-8}$ in background area); application of two sequential quadrupole mass filters for the measurement of ^{135}Cs by setting $m/z = 135$ for both quadrupoles significantly improved the abundance sensitivity. The signal intensity at $m/z = 135$ was measured using a cesium standard solution of 50 $\mu\text{g/mL}$ to evaluate the tailing influence of ^{133}Cs ; the abundance sensitivity (ratio of the count rate at $m/z = 135$ to that at $m/z = 133$) was estimated to be less than 1×10^{-11} . This is about 3 orders of magnitude lower than the $^{135}\text{Cs}/^{133}\text{Cs}$ ratio in normal environmental samples, therefore sufficient good for reliable measurement of ultralow level ^{135}Cs without a correction for the tailing contribution of ^{133}Cs . However, if large amount of stable cesium (e.g., >2 mg) was used as carrier in the chemical separation, the interference of ^{133}Cs tailing will be too serious (corresponding to 20 fg of ^{135}Cs) to obtain an accurate result of low-level ^{135}Cs even for a correction. Therefore, no stable cesium could be added as carrier in the chemical separation of cesium for the determination of ^{135}Cs using ICP-MS/MS.

Detection Limit of the Method for ^{135}Cs . By using Apex-Q sample introduction system, the measurement sensitivity was significantly improved to 3.4×10^5 cps for 1 ng/mL ^{133}Cs standard solution. The count rates at m/z of 135 and 137 in the procedure blanks were measured to be 0.3 ± 0.2 cps. Based on these performances, minimum detectable amount of ^{135}Cs and ^{137}Cs can be estimated to be 1.8×10^{-15}

Table 2. Comparison of the Measured ^{135}Cs Concentrations and $^{135}\text{Cs}/^{137}\text{Cs}$ Atomic Ratios in Standard Reference Materials with the Reported Values

sample	matrix	$^{135}\text{Cs}/^{137}\text{Cs}$ atomic ratio (this work)	^{135}Cs activity (mBq/kg)	$^{135}\text{Cs}/^{137}\text{Cs}$ atomic ratios (in literature)
IAEA-375 ^a	soil	0.30 ± 0.02 (<i>n</i> = 3)	24.3 ± 1.6	0.30 ± 0.01 ²⁹
				0.30 ± 0.01 ¹¹
				0.31 ± 0.01 ¹²
				0.27 ± 0.02 ³⁰
				0.32 ± 0.01 ³¹
IAEA-330 ^a	spinach	0.29 ± 0.02 (<i>n</i> = 1)	8.12 ± 0.56	0.29 ± 0.02 ⁹
IAEA-385 ^b	sediment	1.10 ± 0.17 (<i>n</i> = 3)	0.30 ± 0.05	0.80 ± 0.16 ¹²
				1.21 ± 0.14 ¹⁴

^aDecay corrected to April 1986. ^bDecay corrected to Jan 2015. The results for soil and sediment are the average and 1SD of 3 replicate samples, for spinach are the measurement results with overall analytical uncertainties (*k* = 1).

g/mL using 3 times standard deviation of the procedure blanks. For 60 g of soil sample, the detection limit of the method for ^{135}Cs and ^{137}Cs can be calculated to be 9.1×10^{-17} g/g (or 0.30 mBq/g for ^{137}Cs and 3.9 nBq/g for ^{135}Cs) considering a recovery of 65% for cesium and 2 mL final solution prepared for measurement; this value is more than 1 order of magnitude better than the reported value using ICP-MS/MS.¹⁴

Validation of the Developed Method. There is no standard reference material of similar matrix as soil and sediment with certified for ^{135}Cs concentration or $^{135}\text{Cs}/^{137}\text{Cs}$ ratio available, the standard reference materials of soil (IAEA-375, collected from contaminated area by Chernobyl accident in Russia), sediment (IAEA-385, collected from the Irish Sea) and spinach (IAEA-330, collected from Poleskoe, Kiev, Ukraine) with the reported $^{135}\text{Cs}/^{137}\text{Cs}$ ratios in literatures were therefore analyzed using the developed method. The results (Table 2) show that the measured $^{135}\text{Cs}/^{137}\text{Cs}$ atomic ratios for IAEA-375 and IAEA-330 are in well agreement with the reported data in the literature, indicating the developed method is reliable and accurate for the determination of ^{135}Cs and $^{135}\text{Cs}/^{137}\text{Cs}$ ratio in environmental samples. As for IAEA-385 sediment, only two data were reported,^{12,14} and our result agreed with the value reported by Zheng.¹⁴

$^{135}\text{Cs}/^{137}\text{Cs}$ Ratios in Soil from Denmark, Sweden, and Ukraine. Soil samples collected from Denmark, Sweden, and Ukraine were analyzed for ^{135}Cs concentrations and $^{135}\text{Cs}/^{137}\text{Cs}$ atomic ratios using the developed method. The results (Table 3) show that the measured $^{135}\text{Cs}/^{137}\text{Cs}$ atomic ratios in the soils collected from Gavle, Sweden and Feofaniya, Ukraine are similar (0.65–0.71), although the ^{137}Cs concentrations (140–1650 Bq/kg) are different in these samples by a factor of 2–10. While a much higher $^{135}\text{Cs}/^{137}\text{Cs}$ atomic ratios of (2.08–2.68), but a much lower ^{137}Cs concentrations (3–8 Bq/kg) were observed in soil samples from Denmark. This should be attributed to the different sources of ^{135}Cs and ^{137}Cs in these soil samples. As high as 1650 Bq/kg of ^{137}Cs was measured in the soil collected from Gavle, Sweden and 142–286 Bq/kg in Feofaniya, Ukraine, they are 1–3 orders of magnitude higher than the level in the soil received only global fallout.³² The measured $^{135}\text{Cs}/^{137}\text{Cs}$ ratios (0.65–0.71) (corresponding to 0.30–0.32 with decay correction to April 1986) are similar to those reported ratios in the Chernobyl accident derived radiocesium (0.28–0.32),⁹ indicating that most of radiocesium in these samples were originated from the Chernobyl accident fallout. It has been also reported that Gavle in Sweden and Feofaniya in Ukraine had received heavily radioactive fallout from Chernobyl accident in 1986,

Table 3. Analytical Results of ^{135}Cs Concentrations and $^{135}\text{Cs}/^{137}\text{Cs}$ Ratios in Surface Soil Samples from Denmark, Sweden, and Ukraine^a

sample	^{137}Cs activity (Bq/kg)	$^{135}\text{Cs}/^{137}\text{Cs}$ atomic ratios	^{135}Cs activity (mBq/kg)
Danish soil-1 (<i>n</i> = 2)	3.32 ± 0.16	2.68 ± 0.53	0.12 ± 0.02
Danish soil-2 (<i>n</i> = 2)	8.03 ± 0.31	2.08 ± 0.31	0.23 ± 0.03
Sweden soil (<i>n</i> = 1)	1650 ± 18	0.66 ± 0.06	14.7 ± 1.03
Ukraine soil-1 (<i>n</i> = 1)	283 ± 3	0.65 ± 0.07	2.49 ± 0.27
Ukraine soil-2 (<i>n</i> = 1)	142 ± 3	0.71 ± 0.08	1.36 ± 0.15
Ukraine soil-3 (<i>n</i> = 1)	150 ± 3	0.71 ± 0.08	1.44 ± 0.16

^aAll ^{137}Cs data are decay corrected to 1st Feb 2020. The uncertainties presented here are the overall uncertainty with *k* = 1

and other anthropogenic radionuclides related to Chernobyl accident were observed in these samples.³³ A much lower ^{137}Cs concentrations were observed in the Danish soil samples. The measured $^{135}\text{Cs}/^{137}\text{Cs}$ ratios in two Danish soil samples are 2.68 ± 0.53 (Soil-1) and 2.08 ± 0.31 (Soil-2), which is about 3–4 times higher than the ratios measured in the Chernobyl accident contaminated soil from Gavle, Sweden and Feofaniya, Ukraine. Danish Soil-1 was collected from upper 0–5 cm in Roskilde in 2019, while the Danish Soil-2 is a mixture of soil samples collected from 0 to 5 cm depth at 8 different sites over Denmark in 1987 (Table S2). Based on the assumption of only two sources of ^{137}Cs and ^{135}Cs in Denmark (global fallout and Chernobyl accident fallout) and the reported $^{135}\text{Cs}/^{137}\text{Cs}$ atomic ratios of 3.7 ± 0.4 for the global fallout and 0.65 ± 0.04 for Chernobyl accident releases (decay corrected to Feb 2020),^{9,11} the contributions of radiocesium from Chernobyl accident can be estimated to be 31% for Soil-1 and 51% for Soil-2. The estimated value in Soil-2 agree well with the estimation through the monitoring of ^{137}Cs level in soil at 8 different sites in Denmark collected before Chernobyl accident (1983) and after the Chernobyl accident (1987; Table S2). A less contribution of Chernobyl accident was estimated in Soil-1 compared to the Soil-2. The ^{137}Cs concentration (3.3 Bq/kg) in the Soil-1 is also lower than that in Soil-2 (8.0 Bq/kg) by a factor of 2.4, which supports the less fallout of Chernobyl accident in this site. This is the lowest radiocesium level environmental sample ever used for the radiocesium ratios determination by ICP-MS/MS, which will broaden the

application of $^{135}\text{Cs}/^{137}\text{Cs}$ as a tracer in environmental process studies.

The analytical results of these soil samples demonstrate that the developed method could be successfully applied for the determination of ^{135}Cs concentrations and $^{135}\text{Cs}/^{137}\text{Cs}$ ratios in ordinary environmental samples (with ^{137}Cs concentration down to 3.3 Bq/kg). It also indicates that the $^{135}\text{Cs}/^{137}\text{Cs}$ ratios can be used as a unique fingerprint for discriminating the sources of radioactive contamination and estimating their contributions to the total inventory, which will be useful for nuclear forensics and environmental tracer studies.

CONCLUSION

A method was developed for the determination of ^{135}Cs concentrations and $^{135}\text{Cs}/^{137}\text{Cs}$ atomic ratios in environmental solid samples by coupling AMP-PAN and AG 50W-X8 chromatographic separation with ICP-MS/MS measurement. A selective elution of Cs using NH_4Cl from AMP-PAN without dissolution of the AMP component was proposed for the first time, thereby significantly reducing the burden for removal of a large amount of molybdenum from cesium solution. Ammonium chloride in the eluate of Cs was removed satisfactorily by sublimation at 350 °C with the addition of LiCl to prevent cesium loss during sublimation. In combination with cation exchange chromatographic separation, high decontamination factors for Ba, Mo, Sb, and Sn were achieved, thus eliminating the interference of these elements on the measurements of ^{135}Cs and ^{137}Cs by ICP-MS/MS. A detection limit of 9.1×10^{-17} g/g for ^{135}Cs and ^{137}Cs was achieved for 60 g of soil sample. The analytical method has been validated by analysis of reference materials IAEA-375, IAEA-330, and IAEA-385 and successfully applied to determine $^{135}\text{Cs}/^{137}\text{Cs}$ in environmental soil samples from Denmark, Sweden and Ukraine. The $^{135}\text{Cs}/^{137}\text{Cs}$ atomic ratios in Danish soil (2.08–2.68) are much higher than that in the Swedish and Ukrainian soil (0.65–0.71), suggesting their different sources. About 50% Chernobyl derived radiocesium in the Danish soil was estimated. This demonstrated that the developed method for determination of $^{135}\text{Cs}/^{137}\text{Cs}$ can be used to identify the sources of radioactive contamination and estimate their contribution to the total inventory, and a broad application of $^{135}\text{Cs}/^{137}\text{Cs}$ as a tracer in nuclear forensics and environmental studies can be foreseen.

ASSOCIATED CONTENT

Supporting Information

The Supporting Information is available free of charge at <https://pubs.acs.org/doi/10.1021/acs.analchem.0c01153>.

Three figures and two tables of additional experimental details (PDF)

AUTHOR INFORMATION

Corresponding Author

Xiaolin Hou – Department of Environmental Engineering,
Technical University of Denmark, Roskilde DK-4000,
Denmark; orcid.org/0000-0002-4851-4858; Phone: +45-21325129; Email: xiho@dtu.dk

Authors

Liuchao Zhu – Department of Environmental Engineering,
Technical University of Denmark, Roskilde DK-4000, Denmark

Jixin Qiao – Department of Environmental Engineering,
Technical University of Denmark, Roskilde DK-4000,
Denmark; orcid.org/0000-0001-5409-4274

Complete contact information is available at:
<https://pubs.acs.org/doi/10.1021/acs.analchem.0c01153>

Author Contributions

The manuscript was written through contributions of all authors. L.Zhu designed and implemented the experiment and drafted the manuscript, X.Hou designed research plan and supervised the experiment, and X.Hou and J.Qiao revised the manuscript. All authors have given approval to the final version of the manuscript.

Notes

The authors declare no competing financial interest.

REFERENCES

- (1) Russell, B. C.; Croudace, I. W.; Warwick, P. E. *Anal. Chim. Acta* **2015**, 890, 7–20.
- (2) Masson, O.; Bieringer, J.; Brattich, E.; Dalheimer, A.; Estier, S.; Penev, I.; Ringer, W.; Schlosser, C.; Steinkopff, T.; Steinmann, P.; Tositti, L.; Van Beek, P.; Vismes-Ott, A. d. *J. Environ. Radioact.* **2016**, 162–163, 14–22.
- (3) Rostan, J. C.; Juget, J.; Brun, A. M. *Sci. Total Environ.* **1997**, 193, 251–262.
- (4) Kershaw, P.; Baxter, A. *Deep Sea Res., Part II* **1995**, 42, 1413–1448.
- (5) Bailly du Bois, P.; Laguionie, P.; Boust, D.; Korsakissok, I.; Didier, D.; Fievet, B. *J. Environ. Radioact.* **2012**, 114, 2–9.
- (6) Chao, J.; Tseng, C. *Nucl. Instrum. Methods Phys. Res., Sect. A* **1996**, 372, 275–279.
- (7) Delmore, J. E.; Snyder, D. C.; Tranter, T.; Mann, N. R. *J. Environ. Radioact.* **2011**, 102, 1008–1011.
- (8) Lee, T.; Teh-Lung, K.; Hsiao-Ling, L.; Ju-Chin, C. *Geochim. Cosmochim. Acta* **1993**, 57, 3493–3497.
- (9) Snow, M. S.; Snyder, D. C. *J. Environ. Radioact.* **2016**, 151, 258–263.
- (10) Zheng, J.; Tagami, K.; Bu, W.; Uchida, S.; Watanabe, Y.; Kubota, Y.; Fuma, S.; Ihara, S. *Environ. Sci. Technol.* **2014**, 48, 5433–5438.
- (11) Taylor, V. F.; Evans, R. D.; Cornett, R. J. *J. Environ. Radioact.* **2008**, 99, 109–118.
- (12) Yang, G.; Tazoe, H.; Yamada, M. *Anal. Chim. Acta* **2016**, 908, 177–184.
- (13) Zheng, J.; Bu, W.; Tagami, K.; Shikamori, Y.; Nakano, K.; Uchida, S.; Ishii, N. *Anal. Chem.* **2014**, 86, 7103–7110.
- (14) Zheng, J.; Cao, L.; Tagami, K.; Uchida, S. *Anal. Chem.* **2016**, 88, 8772–8779.
- (15) Snow, M. S.; Snyder, D. C.; Delmore, J. E. *Rapid Commun. Mass Spectrom.* **2016**, 30, 523–532.
- (16) Snow, M. S.; Snyder, D. C.; Clark, S. B.; Kelley, M.; Delmore, J. E. *Environ. Sci. Technol.* **2015**, 49, 2741–2748.
- (17) Karam, L. R.; Pibida, L.; McMahon, C. A. *Appl. Radiat. Isot.* **2002**, 56, 369–374.
- (18) Bu, W.; Tang, L.; Liu, X.; Wang, Z.; Fukuda, M.; Zheng, J.; Aono, T.; Hu, S.; Wang, X. *J. Anal. At. Spectrom.* **2019**, 34, 301–309.
- (19) Shibahara, Y.; Kubota, T.; Fujii, T.; Fukutani, S.; Ohta, T.; Takamiya, K.; Okumura, R.; Mizuno, S.; Yamana, H. *J. Nucl. Sci. Technol.* **2014**, 51, 575–579.
- (20) Dunne, J. A.; Richards, D. A.; Chen, H. W. *Talanta* **2017**, 174, 347–356.
- (21) Russell, B. C.; Warwick, P. E.; Croudace, I. W. *Anal. Chem.* **2014**, 86, 11890–11896.
- (22) Russell, B. C.; Croudace, I. W.; Warwick, P. E.; Milton, J. A. *Anal. Chem.* **2014**, 86, 8719–8726.
- (23) Song, M.; Probst, T.; Berryman, N. *Fresenius' J. Anal. Chem.* **2001**, 370, 744–751.

- (24) McKinley, J. P.; Zeissler, C. J.; Zachara, J. M.; Serne, R. J.; Lindstrom, R. M.; Schaef, H. T.; Orr, R. D. *Environ. Sci. Technol.* **2001**, 35, 3433–3441.
- (25) Livens, F. R.; Howe, M. T.; Hemingway, J. D.; Goulding, K. W. T.; Howard, B. J. *Eur. J. Soil Sci.* **1996**, 47, 105–112.
- (26) Yamaguchi, A.; Tanaka, M.; Kurihara, Y.; Takahashi, Y. *J. Radioanal. Nucl. Chem.* **2018**, 317, 545–551.
- (27) Qiao, J.; Hansen, V.; Hou, X.; Aldahan, A.; Possnert, G. *Appl. Radiat. Isot.* **2012**, 70, 1698–1708.
- (28) Lavrov, V. V.; Blagojevic, V.; Koyanagi, G. K.; Orlova, G.; Bohme, D. K. *J. Phys. Chem. A* **2004**, 108, 5610–5624.
- (29) Snow, M. S.; Snyder, D. C.; Mann, N. R.; White, B. M. *Int. J. Mass Spectrom.* **2015**, 381–382, 17–24.
- (30) Nagy, P.; Vajda, N.; Sziklai-Laszlo, I.; Kovacs-Szeles, E.; Simonits, A. *J. Radioanal. Nucl. Chem.* **2014**, 300, 615–627.
- (31) Cao, L.; Zheng, J.; Tsukada, H.; Pan, S.; Wang, Z.; Tagami, K.; Uchida, S. *Talanta* **2016**, 159, 55–63.
- (32) Schuller, P.; Voigt, G.; Handl, J.; Ellies, A.; Oliva, L. *J. Environ. Radioact.* **2002**, 62, 181–193.
- (33) Hou, X.; Fogh, C. L.; Kucera, J.; Andersson, K. G.; Dahlgard, H.; Nielsen, S. P. *Sci. Total Environ.* **2003**, 308, 97–109.