A new role of proficiency testing in nuclear analytical work

Heydorn, Kaj

Published in:
Nuclear Proficiency Testing

Publication date:
2008

Document Version
Early version, also known as pre-print

[Link back to DTU Orbit](#)

Citation (APA):
Heydorn, K. (2008). A new role of proficiency testing in nuclear analytical work. In *Nuclear Proficiency Testing: Proceedings of the 1st International Workshop* (1036 ed., pp. 38-46). American Institute of Physics. AIP Conference Proceedings No. 1036

General rights

Copyright and moral rights for the publications made accessible in the public portal are retained by the authors and/or other copyright owners and it is a condition of accessing publications that users recognise and abide by the legal requirements associated with these rights.

- Users may download and print one copy of any publication from the public portal for the purpose of private study or research.
- You may not further distribute the material or use it for any profit-making activity or commercial gain
- You may freely distribute the URL identifying the publication in the public portal

If you believe that this document breaches copyright please contact us providing details, and we will remove access to the work immediately and investigate your claim.

DTU

A new role of Proficiency Testing in Nuclear Analytical Work

Kaj Heydorn

Technical University of Denmark

In the beginning was the word.....
(Joh. 1:1)

*and the word was **uncertainty***

- *but some people did not like it, and*
- *those who did could not use it, because
we were not sure of its meaning*

International Vocabulary of Basic and General Terms in Metrology

VIM
2nd Edition
1993

Technical University of Denmark

1. axiom

- A result

without statement of uncertainty

is useless

because no valid conclusions can be reached

Guide to the Expression of Uncertainty in Measurement

GUM
1st Edition
1995

Technical University of Denmark

Evaluation of Sampling Uncertainty

- **Type A**
statistical analysis of actual observations
- **Type B**
any other method

2. axiom

- A result

*with an incorrect statement of
uncertainty*

is dangerous

*because erroneous conclusions may
be reached*

Accreditation after ISO 17025*

- Correct measurement results:
 - no significant bias
 - reliable uncertainty

*or ISO 15189

Proficiency Testing

ISO 13528:

**Statistical methods for use in
proficiency testing by interlaboratory
comparisons**

Laboratory Bias

International Vocabulary of Basic and General Terms in Metrology

VIM
3rd Edition
2007

Technical University of Denmark

quantity intended to be measured

- *this definition differs from VIM 2*
- *must include exact specifications*

Definition of the Measurand I

- The *determinand, i.e.*
the chemical species to be determined
- The *specified amount of material*
to which the measurement should apply

Magnitude of a quantity

*expressed as a product of
a number and a unit*

Information on the measurand consisting of

- *a single quantity value , y and*
- *a measurement uncertainty, u*

Definition of the Measurand II

- A result without corresponding definition of the measurand is worthless
- An uncertainty without corresponding specification of the measurand is misleading

Initial proficiency requirements I

- 1) *Definition of the measurand, incl. identification of the determinand and specification of the system*
- 2) *Choice of analytical measurement method and detailing a procedure yielding traceable results*
- 3) *Development of an uncertainty budget, including correct application of counting statistics*
- 4) *Partial verification of uncertainty budget by replicate analyses*

Final proficiency requirements II

- 5) *Choice of sampling strategy and number of samples to be analyzed*
- 6) *Reporting results of analyses corrected for bias and with specified coverage interval.*
- 7) *Final verification of analytical results and their uncertainties by proficiency testing*
- 8) *Calculation of the E_n number*

Our null hypothesis is now that

All reported measurement results for proficiency testing comply with these stipulations, so that

traceability is consistent with the definition of the measurand

all known biases have been corrected for

uncertainties are based on a verified uncertainty budget with a large number of effective degrees of freedom

Bayesian estimate of mean

$$\hat{\mu} = \frac{\sum_i \omega_i \cdot y_i}{\sum_i \omega_i}$$

where $\omega_i = u_i^{-2}$ and $u_\mu = \sqrt{\frac{1}{\sum_i \omega_i}}$

Test statistic

$$T = \sum_i \frac{(y_i - \hat{\mu})^2}{u_i^2}$$

*Chi-square distribution
with $n-1$ degrees of freedom*

E_n numbers

$$E_n^{(i)} = \frac{y_i - \hat{\mu}}{\sqrt{U_i^2 + U_\mu^2}}$$

$$U \sim 2u$$

“bottom up” strategy

- a) *expanded uncertainties, U , are converted to standard uncertainties, $u = U/k$*
- b) *measurement results are ordered according to decreasing u*
- c) *results are added in this order one at a time, and a value of T is calculated*
- d) *if $T \leq \chi^2_{\alpha, m-1}$ the next measurement result is added*

“bottom up” strategy

- e) *if $T > \chi^2_{\alpha, m-1}$ the result with the largest contribution to T is removed*
- f) *after reaching the end of the list go back to c) and add results previously removed*
- g) *repeat c) to f) until there is no change in the selected group of measurement results*
- h) *calculate the reference value μ and its uncertainty u_μ*

“top down” strategy

- a) *apply robust algorithms A and S [3] to the y_i data for estimating μ , respectively their uncertainties U_i for estimating U_μ*
- b) *calculate E_n numbers and disregard all results with $|E_n| > 1$,*
- c) *calculate the weighted mean of the remaining results, using $1/U_i^2$ as weights*

“top down” strategy

d) *calculate its corresponding uncertainty from*

$$U_{\mu}^{-2} = \sum U_i^{-2}$$

e) *repeat b) to d) until there is no change in the selected group of measurement results*

f) *use their weighted mean as reference value and U_{μ} as its expanded uncertainty.*

REIMEP-18 : Uranium isotopic ratios, U in nitric acid

$n(^{234}\text{U})/n(^{238}\text{U})$ certified range ($\pm U=2u_c$): 0.000056541-0.000056623

Results for the $n(^{234}\text{U})/n(^{238}\text{U})$ ratio for REIMEP 18 A

REIMEP-18 : Uranium isotopic ratios, U in nitric acid

$n(^{236}\text{U})/n(^{238}\text{U})$ certified range ($\pm U=2u_c$): 0.00103326-0.00103414

Results for the $n(^{236}\text{U})/n(^{238}\text{U})$ ratio for REIMEP 18 C

Reference values for Uranium isotopic ratios

Strategy	$^{234}\text{U}/^{238}\text{U}$ value $\pm$ Uncertainty (k=2)	Results accepted	$^{236}\text{U}/^{238}\text{U}$ value $\pm$ Uncertainty (k=2)	Results accepted
Bottom up	0.000056581 $\pm$ 31	42	0.00103368 $\pm$ 51	27
Top down	0.000056609 $\pm$ 37	39	0.00103390 $\pm$ 54	25
Combined	0.000056581 $\pm$ 31	42	0.00103368 $\pm$ 51	27

Correct measurement results

- Participants
- Methods

$^{234}\text{U}/^{238}\text{U}$ proficiency data

$^{236}\text{U}/^{238}\text{U}$ proficiency data

Median of E_n numbers for analytical methods

Technique	Sample A Median	Number results	Sample C Median	Number results
Alpha	-0,31	7	0,56	3
HR-ICP	0,20	4	0,59	4
ICP-IDMS	0,36	1	-0,32	1
ICP-QMS	-0,01	6	-0,24	5
MC-ICP	0,14	16	-1,06	16
SF-ICP	0,71	3	3,70	2
TE	0,35	5	1,45	4
TIMS	0,66	19	0,34	20

VIM 3 is a major challenge in

- Our way of interpreting analytical data
Co-operate with the client to define fitness for purpose
- Our way of treating proficiency data
Accreditation authorities beware of the uncertainty of assigned values

Technical University of Denmark

Question from the audience:

- Would the proposed method lead to substantially different E_n numbers?
- Not for the particular example used here, but for the example used in ISO 13528 the drastic reduction of the uncertainty of the reference value greatly increases the detection capability for too optimistic reported uncertainties.

Comparison with certified values

Value	$^{234}\text{U}/^{238}\text{U}$ value $\pm$ Uncertainty (k=2)	Results accepted	$^{236}\text{U}/^{238}\text{U}$ value $\pm$ Uncertainty (k=2)	Results accepted
Reference	0.000056581 $\pm$ 31	42	0.00103368 $\pm$ 51	27
Certified	0.000056582 $\pm$ 41	Sample A	0.00103370 $\pm$ 44	Sample C

Reference values* for Pb in IMEP-9

Method	Reference value $\pm$ Uncertainty	Number of accepted results	Comment
Synthesis of Precision	617.7 ± 2.7	60	Recommended
E_n numbers	614.1 ± 3.2	59	Alternative
Robust average	605 ± 26	181	ISO 13528 (2005)
ICP-MS	623 ± 13	6	Certified value

*in units of 10^{-10} mol/L