



Quantification of Tissue Trauma following Insulin Pen Needle Insertions in Skin

Jensen, Casper Bo; Larsen, Rasmus; Vestergaard, Jacob Schack; Conradsen, Knut; Kildegaard, Jonas; Præstmark, Kezia Ann

Publication date:
2013

Document Version
Publisher's PDF, also known as Version of record

[Link back to DTU Orbit](#)

Citation (APA):
Jensen, C. B., Larsen, R., Vestergaard, J. S., Conradsen, K., Kildegaard, J., & Præstmark, K. A. (2013). *Quantification of Tissue Trauma following Insulin Pen Needle Insertions in Skin*. Abstract from 13th Annual Diabetes Technology Meeting, San Francisco, California, United States.

General rights

Copyright and moral rights for the publications made accessible in the public portal are retained by the authors and/or other copyright owners and it is a condition of accessing publications that users recognise and abide by the legal requirements associated with these rights.

- Users may download and print one copy of any publication from the public portal for the purpose of private study or research.
- You may not further distribute the material or use it for any profit-making activity or commercial gain
- You may freely distribute the URL identifying the publication in the public portal

If you believe that this document breaches copyright please contact us providing details, and we will remove access to the work immediately and investigate your claim.

Quantification of Tissue Trauma following Insulin Pen Needle Insertions in Skin

Casper Bo Jensen, BSc.Eng; Rasmus Larsen, Professor, PhD; Jacob Vestergaard, MSc.Eng; Knut Conradsen, Professor; Jonas Kildegaard, PhD; Kezia Ann Præstmark, MSc.Eng

Technical University of Denmark, Novo Nordisk A/S
Kgs. Lyngby, Denmark
csbj@novonordisk.com

Objective:

Within the field of pen needle development, most research on needle design revolves around mechanical tensile testing and patient statements. Only little has been published on the actual biological skin response to needle insertions. The objective of this study was to develop a computational method to quantify tissue trauma based on skin bleeding and immune response.

Method:

Two common sized pen needles of 28G (0.36mm) and 32G (0.23mm) were inserted into skin of sedated LYD pigs prior to termination. Four pigs were included and a total of 32 randomized needle insertions were conducted. The affected tissue was removed and fixated in formalin following tissue preparation for histology. Standard immunohistochemical staining procedure was applied with CD-45 and anti-hemoglobin primary antibodies to stain immune cells and red blood cells, respectively. The stained tissue slides were subsequently digitized using 200X magnification. Based on thresholding, morphological masks and blob detection, segmentation of the histology was performed to locate tissue bleeding and immune response. Image-to-image registration was used on images originating from the same tissue, and a quantitative measure of tissue trauma was obtained for each needle insertion.

Result:

Bleeding and immune response were seen for all tested needles. Positive correlation was seen between the needle diameter and the size of the bleeding. The quantitative measure reveal a trend that tissue trauma decreases with decreasing needle diameter.

Conclusion:

A computational and quantitative method has been developed to assess tissue trauma following insulin pen needle insertions. Application of the method is tested by conduction of a needle diameter study. The obtained quantitative measures of tissue trauma correlate positively to needle diameter.