



Encoding of inductively measured k-space trajectories in MR raw data

Pedersen, Jan Ole; Hanson, Christian G.; Xue, Rong; Hanson, Lars G.

Publication date:
2017

Document Version
Publisher's PDF, also known as Version of record

[Link back to DTU Orbit](#)

Citation (APA):
Pedersen, J. O., Hanson, C. G., Xue, R., & Hanson, L. G. (2017). *Encoding of inductively measured k-space trajectories in MR raw data*. Abstract from ISMRM 25th Annual Meeting & Exhibition, Honolulu, Hawaii, United States.

General rights

Copyright and moral rights for the publications made accessible in the public portal are retained by the authors and/or other copyright owners and it is a condition of accessing publications that users recognise and abide by the legal requirements associated with these rights.

- Users may download and print one copy of any publication from the public portal for the purpose of private study or research.
- You may not further distribute the material or use it for any profit-making activity or commercial gain
- You may freely distribute the URL identifying the publication in the public portal

If you believe that this document breaches copyright please contact us providing details, and we will remove access to the work immediately and investigate your claim.

Encoding of inductively measured k-space trajectories in MR raw data

Jan Ole Pedersen^{1,2,3}, Christian G. Hanson, Rong Xue⁴ and Lars G. Hanson^{1,2}

¹Danish Research Centre for Magnetic Resonance, Centre for Functional and Diagnostic Imaging and Research, Copenhagen University Hospital, Hvidovre, Denmark.

²Center for Magnetic Resonance, DTU Elektro, Technical University of Denmark, Kgs Lyngby, Denmark.

³Sino-Danish Center for Education and Research, Aarhus, Denmark

⁴State Key Laboratory of Brain and Cognitive Science, Beijing MR Center for Brain Research, Institute of Biophysics, Chinese Academy of Sciences, Beijing, China

Synopsis

For reconstruction of MRI from raw data, the k-space trajectory is needed. We propose to measure this simultaneously with the MRI signal using novel hardware and the scanner itself. As a first step, we demonstrate real-time processing of a non-MR signal from a gradient pickup coil during scanning, modulation of this signal to a frequency within the MR acquisition bandwidth, and signal extraction from MR raw data. This principle is applied to measure k-space trajectories inductively and perform image reconstruction based on this. The resulting images have comparable quality to images reconstructed using nominal k-space trajectories.

Introduction

By measuring k-space trajectories, $k(t)$, MRI reconstruction can be performed without knowledge of sequence details and hardware imperfections. Using the scanner itself for measuring $k(t)$ alleviates timing challenges arising when using external equipment. This is exploited for MR-based gradient measurement[1]. The time derivative, $\dot{k}(t)$, of the gradient field $G(t)$ can alternatively be measured inductively by a pickup coil. For a gradient echo sequence

$$k(t) = \frac{\gamma}{2\pi} \int_0^t G(\tau) d\tau ,$$

which implies

$$k(t) \propto \int_0^t \int_{t_{co}}^{\tau_1} \epsilon(\tau_2) d\tau_2 d\tau_1 ,$$

where $t=0$ is the center of the relevant excitation pulse, and t_{G0} a time point where the gradients are zero. The signal/trajectory synchronization challenge can be addressed by using the scanner itself for the acquisition, which also limits hardware requirements. Non-MR signals like $k(t)$ can be encoded in the MR data using amplitude modulation (AM) to a frequency in the oversampled range (sampled by the scanner, but not in FOV) without interference with MRI [2]. The encoded signal can be extracted by analyzing raw data, possibly while MRI is performed.

Method

We have developed circuitry for sampling and real-time processing of non-MR signals. The processed signals are amplitude modulated onto a carrier detectable by an MRI scanner. The circuitry has three analog input channels (20 kHz bandwidth) and also measures an RF power correlate for the 50MHz – 1GHz range. The received digitized signals are processed by a Field Programmable Gate Array (FPGA), that controls an RF output channel where amplitude, phase, and frequency of up to 130MHz can be manipulated. Two-way communication for signals and hardware control is supported through a USB connection to a computer.

For proof of concept, the circuitry was used to acquire the k-space trajectory for an EPI sequence using a 3T Philips Achieva MRI system (208x127 receive matrix, 32 slices, TE/TR:45/3000ms, 16% ramp sampling, echospacing: 0.62ms). See figure “setup” for experimental setup. Signal received from a pickup coil placed in the gradient field was integrated twice by the FPGA to obtain $k(t)$ with minimal delay (20us), and send wirelessly to the scanner via its 32 channel head receive coil. To facilitate detection, a central sample in the RF pulse wave form was nulled through pulse programming (see fig.RFpulse). This special RF signature was detected by the power detection of the circuitry, and triggered nulling of t and t_{G0} in the integrals. The k-space contributions from the readout and phase encoding gradient were determined separately by turning off the gradients in the other directions. As a

simple way to compensate the 20us signal delay during FPGA processing, the last 8 MR sample points from each readout period were not used in the image reconstruction. Inaccurate determination of the RF pulse center was corrected, before averaging $k(t)$ across measurements performed over 5 volumes of 32 slices extracted from a single receive channel. Lowpass filtering was performed to avoid noise contributions affecting $k(t)$ estimates. The known FOV was used to determine the otherwise unknown scaling between the double integral and $k(t)$.

Results

Using non-uniform Fourier reconstruction (NUFFT) [3], MRI data acquired in a separate EPI scan of a phantom were reconstructed using measured and nominal k-space trajectories (figure “recon”).

Discussion and Conclusion


Through the use of developed circuitry for non-MR signal sampling, processing and AM modulation, we have demonstrated measurement of k-space trajectories using the MR scanner itself. The method is flexible and requires limited hardware that is of general interest for real-time applications. 20us of each sampling period was discarded which has little influence on reconstructed images, since the missing samples are at the outer regions of sampled k-space. Ramp-sampling reduces the corresponding, limited loss.

RF waveforms were modified to trigger resetting of integrators, but since the scaled RF power is measured and processed real-time, a power calibration can alternatively be used to scale power-integrals and hence calculate nominal tip angles for resonant pulses. This would also facilitate k-space trajectory measurements for sequences with multiple RF pulses per excitation, e.g. spin-echo.

The k-space signal was passed wirelessly to the scanner, thus, simultaneous MRI acquisition was unattractive due to possible interference. Directly connecting the developed circuitry to a receive channel of the scanner alleviate this, and likely also improve SNR. For an EPI sequence the method


performed comparably to using nominal k-space trajectory, but advantages may likely result for challenging trajectories such as spiral-in. No attempts to limit RF noise was made for this prototype hardware implementation, which was reflected in non-white RF noise that was here reduced by averaging and limited filtering. The quality of images reconstructed from nominal and measured gradients were nevertheless of comparable quality, and the results are encouraging.

1. Duyn J et. al. Simple Correction Method for k-Space Trajectory Deviations in MRI. *J. Magn. Reson.*1998;132:150-153
- 2.Hanson et al. *Encoding of Electrophysiology and Other Signals in MR Image. JMRI.* 25.2007;25:1059-1066
3. Fessler A. Nonuniform Fast Fourier Transforms Using Min-Max Interpolation. *IEEEET-SP* 2003;51(2):560-74




(RF pulse)

Waveform of excitation pulses as measured by the developed circuitry, and send through USB connection to a computer. The RF pulse used for MRI data acquisition (red) was modified through pulse programming allowing the circuitry to determine the center of the pulse (blue). When the drop in the RF power of the pulse was detected, a nulling of the integrators used for determining $k(t)$ was triggered.



(recon)

NUFFT reconstructed EPI images using acquired (left) and nominal (right) k-space trajectories. Similar image quality was obtained, demonstrating the feasibility of acquiring k-space trajectories using developed hardware and the scanner itself.



(setup)

Experimental setup used for acquiring k-space trajectories. A pickup coil is placed in the gradient field of the MR scanner, and signal generated by the switching gradients is received by the FPGA. A double integration of the signal is performed to obtain $k(t)$, which is transmitted to the head receive coil using a quarter-wavelength dipole antenna. A similar antenna is connected to measure the RF power transmitted by the scanner. When the RF signature shown in fig “RF pulse” is detected, the integrators are nulled.