



## Liquid-State Polarization of 30% through Photo-Induced Non-Persistent Radicals on $^{13}\text{C}$ Pyruvic Acid

Capozzi, Andrea; Karlsson, Magnus; Petersen, Jan Raagaard; Lerche, Mathilde Hauge; Ardenkjær-Larsen, Jan Henrik

*Publication date:*  
2018

*Document Version*  
Publisher's PDF, also known as Version of record

[Link back to DTU Orbit](#)

*Citation (APA):*  
Capozzi, A., Karlsson, M., Petersen, J. R., Lerche, M. H., & Ardenkjær-Larsen, J. H. (2018). *Liquid-State Polarization of 30% through Photo-Induced Non-Persistent Radicals on  $^{13}\text{C}$  Pyruvic Acid*. Abstract from 59th Experimental Nuclear Magnetic Resonance Conference, Orlando, Florida, United States.

---

### General rights

Copyright and moral rights for the publications made accessible in the public portal are retained by the authors and/or other copyright owners and it is a condition of accessing publications that users recognise and abide by the legal requirements associated with these rights.


- Users may download and print one copy of any publication from the public portal for the purpose of private study or research.
- You may not further distribute the material or use it for any profit-making activity or commercial gain
- You may freely distribute the URL identifying the publication in the public portal

If you believe that this document breaches copyright please contact us providing details, and we will remove access to the work immediately and investigate your claim.

# Liquid-State Polarization of 30% through Photo-Induced Non-Persistent Radicals on $^{13}\text{C}$ Pyruvic Acid

A. Capozzi<sup>1</sup>, M. Karlsson<sup>1</sup>, J. R. Petersen<sup>1</sup>, M. H. Lerche<sup>1</sup>, J. H. Ardenkjær-Larsen<sup>1,2</sup>

<sup>1</sup> Department of Electrical Engineering, Technical University of Denmark, 2800 Kgs. Lyngby, Denmark; <sup>2</sup> GE Healthcare, Denmark




**Figure 1** UV-irradiation stand (A) and its sagittal section view (B). PA UV-vis absorption superimposed to the UV-source irradiation range (C). Radical generation time evolution on a 4  $\mu\text{l}$  frozen pellet of PA:H<sub>2</sub>O (D).

the photo-induced paramagnetic centers inside the DNP samples, we provide a robust protocol to enhance the efficiency and performances of the UV-radical technique.

A new sample irradiation setup was designed to work in a moisture free environment (see Fig 1). The set-up provides a fast generation of the UV-induced radicals thanks to a high power (20 W/cm<sup>2</sup>) broad-band UV-source (Dymax BlueWave 75) shining light on the tail of a quartz cold finger placed inside a reflecting aluminum chamber. On a sample containing a mixture of PA:H<sub>2</sub>O 1:1 (v/v) we were able to induce more than 50 mM of radicals in about 5 min, thus doubling the maximum yield and reducing the irradiation time by a factor of 30 compared to the current state of the art [5, 6].

The distribution of the paramagnetic centers inside the DNP samples was investigated via ESR by changing the radical precursor (PA) concentration, in a solvent mixture of glycerol:water 60:40 (GW64), and the dimension of the frozen pellets. Taking into account the strong absorption of the UV-light while travelling through the spherical frozen samples, a radicals' distribution profile with exponential dumping ( $\sim e^{-x/d}$ ) from the surface to the core of the pellet was assumed and verified experimentally (see Fig 2). The attenuation parameter  $d$  increased from about 0.1 mm to 3 mm by decreasing the PA concentration from 90% to 10% of the total sample volume.

DNP wise the best compromise between PA concentration and homogeneity of the radicals' distribution was found to be a mixture of PA:GW64 1:1 (v/v). In this sample  $^{13}\text{C}$  nuclei polarized up to 30% in about 1 h working at 6.7 T and 1.1 K. These values represent a clear improvement compared to the current state of the art making the UV-radical technique useful for the most recent demanding metabolic imaging applications.



**Figure 2** Photo-induced radicals ESR signal as a function of the PA concentration and frozen pellets radius (A). Calculated attenuation parameter as a function of the PA concentration (B). Simulation of radicals' distribution profile according to the measured  $d$  values (C).

## References

- [1] J. H. Ardenkjær-Larsen et al., PNAS **100**, 18 (2003); [2] S. Nelson et al., Sci. Transl. Med. **198**, 5 (2013); [3] T. R. Eichhorn et al., PNAS **110**, 45 (2013); [4] A. Capozzi et al., JPCC, **119**, 39 (2015); [5] A. Capozzi et al., Nature Communication **8**, 15757 (2017); [6] H. A. I. Yoshihara et al., PCCP **18**, 18 (2016); [7] T. B. Rodrigues et al., Nature Medicine **20**, 1 (2014).