



## Integrated microstructures on chip for ultrasensitive pathogen detection

Ngo Anh, Tien; Kant, Krishna; Huynh, Van Ngoc; Bang, Dang Duong; Wolff, Anders

*Publication date:*  
2018

*Document Version*  
Publisher's PDF, also known as Version of record

[Link back to DTU Orbit](#)

*Citation (APA):*

Ngo Anh, T., Kant, K., Huynh, V. N., Bang, D. D., & Wolff, A. (2018). *Integrated microstructures on chip for ultrasensitive pathogen detection*. Abstract from 44th International conference on Micro and Nano Engineering, Copenhagen, Denmark.

---

### General rights

Copyright and moral rights for the publications made accessible in the public portal are retained by the authors and/or other copyright owners and it is a condition of accessing publications that users recognise and abide by the legal requirements associated with these rights.

- Users may download and print one copy of any publication from the public portal for the purpose of private study or research.
- You may not further distribute the material or use it for any profit-making activity or commercial gain
- You may freely distribute the URL identifying the publication in the public portal

If you believe that this document breaches copyright please contact us providing details, and we will remove access to the work immediately and investigate your claim.

## Integrated microstructures on chip for ultrasensitive pathogen detection

Tien Anh Ngo<sup>a\*</sup>, Krishna Kant<sup>b\*</sup>, Ngoc Huynh Van<sup>b</sup>, Dang Duong Bang<sup>a</sup>, Anders Wolff<sup>b</sup>

<sup>a</sup> Laboratory of Applied Micro and Nanotechnology (LAMINATE), National Food Institute, Technical University of Denmark, 2800 Kgs. Lyngby, Denmark

<sup>b</sup> Department of Micro- and Nanotechnology, Technical University of Denmark, 2800 Kgs. Lyngby, Denmark

E-mail: [tinan@food.dtu.dk](mailto:tinan@food.dtu.dk)

(\* T. A. Ngo and K. Kant contributed equally, and both should be considered as first authors)

Keywords: Solid-phase PCR, SAF microstructure, pathogen detection, lab-on-a-chip

### Abstract

Nucleic acid amplification based polymerase chain reaction (PCR) technique is popular and gold standard method in molecular biology to amplify a single copy or a few copies of target DNA. It is widely used in molecular diagnostics to detect the wide range of pathogens. To address the huge demand of the molecular diagnostics the development of new integrated rapid and sensitive method for point of care devices is needed. Earlier pathogens detection based on nucleic acid amplification techniques have been more and more attractive and successful. Since it is easy to transfer nucleic acid amplification from liquid to solid phase into lab-on-chip (LOC) devices. Concerning the issues of molecular diagnostics, a solid phase PCR (SP-PCR) technique has been developed and become popular for molecular diagnostics[1–4]. Recently, we have successfully developed a LOC platform based on a combination of SP-PCR with supercritical angle fluorescence (SAF) microlens array embedded in a microchip for rapid foodborne pathogen detection[5]. By integrating the SAF microstructure array into a polymeric chip, the sensitivity of the test was increased to 46 folds compared to a conventional array without SAF[6]. Hence, the SAF microstructures array embedded microfluidic chip is a good strategy to develop a portable device for onsite, online, and rapid pathogen detection. However, to enhance the sensitivity of detection and multiplexing of samples, it is required to increase the number of SAF microstructures arrays on a chip. In this report, we addressed this challenge by fabricating small size of SAF microlens array in the microfluidic chamber of a disposable polymer (Cyclic olefin copolymers) chip. The limitation of detection obtained from SAF microstructures array will determine the sensitivity of the developed system. The advantages of increasing the number of sample site by increasing number of SAF and advantages of reducing the size of SAF will be discussed.

- [1] J. Hoffmann, M. Trotter, F. von Stetten, R. Zengerle, G. Roth, Solid-phase PCR in a picowell array for immobilizing and arraying 100 000 PCR products to a microscope slide, *Lab Chip*. 12 (2012) 3049. doi:10.1039/c2lc40534b.
- [2] Y. Sun, R. Dhumpa, D.D. Bang, J. Høgberg, K. Handberg, A. Wolff, A lab-on-a-chip device for rapid identification of avian influenza viral RNA by solid-phase PCR, *Lab Chip*. 11 (2011) 1457.

doi:10.1039/c0lc00528b.

- [3] Y. Sun, R. Dhumpa, D.D. Bang, K. Handberg, A. Wolff, DNA microarray-based solid-phase RT-PCR for rapid detection and identification of influenza virus type A and subtypes H5 and H7, *Diagn. Microbiol. Infect. Dis.* 69 (2011) 432–439. doi:10.1016/j.diagmicrobio.2010.11.008.
- [4] W.H. Chin, Y. Sun, J. Høgberg, T.Q. Hung, A. Wolff, Solid-phase PCR for rapid multiplex detection of *Salmonella* spp. at the subspecies level, with amplification efficiency comparable to conventional PCR, (2017). doi:10.1007/s00216-017-0216-y.
- [5] T. Quang, W. Hoe, Y. Sun, A. Wol, D. Duong, A novel lab-on-chip platform with integrated solid phase PCR and Supercritical Angle Fluorescence (SAF) microlens array for highly sensitive and multiplexed pathogen detection, *Biosens. Bioelectron.* 90 (2017) 217–223. doi:10.1016/j.bios.2016.11.028.
- [6] T.Q. Hung, Y. Sun, C.E. Poulsen, T. Linh-Quyen, W.H. Chin, D.D. Bang, A. Wolff, Miniaturization of a micro-optics array for highly sensitive and parallel detection on an injection moulded lab-on-a-chip, *Lab Chip*. 15 (2015) 2445–2451. doi:10.1039/C5LC00176E.