



Comparison campaign of XCT systems using machined standards representative of additively manufactured parts

Obaton, Anne-Françoise; Quagliotti, Danilo; Yardin, Catherine; Liltorp, Klaus; De Chiffre, Leonardo

Publication date:
2022

Document Version
Publisher's PDF, also known as Version of record

[Link back to DTU Orbit](#)

Citation (APA):
Obaton, A.-F., Quagliotti, D., Yardin, C., Liltorp, K., & De Chiffre, L. (2022). *Comparison campaign of XCT systems using machined standards representative of additively manufactured parts*. Paper presented at 11th Conference on Industrial Computed Tomography.

General rights

Copyright and moral rights for the publications made accessible in the public portal are retained by the authors and/or other copyright owners and it is a condition of accessing publications that users recognise and abide by the legal requirements associated with these rights.

- Users may download and print one copy of any publication from the public portal for the purpose of private study or research.
- You may not further distribute the material or use it for any profit-making activity or commercial gain
- You may freely distribute the URL identifying the publication in the public portal

If you believe that this document breaches copyright please contact us providing details, and we will remove access to the work immediately and investigate your claim.

Comparison campaign of XCT systems using machined standards representative of additively manufactured parts

Anne-Françoise Obaton¹, Danilo Quagliotti², Catherine Yardin¹, Klaus Liltorp², Leonardo De Chiffre²

¹LNE, 1 rue Gaston Boissier, 75015 Paris, France, e-mail: anne-francoise.obaton@lne.fr, catherine.yardin@lne.fr,

²DTU Mekanik, Produktionstorvet, 2800 Kgs. Lyngby, Denmark, e-mail: danqua@mek.dtu.dk, klli@mek.dtu.dk, ldch@mek.dtu.dk

Abstract

Additive manufacturing enables the production of complex geometries, both internally and externally. This poses a challenge to quality control. Non-destructive volumetric testing methods are required. Among these methods, X-ray computed tomography is currently the most promising. It enables not only to detect inner and outer defects but also to perform dimensional measurements. However, the method lacks of metrological tracability. To fill this gap, an EMPIR project 17IND08 AdvanCT involving several European countries was proposed. In this frame, DTU Mekanik and LNE have conducted a comparison campaign of XCT systems using machined material measures representative of additively manufactured parts. The purpose of the comparison was to investigate the performances of industrial X-ray computed tomography with respect to dimensional measurements on external but also on internal features.

Keywords: Comparison campaign, round robin test, X-ray computed tomography (XCT), dimensional metrology, material measure, additive manufacturing (AM)

1 Introduction

A growing number of industrial sectors are considering the potential of additive manufacturing (AM) as an asset to improve their production capabilities. Indeed, AM enables the fabrication of very complex shapes and inner cavities that cannot be manufactured with conventional techniques. Nevertheless, AM has been shown to produce specific types of defects in as-built parts [1]. Thus, before they can be used in safety-critical areas, such as aeronautics, medical sector, etc, the integrity of the AM parts and their compliance with specifications must be ensured to be certified. This requires stringent quality controls to be implemented. With such complex geometries, inner cavities, and high surface roughness, it is particularly challenging to characterise the parts' geometry, the dimensional accuracy and the inside of the parts in terms of flaws, without damaging them. The characterisation methods need to be non-destructive as well as volumetric and unaffected by surface conditions. X-ray computed tomography (XCT) meets these requirements and is presently the most suitable method [2]. However, the traceability and the evaluation of the uncertainty for XCT dimensional measurement are still critical tasks, above all in the case of internal features and structures.

The overall objective of the EMPIR project 17IND08 AdvanCT for "Advanced Computed Tomography for dimensional and surface measurements in industry" [3] was to improve the metrological quality of dimensional XCT measurements. In the frame of this project, DTU Mekanik and LNE designed and manufactured a number of material measures (also called transfer standards in the following) to lead a comparison campaign of XCT systems. As both laboratories are involved in quality control of AM parts, one decided to develop transfer standards representative of AM parts in terms of flaws and material, meeting the needs of the industry. These transfer standards as well as the chosen measurands were fully described in the previous iCT2020 paper [4].

Several comparison campaigns of XCT systems have been conducted and the results published. Their aims as well as the investigated transfer standards were different. A. du Plessis et al., in 2019 [5], aimed to advance the efficient use of micro-XCT facilities for process optimization and quality inspections for AM products in order to provide confidence in their use but also to indicate the need for further development of transfer standards, protocols and image analysis workflows for quantitative assessment, especially for direct and quantitative comparisons between different laboratories. Whereas, the different comparison campaigns presented by M. Bartscher et al., in 2018 [6], focused on dimensional measurements. However, none of these interlaboratory comparisons aimed to investigate specifically the performance of industrial XCT with respect to dimensional measurements on external but also on internal features.

After a summary of the previous iCT2020 paper, with a brief review of the transfer standards and of the measurands, this paper will present the metrological calibration of the transfer standards and the statistical analysis of the results of the comparison of XCT systems. The XCT campaign involved several laboratories and companies from all over Europe: Zeiss (DE), Novo Nordisk (DK), DTI (DK), NPL (UK), Messtronik (DE), CEA-List (FR), MTC (UK) and the University of Nottingham (UK).

2 Summary of iCT2020 paper

A list of XCT existing material measures was presented in the iCT2020 paper [4]. Then the design (shape, dimensions, and materials), aim and fabrication of the proposed transfer standards were detailed, and the measurands specified.



2.1 Physical standards for the comparison campaign of XCT systems

Three material measures, such as the ones presented in Figure 1, designed with the purposes of being representative of the AM parts and of evaluating the capability of XCT to detect internal features in mono- or multi-materials, were machined. In order to be representative of AM parts these standards contain external but also internal spherical calottes, of different diameters, simulating internal AM flaws, internal hollow cylinders simulating internal structures and external grooves simulating cracks. In addition, they are in acrylonitrile butadiene styrene (ABS), stainless steel and aluminium, materials used a lot in the AM industry [4]. Internal plugs allow for evaluating the capability of XCT to detect internal features in mono- or multi-materials, and, as they are removable, their metrological qualification with surface reference methods such as coordinate measuring machine (CMM) is possible.

2.2 Measurands for the comparison campaign of XCT systems

The chosen measurands are presented in Figure 1 and described in Table 1.




Figure 1: Transfer standard and measurand presentation.

Table 1: Measurand definition (cf. Figure 1).

Identification	Description
M1	Distance between the centres of the spheres which fit (GG) calotte S1-6 (calotte 6 on step 1) and calotte S1-0 (calotte 0 on step 5)
M2	Width of the external insert groove 1 on step 3 i.e. distance between Point 1 and Point 2. Point 1 corresponds to the intersection of the x-axis and edge 1: intersection of planes (GG) XY1 and YZ1. Point 2 corresponds to the intersection of the x-axis and edge 2: intersection of planes (GG) XY2 and YZ2. The coordinate system is defined as follows: 0 is the centre of the hole used for the screw holding the external insert, the y-axis is parallel to edge 1, the x-axis is in plane XY1 and the z-axis perpendicular to XY1 plane.
M3-1 to 4	Diameters of the 0.6 mm bore at 4 heights on the internal insert on step 4. These diameters correspond to the diameters of the circles fitting (GG) the bore at Z=-0.1 mm (M31), Z=-0.58 mm (M32), Z=-0.82 mm (M33) and Z=-1.3 mm (M34)
M4	Diameter of the sphere which fits (GG) the calotte P3 on step 4 internal insert

3 Metrological calibration of the transfer standards

The reference measurements of M1, M3 and M4 were performed at DTU with a tactile CMM, and the measurements of M2 were performed at LNE with an optical CMM. They were carried out either before or after circulation of the standards.

The DTU machine is a mechanical CMM equipped with a static probe placed in a temperature controlled room ($T = 20\text{ °C} \pm 1\text{ °C}$). This CMM is of the type Zeiss PRISMO (**Erreur ! Source du renvoi introuvable.** 2) with a VAST XT-GD1 measuring sensor. The calibration equipment was used with CALYPSO 6.8 software.

The measurements to evaluate M1 were performed with a $\varnothing 0.8\text{ mm}$ ruby sphere, a $\varnothing 0.6\text{ mm}$ tungsten shaft, a 8 mm shaft length probe and a measuring force of 50 mN. The position of each calotte was measured by scanning 4 circles picking up 32 points on each circle at 0.125 mm/s. The measurements to evaluate M3 and M4 were performed with a $\varnothing 0.3\text{ mm}$ ruby sphere, $\varnothing 0.2\text{ mm}$ tungsten shaft, 3 mm shaft length probe and a measuring force of 50 mN. The bore circles related to M3 were measured by scanning 2 circles of 32 points at 0.125 mm/s. The calotte related to M4 was measured by scanning 2 circles of 32 points at 0.125 mm/s.




Figure 2: CMM of the type ZEISS PRISMO placed at DTU.

The LNE machine used was an optical CMM. It is of the type MICROVU EXCEL502UM (Figure 3) equipped with an optical zoom from x25 to x600 and digital zoom up to x1800. The CMM is placed in a temperature controlled room ($T = 20\text{ }^{\circ}\text{C} \pm 0.5\text{ }^{\circ}\text{C}$) with a hygrometry of $55\% \pm 10\%$. The calibration equipment is equipped with INSPEC 2.97.2 software.

The groove widths of the external inserts were measured in episcopy at x300 magnification on the upper edge of the groove at mid-length.



Figure 3: CMM of the type MICROVU EXCEL502UM placed at LNE.

Traceability of the reference measurements was documented by the general condition of instruments and repetition of the measurements, which showed good consistency.

4 Comparison campaign of XCT

Ten laboratories and companies, listed in Table 2, participated to the XCT measurement's campaign. These measurements were performed with different XCT systems, and post-processed by different software tools, also given in Table 2.

A technical XCT measurement protocol was drafted and sent to all participants beforehand. However, no strict instructions were given for the scanning such as the XCT parameters and the number of scans. The transfer standards were to be measured by selecting all XCT parameters according to the experience of the participant to give the best of his XCT system in term of dimensional measurement uncertainty in order to be placed in the conditions of XCT service providers. In addition to the protocol, a template table, specifying the information to give on the XCT systems, XCT scans, etc, was also sent.

The comparison was based on the circulation of three transfer standards, one in ABS (ABS1 then ABS2—severe cracks in ABS1 forced to switch to ABS2), one in aluminium (A11) and one in stainless steel (SS1). According to the available energies of the systems, the stainless steel standard was measured with a few systems and according to the chosen XCT parameters, some measurands could not be determined. Those, which were measured, are listed in Table 2.

Table 2: Participants, countries, XCT systems and post-processed softwares, and evaluated measurands according to the transfer standards of the comparison campaign of XCT.

Participant	Country	XCT system and post-processed software	ABS1 Evaluated measurands	ABS2 Evaluated measurands	A11 Evaluated measurands	SS1 Evaluated measurands
Zeiss	Germany	Zeiss Metrotom 800 (max 225 kV) with VGStudioMax	not scanned	not scanned	M1, M2, M3, M4	not scanned
Novo Nordisk	Denmark	Zeiss, Metrotom 800 (max 225 kV) with VGStudioMax	M1, M2, M3, M4	M1	M1, M2, M3, M4	not scanned
		Zeiss Metrotom 1500 with VGStudioMax	M1, M2, M3, M4	not scanned	M1, M2, M3, M4	M1, M2, M3, M4
DTI	Denmark	Zeiss Metrotom 800 (max 130 kV) with Calypso	not scanned	M1, M2, M3	M1, M2, M3	not penetrated
		Zeiss Metrotom 1500 (max 180 kV) with Calypso	not scanned	M1, M3	M1, M2	not penetrated
NPL	UK	Nikon XT H 225M (max 225 kV) with VGStudioMax	not scanned	M1, M2, M3, M4	M1, M2, M3, M4	not penetrated
Messtronik	Germany	Werth TomoScope XL (max 300 kV) with Winwerth	not scanned	M1, M2, M3, M4	M1, M2, M3, M4	M1, M2, M3, M4
CEA	France	Home made (max 225 kV) with VGStudioMax	not scanned	M1, M2, M3, M4	M1, M2, M3	M1, M2
MTC	UK	Diondo d2 (max 240 kV) with VGStudioMax	not scanned	M2, M3, M4	M1, M2, M3, M4	not penetrated
UNOTT	UK	Nikon MCT225 (max 225 kV) with VGStudioMax	not scanned	M1, M2, M3, M4	M1, M2, M3, M4	Not penetrated

5 Statistical analysis

Firstly, the results were analysed in accordance with ISO 5725 parts 2 and 4 [7, 8]. These ISO standards provide practical guidance for evaluating the accuracy of a measurement method. Although the accuracy is a qualitative characteristic, in these ISO standards, and generally in the validation method field, it is designated by the two properties *precision* [7] and *trueness* [8]. According to these documents, trueness refers to the bias between the arithmetic mean of a large number of test results and a reference value, whereas precision evaluates the closeness of agreement between test results.

Secondly, the z scores and the zeta scores, defined in ISO 13528 [9], were calculated for each laboratory. Using the bias of one laboratory, these criteria allow to evaluate the overall performance of the laboratory. Moreover, zeta scores may reveal underestimated uncertainties.

5.1 Application of the ISO 5725-parts 2 and 4

ISO 5725-part 2 is dedicated to *precision* and ISO 5725 part 4 to *trueness*.

5.1.1 Precision: repeatability and reproducibility

Precision is quantified by two uncertainty contributors, viz. *repeatability* and *reproducibility*. Repeatability represents the dispersion of the results obtained under unchanged measurement conditions. Reproducibility represents the dispersion of the results where at least one change has been introduced [10].

The variances of the results were initially analysed with a Cochran test. If the test failed, the individual results associated to the maximum variance were successively analysed with a Grubbs test for the detection of outliers in the data set [7].

After the outlier's removal, the repeatability S_r was evaluated as weighted mean of the standard deviation of each laboratory's results S_{ri} , according to Eq. 1

$$S_r = \left(\frac{1}{\sum dof_i} \sum dof_i \times S_{ri}^2 \right)^{1/2} \quad (1)$$

where dof_i represents the number of degree of freedom associated to the standard deviation S_{ri} .

Eventually, an ANOVA test (ANalysis Of VAriance) was performed on the participant's mean values to highlight systematic differences among laboratories. Measurement results affected by a systematic behaviour could be accounted for as *laboratory effect* S_L^2 evaluated according to Eq. 2

$$S_L^2 = S_d^2 - S_r^2 \quad (2)$$

where S_d^2 corresponds to n times the variance of the mean— n is the number of repeated measurements of each laboratory.

Finally, the reproducibility standard deviation S_R was obtained using Eq. 3

$$S_R^2 = S_r^2 + S_L^2 \quad (3)$$

5.1.2 Bias of the method

Trueness is normally expressed in terms of bias [8].

According to ISO 5725-4, an overall mean value x_{mean} representative of the comparison results (*method* in the ISO standard) was evaluated for each measurand after applying Cochran and Grubbs tests. Thus, the mean value of each measurand was compared with the corresponding reference value x_{ref} to evaluate the *bias of the method* (see Eq. 4)

$$Bias = x_{mean} - x_{ref} \quad (4)$$

The reference value x_{ref} is discussed in § 5.4 of ISO 5725-4 [8] and in § 7 of ISO 13528 [9].

The standard uncertainty of the bias was evaluated according to Eq. 5 (see § 5.5.3.2 of ISO 5727-4)

$$u(Bias) = \sqrt{\frac{S_R^2 - (1 - \frac{1}{n})S_L^2}{p} + u^2(x_{ref})} \quad (5)$$

The two components in the equation are: the precision of the method estimated with standard deviation S_R and the measurement uncertainty $u(x_{ref})$ of the reference value. It should be noted in Eq. 5 that the number p of participating laboratories plays a central role, in particular for the significance of the estimate.

5.2 Application of the ISO 13528

The proficiency test result of a laboratory can be assessed by a z score according to Eq. 6

$$z \text{ score} = \frac{x_{lab} - x_{ref}}{\sqrt{u^2(x_{ref})}} \quad (6)$$

Zeta scores can also be useful to evaluate a laboratory's capability to have results close to the reference value within its stated uncertainty. It was computed using Eq. 7

$$zeta \text{ score} = \frac{x_{lab} - x_{ref}}{\sqrt{u^2(x_{lab}) + u^2(x_{ref})}} \quad (7)$$

Z score indicates agreement among the results (or the lack of it), whereas zeta score shows coherence of the measurement uncertainty assessment with respect to the reference uncertainty (e.g. an underestimated uncertainty).

Z scores and zeta scores lying outside the interval $[-2, 2]$ are considered questionable, while those lying outside the interval $[-3, 3]$ need further analysis for explaining the causes, and possibly correcting the deviation from the reference value.

5.3 Statistical analysis of the results of the comparison of XCT systems

LNE analysed the measurement results provided by the participants using ISO 5725 parts 2 and 4 to evaluate the performance of XCT systems, and some criteria of ISO 13528 to evaluate the performance of each participant.

In this study, to guarantee the confidentiality of the participants, laboratories were identified by numbers.

When an outlier was found using Cochran and Grubbs tests, the interested laboratory was contacted for explanation, and to discuss the possibility of removal.

The uncertainties given by the laboratories were compared with those associated with the reference values, to check that the first ones were higher than the last ones. When possible, we analyzed the uncertainty factors identified by the laboratories.

For the reference values, it was recommended to measure the transfer standards with the CMM reference method before and immediately after the inter-laboratory test (§ 7 of ISO 13528 [9]). However, this could not be possible because of the Covid pandemic. Therefore, we did not follow the standard procedure, which proposes to estimate additional uncertainty components

associated to these instabilities.

When, in this study, too few laboratories provided results and repeated measurements for some of the transfer standards, we did not combined S_R and $u(x_{ref})$ to compute the uncertainty associated to the bias, $U(B)$. Instead, we only compared the magnitude of these two components to the bias. For example, a bias lower than the reference value uncertainty may be considered as non-significant. Also, a bias higher than the reference value uncertainty but lower or similar to S_R may be considered as non-significant.

The analysis results for the standard A11 are given below in Tables 3 to 6. Laboratory 9 has been eliminated with the Grubbs test for measurands M1, M31 and M32. Thus, its results were not taken into account to compute the general mean and the reproducibility standard deviation S_R .

Table 3 gives the overall mean values x_{mean} of all laboratories' measurements, the reproducibility S_R , the reference values x_{ref} as well as their uncertainties $U(x_{ref})$ and bias and Table 4 the uncertainty associated with the mean value for A11.

Table 3: General means and bias of the comparison for A11.

Measurands	M1	M2	M31	M32	M33	M34	M4
x_{mean} /mm	28.5723	0.2849	0.6026	0.6082	0.6112	0.6161	0.3989
S_R /mm	0.006	0.006	0.005	0.005	0.012	0.014	0.01
x_{ref} /mm	28.5696	0.2795	0.6038	0.6107	0.616	0.6211	0.3909
$u(x_{ref})$ /mm	0.0011	0.0025	0.0038	0.0107	0.016	0.0211	0.0091
$Bias$ /mm	0.001	0.007	0.0005	-0.002	-0.007	0.007	0.004

Table 4: Uncertainty associated with the mean value for A11. Note: the uncertainty associated to reference value M1 is given with $k = 2.4$ but, to compute $U(B)$, $k = 2$ was used.

Measurands	M1	M2	M31	M32	M33	M34	M4
nb lab	10	10	10	10	10	10	7
eliminated lab	1	0	1	1	0	0	0
nb lab kept p	9	10	9	9	10	10	7
average n	2.1	2	2.1	2	2	2	2.4
S_L^2/p /mm ²	3.7E-06	3.6E-06	2.9E-06	3.0E-06	1.5E-05	2.0E-05	1.0E-05
$S_r^2/(n \times p)$ /mm ²	2.0E-07	1.1E-08	1.4E-07	4.0E-08	9.5E-08	7.8E-08	1.5E-06
$S_R^2(x_{mean})$ /mm ²	3.9E-06	3.6E-06	3.0E-06	3.0E-06	1.5E-05	2.0E-05	1.2E-05
$S_R(x_{mean})$ /mm	0.0020	0.0019	0.0017	0.0017	0.0039	0.0044	0.0034
$u(x_{ref})$ /mm	0.0005	0.0013	0.0019	0.0054	0.0080	0.0106	0.0046
$u(Bias)$ /mm	0.0020	0.0023	0.0026	0.0056	0.0089	0.0114	0.0057
$U(Bias)$ /mm	0.0041	0.0045	0.0052	0.0113	0.0178	0.0229	0.0114

The z scores and zeta scores evaluated for each laboratory are given in Tables 5 and 6 for the aluminium transfer standard A11. In addition, the zeta scores are represented graphically in Figure 4.

Table 5: Z scores for All.

Laboratory	M1	M2	M31	M32	M33	M34	M4
1	18.81	3.12	-0.82	-0.27	-0.06	-0.18	1.27
2	7.42	-0.4	-2.53	-0.5	-0.13	-0.48	-0.64
3	-27.49	0.4	3.25	1.74	0.5	1.79	2
4	-0.58	2.53	-3.05	-1.02	-0.54	-0.38	0.73
5	7.42	4.4	0.63	0.24	0.13	-0.01	-2.62
6	5.24	0.05	-0.42	-0.69	-0.13	-0.2	0.68
7	-5.67	12.4	5.89	0.06	-1.38	-2.19	
8	13.96	14,00	0.63	-1.81	-2.5	-2.85	
9	-304.58	8.4	-23.05	-5.74	-4.5	-2,00	
10	3.34	3.71	-1.25	-0.68	-0.41	-0.31	4.39

Table 6: zeta scores for All.

Laboratory	M1	M2	M31	M32	M33	M34	M4
1	0.86	0.43	-0.17	-0.14	-0.04	-0.13	0.55
2	0.68	-0.07	-0.82	-0.42	-0.11	-0.44	-0.25
3	-2.79	0.11	1.27	1.33	0.44	1.65	1.42
4	-0.08	0.39	-1.31	-0.82	-0.48	-0.35	0.29
5	0.17	0.27	0.06	0.06	0.05	0,00	-0.58
6	0.94	2,00	-0.2	-0.58	-0.11	-0.19	0.57
7	-1.02	5.55	3.57	0.05	-1.31	-2.13	
8	1.41	3.75	0.25	-1.39	-2.18	-2.62	
9			-1.75	-1.2	-1.37	-0.78	
10	0.42	1.26	-0.6	-0.57	-0.38	-0.3	3.46

The zeta scores of the laboratories 7, 8 and 10 are out of the interval [-3,3] denoting an underestimation of their uncertainty. Other six laboratories (1, 2, 4, 5, 6 and 9) have zeta scores included in the interval [-2,2]. The results given by these laboratories were considered in compliance with the reference values. Thus, the overall comparison outcome could be considered satisfactory.




Figure 4: Zeta scores for All

Individual analysis for each standard:

- A11: the method seems biased only for M2. In addition, there is a strong dispersion of the results because the relative standard deviations are greater than the biases. Six laboratories out of nine have zeta scores in the interval [-2,2]. The XCT method can be considered well controlled with the A11 transfer standard.
- ABS2: the method seems biased only for measurand M2. The results are more dispersed than biased. The laboratory 5 has shown good performances with zeta scores in the range [-2,2].
- SS1: due to the high density of the material, only three laboratories were able to scan the transfer standard and provide the measurements. Consequently, it is difficult to draw conclusions. However, we noticed that the method seems biased for measurand M4 due to a high bias and a reference value lying outside the range of the laboratory's results. For the other measurands, the method provided results more dispersed than biased. Laboratory 5 has shown good performances with zeta scores in the interval [-2,2].

6 Conclusions

This paper presents the statistical analysis of an XCT dimensional measurement campaign involving several participants from all over Europe: Zeiss (DE), Novo Nordisk (DK), DTI (DK), NPL (UK), Messtronik (DE), CEA-List (FR), MTC (UK) and the University of Nottingham (UK). Three transfer standards were designed and machined with the purpose of being representative of AM-manufactured components. Particular relevance was given to the capability of XCTs to detect internal features in mono- or multi-material elements. A technical XCT measurement protocol was drafted and sent to all the participants beforehand. No strict instructions were given concerning the scanning, such as the XCT parameters or the number of scans. The participants were to measure according to their experience, selecting all XCT parameters for the best performance in terms of dimensional measurements and uncertainty, i.e. in usual conditions of XCT service providers.

Most of the laboratories stated values and uncertainties consistent with the reference values. Nevertheless, mainly considering the different available maximum voltage of the instruments involved in the comparison, this study also highlighted the following:

- The number of results available for SS1 was insufficient to provide statistically meaningful results.
- The XCT systems used exhibited different performances. This led either to the elimination of some laboratory's results, or to a high reproducibility.
- Except few cases, no repeated measurements were available. Thus, it was necessary to adapt the calculations described in the series ISO 5725, resulting in larger estimated uncertainties for the bias and the reproducibility. This suggests that a minimum number of XCT scans should have been indicated explicitly in the technical protocol, to be sufficient for a rigorous statistical analysis.
- Very heterogeneous uncertainties were stated by the participants, which in some cases were also underestimated. Even though the sources of uncertainty were explicitly requested, they were provided by only few participants. Therefore, it is not possible to add specific information in this regard.

Acknowledgements

The authors would like to thank Zeiss, Novo Nordisk, DTI, NPL, Messtronik, CEA-List, MTC and the University of Nottingham for participating to the comparison campaign of XCT machines to perform dimensional measurements.

This work was supported by the EMPIR programme, co-financed by the Participating States and from the European Union's Horizon2020 research and innovation programme.

References

- [1] ISO TC261/ASTM F42, JG59, Additive manufacturing—General principles—Non-destructive testing of additive manufactured products, International Organization for Standardization, On balloting 2020.
- [2] A. Thompson, I. Maskery, R. K. Leach, X-ray computed tomography for additive manufacturing: a review, *Meas. Sci. Technol.* 27 (2016) no. 7, 072001, doi: 10.1088/0957-0233/27/7/072001.
- [3] <https://www.ptb.de/empir2018/advanct/project/overview/> 2021 (accessed 23 June 2021).
- [4] A-F. Obaton, C. Gottlieb Klingaa, C. Rivet, K. Mohaghegh, S. Baier, J. Lasson L. Carli, L. De Chiffre, Reference Standards for XCT Measurements of Additively Manufactured Parts, *ndt.net*, id152 (2020), https://www.ndt.net/article/ctc2020/papers/ICT2020_paper_id152.pdf.
- [5] A. du Plessis et al., Laboratory X-ray tomography for metal additive manufacturing: Round robin test, *Add. Manuf.*, 30 (2019), 100837.
- [6] S. Carmignato, W. Dewulf, R. Leach, *Industrial X-Ray Computed Tomography*, first ed., Springer, Switzerland, 2018.
- [7] ISO 5725, Accuracy (trueness and precision) of measurement methods and Results, Part 2: Basic method for the determination of repeatability and reproducibility of a standard measurement method (2019).

- [8] ISO 5725, Accuracy (trueness and precision) of measurement methods and Results, Part 4: Basic methods for the determination of the trueness of a standard measurement method (2020).
- [9] ISO 13528, Statistical methods for use in proficiency testing by interlaboratory comparison, 2015.
- [10] BIPM, International vocabulary of metrology – Basic and general concepts and associated terms (VIM), third ed., France, JCGM 200:2012.