



## Controlling the thermal gradient during in situ transmission electron microscope heating experiments

Yang, Yi-Chieh; Vijayan, Sriram; Jinschek, Joerg

*Publication date:*  
2022

*Document Version*  
Publisher's PDF, also known as Version of record

[Link back to DTU Orbit](#)

*Citation (APA):*

Yang, Y.-C., Vijayan, S., & Jinschek, J. (2022). *Controlling the thermal gradient during in situ transmission electron microscope heating experiments*. Abstract from 4th East-Asia Microscopy Conference, Taipei, Taiwan, Province of China.

---

### General rights

Copyright and moral rights for the publications made accessible in the public portal are retained by the authors and/or other copyright owners and it is a condition of accessing publications that users recognise and abide by the legal requirements associated with these rights.

- Users may download and print one copy of any publication from the public portal for the purpose of private study or research.
- You may not further distribute the material or use it for any profit-making activity or commercial gain
- You may freely distribute the URL identifying the publication in the public portal

If you believe that this document breaches copyright please contact us providing details, and we will remove access to the work immediately and investigate your claim.

# Controlling the thermal gradient during in situ transmission electron microscope heating experiments

Yi-Chieh Yang<sup>1</sup>, Sriram Vijayan<sup>2</sup> and Joerg R. Jinschek<sup>1</sup>

Presenting author: Yi-Chieh Yang  
Email: yicyang@dtu.dk

<sup>1</sup> National Center for Nano Fabrication and Characterization (Nanolab), Technical University of Denmark (DTU), Kgs. Lyngby, Denmark

<sup>2</sup> Center for Electron Microscopy & Analysis, The Ohio State University, 1305 Kinnear Road, Columbus, OH 43212, USA.

**Keywords:** in situ heating, in situ TEM, nonequilibrium processes, thermal gradient, thermometry

Recent advances in in situ transmission electron microscopy (TEM) technology enable routine high-resolution observation of materials under conditions close to their application environment, thereby providing profound insights into the correlations between materials microstructures and properties at elevated temperatures. Nowadays, microelectromechanical system (MEMS)-type microheaters enable very accurate control of local temperature control during the in situ TEM experiment. So far, however, these experiments can only be carried out with a homogeneous temperature distribution over the area of interest.

Here, we report a method to introduce thermal gradients in in situ experiments, relevant, e.g. to study metal materials under far-from-equilibrium process conditions, by modifying a commercially available MEMS-based microheater. Using a focused ion beam (FIB), the sample is placed over a specially cut window not on top but near the heater <sup>[1]</sup>.

COMSOL simulations are performed to study the effect of window geometry on the established thermal gradient and to tune and control the thermal gradients across the TEM sample. To verify the introduction of the thermal gradient on a Si FIB lamella, the Ag sublimation approach <sup>[2]</sup> is applied for profiling of the temperature across the specimen. The measured gradient was found to be  $\sim 6.3 \times 10^6$  K/m <sup>[3]</sup>. Further simulations indicate that the thermal gradients can reach up to  $\sim 10^7$  K/m at the setting temperature of 1400 K. The good agreement between simulation and experiment encourages further detailed investigations of microstructures under controlled thermal gradients.

## References:

[1] Vijayan, S., Jinschek, J.R., Kujawa, S., Greiser, J. and Aindow, M.. Focused ion beam preparation of specimens for micro-electro-mechanical system-based transmission electron microscopy heating experiments. *Microscopy and Microanalysis*, 23 (2017) 708-716.

[2] Vijayan, S., Aindow M., Temperature calibration of TEM specimen heating holders by isothermal sublimation of silver nanocubes, *Ultramicroscopy* 196 (2019) 142-153.

[3] Vijayan, S., Wang, R., Kong, Z., Jinschek, J. R., Quantification of extreme thermal gradients during in situ transmission electron microscope heating experiments, *Microscopy Research and Technique*, 85.4 (2022) 1527-1537.